



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 173946

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>CHC palode</i>	Call Registration Date : <i>16/06/2024</i>
Address <i>Trivandrum</i>	Caller ID : <i>117472</i>
Ph : <i>9496279239</i>	Date of Visit : <i>23/06/2024</i>
	Asset No. : <i>0130644</i>
	EQPT Name : <i>Nebulizer</i>
	Manufacture <i>Troucheik</i> Model : <i>NA</i>
	S. No. <i>NA</i> Dept. <i>OP</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *not working
sums not getting*

Action Taken : *checked the machine and found piston broken, power cord damaged, nozzle broken need to replace these spares for further checking and working condition of the equipment.*

Completed Date : *23/06/2024* Time : *6:30pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. _____	_____	_____	_____
2. _____	_____	_____	_____
3. _____	_____	_____	_____

Cyrux Engineer	Date	Start Time	End Time
<i>midhun BM</i>	<i>23/06/24</i>	<i>6:00pm</i>	<i>6:30pm</i>

Customer Remark Completed Pending

Office closed.

Service Engineer Name : *midhun BM*
Signature : *[Signature]*
Date : *23/06/2024*
Contact Number : *9847467987*

Customer Name : *Almirah*
Signature : *[Signature]*
Date : *23/06/24*
Contact Number : *9847467987*
Designation : *Nursing Officer*
Hospital Seal : *[Seal]*

DR. P.R. MANOJ KUMAR
Medical Officer, Asst. Surgeon
BLOCK FHC, PALODE

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	CHC PALODE
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Nebuliser Make : Truecheck Model : N8001 Serial No : NA
4	Equipment ID/ Barcode	117472/0130644
5	Date of purchase/ Year of manufacture/Installation Date	13/10/2020
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	16/06/2024(Toll free)
9	Action taken	Checked and found piston, nozzle and powercord defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

DR. P. S. MANOJ KUMAR
 Medical Officer
 BLOCK III, PALODE

13	Asset Value	1353/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found piston, nozzle and powercord defective. The equipment was installed on 13/10/2020 and covered upto 3+ years. Since the spares are not available in the market. Recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	MIDHUN BM

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NIIM:


Since the major parts are damaged, this can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


10.7.2024
Signature of JC BM (NIIM)

Date:




9/07/24
Signature of
Superintendent/Medical Officer (i/c)
Dr. P.R. MANOJ KUMAR
Rg.No, 37562
Medical Officer, Asst. Surgeon
BLOCK FHC, PALODE




Nebuliser Electric ✓

Date	Description	Receipt	Account	Balance	Remarks
12/10/2020	Rd from store Machine No. 80013301	1	(Cognate No. 013064)	3	
	To: out to unmovable assets			1	
27/11/2020	Rd from store [0130806] Contract pay No. 9 (10p, 100w)	1		2	PKI S.No.
27/11/2020	Rd from store	1		3	
27/11/2020	To: out to FW - Equipment Register page (61)			2	PKI S.No.

[Signature]

Dr. P.R. MANOJ KUMAR
 Rg: No, 37562
 Medical Officer, Asst. Surgeon
 BLOCK FHC, PALODE





Dr. P.R. MANOJ KUMAR
Rg:No, 37562
Medical Officer, Asst. Surgeon
BLOCK FHC, PALODE