



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. :

223851 544

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>Phc</u></p> <p>Address : <u>Vamanapuram</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>9766340677</u></p>	<p>Call Registration Date : <u>13/06/2024</u></p> <p>Caller ID : <u>116972</u></p> <p>Date of Visit : <u>22/06/2024</u></p> <p>Asset No. : <u>01H1438</u></p> <p>EQPT Name : <u>Pharmacy Refrigerator</u></p> <p>Manufacture : <u>Samsung</u> Model : <u>NA</u></p> <p>S. No. : <u>NA</u> Dept. : <u>Observation Room</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** NOT WORKING

**Action Taken :** Checked and found Compressor, Freezer, Thermostat defective - Need to Replace these spare for further checking and working condition of the Equipment - Enquired spares with OEM


Completed  Date : 22/6/24 Time : 12PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Deepak. E.M</u>	<u>22/6/24</u>	<u>11:30</u>	<u>12PM</u>

**Customer Remark** Completed  Pending



**Service Engineer Name :** Deepak. E.M  
Signature : [Signature]  
Date : 22/6/24  
Contact Number : 7902881038


**Customer Name :** [Signature]  
Signature : [Signature]  
Date : 22/6/24  
Contact Number : 8547059573  
Designation : Nursing Officer Govt  
Hospital Seal : [Signature]

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

SL No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	PHC VAMANAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name: Pharmacy refrigerator Make : Samsung Model : NA SL : NA
4	Equipment ID/ Barcode	116972 / 0141438
5	Date of purchase/ Year of manufacture/Installation date	14/03/2007
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	13/06/2024
9	Action taken	Checked and found compressor, freezer, Thermostat defective. Need replace these spares for further checking and working condition of the equipment. Enquired spares with OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Compressor - 3200/- Freezer - 1300/- Thermostat - 500/- Gas charging - 1000/- Transportation - 500/- Service charge - 1500/- Total - 8000/-

13	Asset Value	11990/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.7%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and Vendor Quotation attached.
16	Reasons for recommending the equipment as BER	Checked and found compressor, freezer and Thermostat defective. The equipment was installed on 14/03/2007 and outlived upto 17+ years. The repairing cost is 66.7%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	DEEPAK EM 


\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:



Compressor, freezer and thermostat defective.  
So can be BER as per guidelines.

**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
6-8-24  
Signature of JC BM (NHM)

Date: 6/8/24





  
**Signature of**   
**Superintendent/Medical Officer (i/c)**

# STOCK BOOK OF MEDICINES

61

Name Of article.....

Refecagatorol

Date	Receipt	Issue	Balance	Date of Expiry	Price Rs.    Ps	Name of the Firm and Date of supply	Remarks
							P/F 79
14-3-2007			1			14.3.07	op.
		Donated by Dr. Jambala Jeevi					
							Bar code 69013726 Number 0141438
				 6/8/24.			
				മലയാളം ഹിന്ദി സർ ഇൻ ചാർജ്ജ് മെഡിക്കൽ ഓഫീസർ, മെഡിക്കൽ കോളേജ് വാമനപുരം			

Phone. 9188253055, 9745023891

# SUPER COOL

AIR CONDITION & REFRIGERATION  
REPAIRING & SERVICING  
Near Police Station, Venjaramoodu.

No 11

**CASH BILL**

Date 22/06/24

Customer Name:


Address:

Phone:

*Quotation only.*

Sl No	Particulars	Amount	
		Rs	Ps
1.	Aircondition Unit Service/ Repair		
2.	A.C. Unit Gas Charging		
3.	Fridge Repairing And Gas Charger		
4.	Freezer/Chiller/Water Cooler		
	Repairing and Servicing		
5.	Spare Parts.....		
6.	Labour Charge		
7.	Job Attend Charge		
8.	Others.....		
<b>GRAND TOTAL</b>			

*Handwritten notes in table:*  
 Compressor - 3200  
 Freezer - 1300  
 Thermostat - 500  
 gas - 1000  
 Transport - 500  
 Service charge - 1500  
 Total: 2500




*Signature:* 

Name: *Shaji*



*Handwritten signature*

ബഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്  
സ്റ്റാഫ് ക്ലിനിക്കൽ ഓഫീസർ  
വെണ്ണമൂട്



*(Signature)*  
 ഓഡിറ്റർ റെ-വീസർ ഇൻ ചാർജ്ജ്  
 കേരള ഓഡിറ്റർ ജനറൽ ഓഫീസ്  
 തിരുവനന്തപുരം