


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Thrissur
2	Name of Hospital	CHC VATANAPPALLY
3	Name of Equipment with Make, Model and Serial Number	Oxygen Concentrator Make: N/A Model: NA SN: 2106223422016
4	Equipment ID/ Barcode	0831028 #97578
5	Date of purchase/ Year of manufacture/Installation Date	29-07-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 22-02-24
9	Action taken	Checked the machine and found that sieve bed, compressor motor and main board defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. Installation date is 29-07-2021 Unfortunately, the supplier information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Details of Supplier not available


 Consultant (Biomedical)
 NHM, Thrissur



Signature of JC BM (NHM)

Date:



Signature of
 Superintendent/Medical Officer (i/c)


 SUPERINTENDENT
 BLOCK FAMILY HEALTH CENTRE
 VATANAPPALLY, P. THRITHALLUR
 PIN: 680 619



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
214080

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC</u></p> <p>Address <u>Vadanapilly</u> <u>Thirissur</u></p> <p>Ph : <u>9656854355</u></p>	<p>Call Registration Date : <u>22/02/24</u></p> <p>Caller ID : <u>97578</u></p> <p>Date of Visit : <u>24/02/2024</u></p> <p>Asset No. : <u>0831028</u></p> <p>EQPT Name : <u>O2 concentrator</u></p> <p>Manufacture <u>N/A</u> Model : <u>2Y-102W</u></p> <p>S. No. <u>210622342</u> Dept. <u>Pain & Palliative</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : equipment not getting on

Action Taken : checked the equipment and found that sive bed, compressor motor and main board defective. need to replaced these spares for further checking.

Completed Date : 24/2/24 Time : 3:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>24/2/24</u>	<u>2:00pm</u>	<u>3:00pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Christy</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>24/02/24</u></p> <p>Contact Number : <u>6238844347</u></p>	<p>Customer Name : <u>Liji V.S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>24/2/24</u></p> <p>Contact Number : <u>9656854</u></p> <p>Designation : <u>Palliative</u></p> <p>Hospital Seal :</p>
---	--



SUPERINTENDENT
[Signature]
BLOCK FAMILY HEALTH CENTRE
VATANAPILLY, PO. PUNITHURA
TRIP 680619

supply details enquiry - 0832028 - chc vatanapally

2 messages

Bijo T Joy <bijocyrix@gmail.com>

Thu, 4 Jul, 2024 at 8:21

To: KMSCL <kmsclbmecentral1@gmail.com>, bmetsr2020@gmail.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

dear mam,

kindly provide the supply detail of below mentioned oxygen concentrator equipment - oxygen concentrator(donation)

hospital - chc vatanapally

barcode - 0831028

caller id - 97578

make - nil

model - ZY-10ZW

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

KMSCL <kmsclbmecentral1@gmail.com>

Thu, 4 Jul, 2024 at 10:03

To: Bijo T Joy <bijocyrix@gmail.com>

The supply details for the above mentioned equipment were not available with us. Please refer to the hospital stock data for additional details.

[Quoted text hidden]

Regards,

SALU U S

DISTRICT BIOMEDICAL ENGINEER

KMSCL

Contact No: +91 91889 10013



+

-

T


POWER

10
8
6
4
2
0
L/min

BLOCK FAMILY HEALTH CENTRE
THRITHALLUR
680 619
VATANAPPALLY

SUPERINTENDENT
FAMILY HEALTH CENTRE
P O THRITHALLUR
680 619

BLOCK
VATANAPPALLY




CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
HEALTH CARE UNIT LTD PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0831028**




SUPERINTENDENT
BLOCK FAMILY HEALTH CENTRE
VATANAPPALLY, P O THRITHALLUR
PIN: 680 619



Model No : 10ZW

Date Of Manufa : 22th. 06. 2021



Serial Number: 2106223422016 2204/50Hz


SUPERINTENDENT
BLOCK FAMILY HEALTH CENTRE
VATANAPPALLY, P O THRITHALLUR
PIN: 680 619



Slno	Date	Reg no	Name & Address	Age	Whom Received Photo Address no, Signature
	29/07/21	RLO	From HI		
1	22/09/21	27/21	Shameer chekkampadivil P.O. Thottattur	51/2	Manju prembal Ward member 9526901835
2	01/10/21	13/18	Ranchana Tharayampalambil P.O. Vattarappilly ph:- 9847091936	62/2	Rishid 807 ph 9847091936
3	11/4/22	63/22	Sunitha Peringal P.O. Vattarappilly ph	40/2	Anjana Daughter in law 8590068964
4	10/10/22 10/10/22	166/22	Shanmugas Chennara P.O. Vattarappilly ph:- 7510244103	67/2	Raman 7025338604
	21/10/22		RLO from HI Page no:- 99.		
5	21/10/22	174/22	Santha Mepalambil P.O. Vattarappilly W-15	68/2	Sunil
	17/08/23		Returned by Santha mepalambil		



SUPERINTENDENT
BLOCK FAMILY HEALTH CENTRE
VATANAPPALLY, P.O. THRISUR
PIN- 680 639