


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC NALUKETTU
3	Name of Equipment with Make, Model and Serial Number	Weighing Machine Make : SAMSO Model : NA Sr.No- NA
4	Equipment ID & Barcode	0841327 - #123108
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date – NA Actual installation date - 13-11-2019 (Invoice attached for reference)
6	Warranty details (Yes/No)	NA
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 15-07-2024
9	Action taken	Checked the machine and found board and sensor are faulty. Need to replace the board and sensor for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	5000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report
16	Reasons for recommending the equipment as BER	Checked equipment found board and sensor damaged, equipment installed on 13-11-2019 and covered up to 4 years 7 months, Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	BIJO T JOY 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Space not available*

  
Signature of JC BM (NHM)

  
Superintendent Medical Officer (i/c)  
Medical Officer in Charge  
Family Health Centre  
Nalukettu, Pin: 680 308

  
Signature of seal

Date





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :**  
224671

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>phc</u></p> <p>Address : <u>nalukettu</u> <u>Thrissur, Kerala</u></p> <p>Ph : <u>8590640067</u></p>	<p>Call Registration Date : <u>15/7/24</u></p> <p>Caller ID : <u>123108</u></p> <p>Date of Visit : <u>16/7/24</u></p> <p>Asset No. : <u>0841327</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>SAMCO</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Public health</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** weight reading error

**Action Taken :** checked the machine and found board and sensor are faulty. need to replace the board and sensor for further checking.


Completed  Date : 16/7/24 Time : 1:30PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. <u>board</u>	1		
2. <u>sensor</u>	1		
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jinto Joy</u>	<u>16/7/24</u>	<u>12:30PM</u>	<u>1:30PM</u>


**Customer Remark** Completed  Pending



<p><b>Service Engineer Name :</b> <u>Jinto Joy</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>16/7/24</u></p> <p><b>Contact Number :</b> <u>7593847125</u></p>	<p><b>Customer Name :</b> <u>MERCY. P.V</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>16/7/24</u></p> <p><b>Contact Number :</b> <u>9995510132</u></p> <p><b>Designation :</b> <u>[Signature]</u></p> <p><b>Hospital Seal :</b></p>
--	--

Medical Officer in Charge  
Family Health Centre  
Nalukettu, Pin: 680









**KARUNYA MEDICINE DEPOT**

**SWARAJ ROUND EAST, TRISSUR**  
 PH: 0487-2975601 EMAIL : kcpdepotstr@gmail.com

State : Ker  
 State Code : 32  
 DL No1: KL-TSR-1394  
 DL No2: KL-TSR-1394  
**[ORIGINAL]**

**INVOICE**  
**GSTIN : 32AADCK4029M1ZK**

**Name** : TSR-Nalukettu-PHC  
**Cust.Code** : Naluk  
**Address** : Medical Officer I/c  
 Primary Health Centre - Nalukettu  
 Mukundapuram Taluk  
 Thrissure-680308  
**Phone No** :  
**GSTIN** :  
**DL No1** :  
**DL No2** :

**Invoice No** : 5492  
**Date** : 13/11/2019  
**Time** : 11:20 AM  
**Prepared By** : SINDHU K B

Page 2 of 2

Sl#	MFR	Product Name	KCP S.P	Exp.Dt	MRP	Batch	Unit	Qty	Rate	CGST %	CGST Amt	SGST %	SGST Amt	Amount
2	GE	WEIGHING MACHINE (DIGITAL PAEDIATRIC) - PE	4239.44	Dec/2030	6000.00	111	NOS	1.0	4239.44	9.00	381.55	9.00	381.55	4239.44
3	GE	WEIGHING MACHINE (DIGITAL ADULT - SCALE CAPACITY 150 KGS) - PE	1608.04	Jan/2028	2590.00	113	NOS	1.0	1608.04	9.00	144.72	9.00	144.72	1608.04




<b>CGST:</b> 1352.56	<b>SGST:</b> 1352.56	<b>CESS:</b> 178.06	<b>Total:</b> 20688.69
			<b>Handling Charge:</b> 1246.39
			<b>Handling Charge Gst18% :</b> 224.35
			<b>Round off:</b> -0.43
<b>P Amount: 44465.00</b>			<b>Net Amount: 22159.00</b>

**Words:** Twenty Two Thousand One Hundred and Fifty Nine

**Terms & Conditions**

Less returns only within 10 days from the date of billing  
 Less return -10AM to 4PM Only  
 Sales return on Sunday  
 Products to be kept in refrigerator will not be taken back after sales.



**AUTHORISED SIGNATORY**