


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

| Sl. No | Particulars  | Details   |
|--------|--|---|
| 1      | Name of District   | THRISSUR  |
| 2      | Name of Hospital   | GH THRISSUR   |
| 3      | Name of Equipment with Make, Model and Serial Number                           | Dialysis Machine<br>Make: B Braun<br>Model: Dialog plus<br>SN:106292  |
| 4      | Equipment ID/ Barcode  | 0810427 #121814   |
| 5      | Date of purchase/ Year of manufacture/Installation Date                        | Ticket master installation date is 09-10-2013<br>As per stock book installation date is 01.08.2013  |
| 6      | Warranty details (Yes/No)  | No warranty   |
| 7      | *AMC/ CAMC Period agreed at the time of Purchase                               | No AMC/CAMC   |
| 8      | Date of breakdown (Date of registration of complaint through email/ Toll free) | Toll free-08-07-2024  |
| 9      | Action taken   | Checked the machine and Found switch mode power supply, pressure sensor, gear pump FPA, membrane, motor for gear pump, pump body, motor for piston pump, power board, digital board, gear pump EP, gear pump FPE, mother board, compact flash card and analog board and pressure monitor module are defective. Need to replace these spares for further checking.   |
| 10     | Present status of the equipment (Fully damaged / partially damaged)            | Fully damaged   |
| 11     | Recommendations for repair (required service details)                          | No recommendations  |
| 12     | Cost of spares (specify parts and cost)  | Switch mode power supply -189332/-<br>Pressure sensor -28332/-<br>Gear pump FPA -50413/-<br>Balance chamber membrane- 9323/-<br>Motor for gear pump -118089/-<br>Pump body - 80382/-<br>Motor for piston pump - 35014/-<br>Power board (valves) - 52864/-<br>Digital board version6 - 82308/-<br>Gear pump EP - 64988/-<br>Gear pump FPE - 48007/-<br>Mother board LX400 -115895/-<br>Compact flash card - 12986/-<br>Analog board - 60230/-<br>Pressure monitoring module - 50463/-<br><b>Total - 998626/-</b> |

|    |  |   |
|----|--|---|
| 13 | Asset Value  | 506667/-  |
| 14 | #Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value                | 197.21%   |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Service report and OEM quotation attached   |
| 16 | Reasons for recommending the equipment as BER  | Checked the machine and Found switch mode power supply, pressure sensor, gear pump FPA, membrane, motor for gear pump, pump body, motor for piston pump, power board, digital board, gear pump EP, gear pump FPE, mother board, compact flash card and analog board and pressure monitor module are defective. In ticket master installation date is 09.10.2013, As per stock book Equipment installed on 01-08-2013 (stock book attached for reference) aged up to 10 years 8 months. Spare cost is 197.21% of asset value. Both criteria for RBER met, So as per tender clause 5.3.14.1 the equipment recommending for condemnation |
| 17 | Name & Signature of CYRIX Authority  | BIJO T JOY   |


\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Spare cost ₹ 9,98,626/-

**Dr. Consultant (Biomed)**  
NHM, Thrissur

  
Signature of JC BM (NHM)



Date: \_\_\_\_\_

  
SUPERINTENDENT (Dy. DHS)  
GENERAL HOSPITAL  
THRISSUR - PIN: 680 001



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. :  
**224522**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

|  |  |
|--|--|
| <p>Health Facility ..... <u>GH</u> .....</p> <p>Address ..... <u>Thirissur</u> .....</p> <p>Ph : .....</p> | <p>Call Registration Date : ..... <u>08-07-2024</u> .....</p> <p>Caller ID : ..... <u>121814</u> .....</p> <p>Date of Visit : ..... <u>9-07-2024</u> .....</p> <p>Asset No. : ..... <u>0810427</u> .....</p> <p>EQPT Name : ..... <u>Dialysis machine</u> .....</p> <p>Manufacture ..... <u>B Braun</u> ..... Model : ..... <u>Dialog plus</u> .....</p> <p>S. No. ..... <u>106292</u> ..... Dept. ..... <u>Dialysis</u> .....</p> |
|--|--|

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... It's showing malfunction error .....

**Action Taken :** ..... checked the machine and found switch mode power supply, pressure sensor, gear pump, FPA, membrane, motor for gear pump, pump body, motor, PSIR pump, power board, digital board, gear pump E.P., gear pump, F.P.E, mother board, compact flash card and analog board and pressure monitor module are defective. Need to replace these. Spare for further service .....

Completed  Date : 9-07-2024 Time : ..... 2:00pm ..... Spare Required

Spare Replaced  Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1.          |      |             |           |
| 2.          |      |             |           |
| 3.          |      |             |           |

| Cyrix Engineer | Date             | Start Time    | End Time      |
|----------------|------------------|---------------|---------------|
| <u>Byo</u>     | <u>9-07-2024</u> | <u>1:00pm</u> | <u>2:00pm</u> |

**Customer Remark** ..... Completed  Pending

|  |   |
|--|---|
| <p>Service Engineer Name : <u>Byo</u></p> <p>Signature : </p> <p>Date : <u>9-07-2024</u></p> <p>Contact Number : <u>9921970910</u></p> | <p>Customer Name : <u>9</u></p> <p>Signature : </p> <p>Date : <u>9</u></p> <p>Contact Number : <u>9</u></p> <p>Designation : <u>214</u></p> <p>Hospital Seal : </p> |
|--|---|



**SUPERINTENDENT (Dy. DHO)**  
**GENERAL HOSPITAL**  
**THIRISSUR - PIN: 680 001**

Installation

PMS / TSC

Break Down

|   |  |   |  |
|---|--|---|--|
| Customer Name : <u>General Hospital</u>   |  | Service Request Category:<br><b>Installation</b>  |  |
| Customer Address : <u>Thrikkur</u>  |  | User Training checklist signed: <input type="checkbox"/> Y <input type="checkbox"/> N   | Installation checklist signed: <input type="checkbox"/> Y <input type="checkbox"/> N |
| Contact Number : _____  |  | Warranty End Date: _____  |  |
| Division  | BA <input checked="" type="checkbox"/> HC <input type="checkbox"/> AE <input type="checkbox"/>               | TSC cum PMS   |  |
| Service Place   | Onsite <input checked="" type="checkbox"/> Workshop <input type="checkbox"/> Office <input type="checkbox"/> | If Yes PM KIT replacement date: _____   | User Training: <input type="checkbox"/> Y <input type="checkbox"/> N                 |
| Model : <u>Dialox + FVO</u>   |  | User Training checklist signed: <input type="checkbox"/> Y <input type="checkbox"/> N   |  |
| Serial Number: <u>106292</u>  |  | Breakdown call  |  |
| Software version: <u>9.12</u> Working hours: _____  |  | Software: <input type="checkbox"/> Y <input type="checkbox"/> N   |  |
| Observations :  |  | Electrical Hardware: <input type="checkbox"/> Y <input type="checkbox"/> N  |  |
| RO TDS: ___ (0-10PPM Acceptable limits) <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA   |  | Mechanical Hardware: <input type="checkbox"/> Y <input type="checkbox"/> N  |  |
| Neutral to Earth Voltage: ___ (0-2 Volt)  |  | Trend backup: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA                               | Service Type   |
| Use of UPS / Stabilizer: ___ (recommended 3KVA per machine in Dialog mode) <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA  |  | 50% Citric Disinfection after each Therapy: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA | Warranty / Extended warranty: <input type="checkbox"/>                               |
| Bicarb mixer : <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA  |  | 50% Citric Disinfection after end of day: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA   | CAMC : <input type="checkbox"/>  |
|   |  |   | AMC : <input type="checkbox"/>   |
|   |  |   | Oncall & Chargable : <input checked="" type="checkbox"/>                             |
| Reported Problem: <u>machine not switching on</u>   |  |   |  |
| Observed Problem: _____   |  |   |  |
| Service Resolution: <u>checked the machine need to replace the S.M.P.S.I, Pressure sensor - 2, Gear Pump EP - 1, Gear Pump PPA - 1, Gear Pump PPF - 1, motor for Gear Pump - 3, membran for dialy chamber - 1, Digital board - 1, Analogy board - 1, Power board valve - 1, Pump body - 2, motor for piston pump - 2, motor board - 1, P.F/P.V. modul - 1, motor board L200 - 1, C.F. card - 1,</u> |  |   |  |
| Following spareparts are :  |  | Replaced : <input type="checkbox"/> Y <input type="checkbox"/> N  | To be replaced : <input type="checkbox"/> Y <input type="checkbox"/> N               |
|   |  | To be quoted: <input type="checkbox"/> Y <input type="checkbox"/> N   |  |
| SN  | Article number   | Article Name  | Quantity   |
|   |  |   |  |
|   |  |   |  |
|   |  |   |  |
| Date  | Start time (Hrs : Min)   | End time (Hrs : Min)  | Labour Hours Spent: (Hrs : Min)  |
|   |  |   |  |
| This is to certify that above mentioned spareparts of B Braun machine are consumed / to be consumed on our in hospital for satisfactory working of machine.   |  |   |  |
| Status of Service request:  | Completed: <input type="checkbox"/> Y <input type="checkbox"/> N   | Under Observation: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA                          | Followup Date: _____   |
| Contact person: <u>Kijo Jon</u>   | Signature with Date & Time: <u>[Signature] 29/7/24 1:50pm</u>  | Stamp: <b>DIALYSIS CENTRE GENERAL HOSPITAL THRISSUR</b>   |  |
| Customer remark / Feedback: _____   |  |   |  |
| Engineer name & Signature: <u>[Signature]</u>   |  | Engineer Contact No.: <u>99160310</u>   |  |

To,  
Cyrix Healthcare  
C/O Thrissur

**Sub: - Quotation for Spare Parts**

**Our Ref.BB/BA-W-241/24 Date: 01.08.2024**

Dear Sir/ Madam,

Greetings from BBraun Medical India Pvt. Ltd.

Kindly find our most competitive rate for Spare part for Hemodialysis machine.

| Art. No. | Description                      | Qty | Unit price<br>( Excl. Tax) | Total Price<br>( Excl. Tax) |
|----------|----------------------------------|-----|----------------------------|-----------------------------|
| 3461076  | Switch mode power supplyMc2      | 1   | 189,332                    | 189,332                     |
| 3456124A | Pressure Sensor                  | 2   | 14,166                     | 28,332                      |
| 34562331 | Gear pump Fpe                    | 1   | 50,413                     | 50,413                      |
| 3451615A | Membrane for balance chamber     | 1   | 9,323                      | 9,323                       |
| 34560718 | MOTOR FOR GEAR PUMP              | 3   | 39,363                     | 118,089                     |
| 34560602 | PUMP BODY                        | 2   | 40,191                     | 80,382                      |
| 3456055B | MOTOR FOR PISTON PUMP            | 2   | 17,507                     | 35,014                      |
| 34562150 | POWER BOARD (VALVES)             | 1   | 52,864                     | 52,864                      |
| 34562303 | Digital board, version 6         | 1   | 82,308                     | 82,308                      |
| 34562332 | Gear pump Ep                     | 1   | 64,988                     | 64,988                      |
| 34562330 | Gear pump Fpa                    | 1   | 48,007                     | 48,007                      |
| 34560610 | MOTHERBOARD LX 800               | 1   | 115,895                    | 115,895                     |
| 3456088A | Compact Flash Card, 1 GB         | 1   | 12,986                     | 12,986                      |
| 34560734 | ANALOG BOARD                     | 1   | 60,230                     | 60,230                      |
| 34561641 | PRESSURE MONITORING MODULE PA/PV | 1   | 50,463                     | 50,463                      |
|          |                                  |     | <b>Total:</b>              | <b>998,626</b>              |
|          |                                  |     | <b>GST: 18%</b>            | <b>179,753</b>              |
|          |                                  |     | <b>Total Amount:</b>       | <b>1,178,379</b>            |

**Terms & Conditions:**

1. Taxes : As per GST Rate w.r.t. HSN code of spare part.
2. Payment : 100% Advance Cheque/DD in Favour of M/s B.Braun Medical (I)  
Pvt. Ltd.
3. Delivery : Within 2-3 weeks from PO.
4. Quotation : Validity 30 days


Thanking you,

Yours faithfully,

For B|Braun Medical (India) Pvt Ltd.



**SUSMEETA PANDEY**  
Senior Executive - Technical Service  
E Mail ~ [susmeeta.mazumdar@bbraun.com](mailto:susmeeta.mazumdar@bbraun.com)  
Phone No: 022 66682205



**BABU MURUGESAN**  
Manager -Training & Quality  
E Mail ~ [babu.murugesan@bbraun.com](mailto:babu.murugesan@bbraun.com)  
Mob No: 9884450417



**Purchase Order -KL 24 25 101194 -B. Braun Medical (India) Pvt. Ltd - Mumbai -GH THRISSUR-0810471**

8 messages

Anjitha T L <purchasecyrix2@gmail.com>

Tue, 16 Jul, 2024 at 9:34 am

To: Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Bijo T Joy <bijocyrix@gmail.com>

Dear Sir,

Please find the attached Purchase Order and do the needful. Kindly send a confirmation mail with details of material delivery (Docket details) along with the invoice Copy mentioning our PO No. (For reference) with Seal and Signature.

EDD as per payment term:15/07/24

**For Queries and Clarifications**

1. Technical: [dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in](mailto:dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in)
2. Accounts: [paymentcyrix@gmail.com](mailto:paymentcyrix@gmail.com)
3. Logistics: [logistics@cyrix.in](mailto:logistics@cyrix.in)
4. PO Clarification: [ampurchase@cyrix.in/purchasemgr@cyrix.in](mailto:ampurchase@cyrix.in/purchasemgr@cyrix.in)
5. Delivery address: **Cyrix Healthcare PVT LTD**

**Methanam Road**  
**Door No: 312/a Eloor North,**  
**Kerala - 683501- India**  
**Nikesh V**  
**Mob: 9072522014**

**@ Cyrix Accounts**

Payment Term: 100% adv

Thanks & Regards

Anjitha T L  
Sr. Purchase Executive  
PH:- 9947719111  
[purchasecyrix2@gmail.com](mailto:purchasecyrix2@gmail.com)  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

**Cyrix Healthcare Pvt Ltd**  
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India  
**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

Babu Murugesan <babu.murugesan@bbraun.com>

Wed, 24 Jul, 2024 at 5:18 pm

To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>  
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Bijo T Joy <bijocyrix@gmail.com>

Dear Anjitha,

Please share payment transfer details.

Thanks & Regards  
Babu Murugesan

B. Braun India is recognized as one of the Best Workplaces to Work in India



From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: Tuesday, July 16, 2024 9:34 AM

To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Afsal Yunuze <zm1.klbemp@cyrix.in>; DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Bijo T Joy <bijocyrix@gmail.com>

Subject: Purchase Order -KL 24 25 101194 -B. Braun Medical (India) Pvt. Ltd - Mumbai -GH THRISSUR-0810471

**CAUTION:** This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

Babu Murugesan

Manager - Technical Services

+919884450417

babu.murugesan@bbraun.com

B. Braun Medical (India) Pvt. Ltd. (CIN : U33112MH1984PTC214514)

WorkEZ - Block-B, Door No.3, 2nd Floor, Hansa Building , RK Swamy Centre, 147, Pathari Rd, Thousand Lights, Chennai, 600006 Tamil Nadu, IN, India

<http://www.bbraun.co.in>

[www.linkedin.com/company/bbraun-group](http://www.linkedin.com/company/bbraun-group)

As a true partner  
we develop smart solutions  
and set standards  
to improve health care.

B. Braun – Sharing Expertise

The information contained in this communication is confidential, may be attorney-client privileged, may constitute inside information, and is intended only for the use of the addressee. It is the property of the company of the sender of this e-mail. Unauthorized use, disclosure, or copying of this communication or any



part thereof is strictly prohibited and may be unlawful. If you have received this communication in error, please notify us immediately by return e-mail and destroy this communication and all copies thereof, including all attachments.

Cyrix Payments <paymentcyrix@gmail.com>

Wed, 24 Jul, 2024 at 5:24 pm

To: Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Bijo T Joy <bijocyrix@gmail.com>

Dear Sir/Mam,

Please note that we had transferred Rs. 27840/- as advance against PO101194 after deducting TDS Rs480/-

|  |            |           |                  |      |           |
|--|------------|-----------|------------------|------|-----------|
| B. Braun Medical<br>(India) Pvt. Ltd -<br>Mumbai | 0062620000 | 27,840.00 | AXISP00520948102 | PAID | 24-JUL-24 |
|--|------------|-----------|------------------|------|-----------|

[Quoted text hidden]

Thanks & Regards,

Tincy T Jose

Accounts Department.

**Cyrix Healthcare Pvt Ltd**

Office : 30/641 B, First Floor, Petta Junction, Poonithura, Cochin, Kerala - 682 038

Website:- [www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Mob : 7593847123

Anjitha T L <purchasecyrix2@gmail.com>

Thu, 25 Jul, 2024 at 10:06 am

To: Babu Murugesan <babu.murugesan@bbraun.com>, Anoop <anoop.karthikeyan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Bijo T Joy <bijocyrix@gmail.com>

Dear sir,

Please find the machine serial No: 104724

[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>

Thu, 1 Aug, 2024 at 4:11 pm

To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Bijo T Joy <bijocyrix@gmail.com>

Dear Anjitha,

Please find Service report, spares estimate & Invoice copy in the attached files.

Kindly issue PO for the required spare parts.

Thanks & Regards

**Babu Murugesan**

From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: Tuesday, July 16, 2024 9:34 AM

To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Afsal Yunuze <zm1.klbemp@cyrix.in>; DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP

<dm1.klbemp@cyrix.in>; **Cyrix Logistics** <logistics@cyrix.in>; **Kerala GRN** <keralagr@cyrix.in>; **Izrath K K** <purchasecyrix5@gmail.com>; **Bijo T Joy** <bijocyrix@gmail.com>

**Subject:** Purchase Order -KL 24 25 101194 -B. Braun Medical (India) Pvt. Ltd - Mumbai -GH THRISSUR-0810471

**CAUTION:** This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

**Dear Sir,**

[Quoted text hidden]

[Quoted text hidden]

**Izrath K K** <purchasecyrix5@gmail.com>

Fri, 2 Aug, 2024 at 4:39 pm

To: Kerala GRN <keralagr@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Bijo T Joy <bijocyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Hi,

Kindly take GRN

po:101194

@Bijo T Joy kindly give confirmation

Thanks & Regards

Izrath KK

Jr Purchase Executive

PH: 8848002238

[Quoted text hidden]

**Izrath K K** <purchasecyrix5@gmail.com>

Sat, 3 Aug, 2024 at 1:42 pm

To: Bijo T Joy <bijocyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>

Reminder

Thanks & Regards

Izrath KK

Jr Purchase Executive

PH: 8848002238

[Quoted text hidden]

**Bijo T Joy** <bijocyrix@gmail.com>


Sat, 3 Aug, 2024 at 2:49 pm

To: Izrath K K <purchasecyrix5@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>

Confirmed

[Quoted text hidden]



16

Stock Account of Furniture and other Office Stores

Name of article... Dialogues... Haemo Dialysis machines Office of.....

| Date   | Number and date of contingent voucher, invoice etc.   | Nature of transaction  | Receipts | Issues                         | Balance |
|--------|---|--|----------|--------------------------------|---------|
| 1/5/13 | KMSCL<br>TIN N. 9310000818 dated 30.11.13<br>B. Brown   | Order NO: KMSCL/EP/266<br>281/12 dated 30-1-13<br>No. 1802394/52 |          | (Entered page No 179 ud<br>70) |         |
|        | #10200N Dialog Blood pump 210   |  |          | RS 5.37 lac<br>26850           |         |
|        | D.O. NO: #1850171<br>Batch no 105898 (M10) Spt. Code  |  |          | (Per each machine)             |         |
| - 11   | SLNO: 105898 (M10) → 281-25   | Equipments   |          | Code - 281                     |         |
| - 2    | SLNO: 106358 (M8) → 281-27  |  |          |                                |         |
| - 3    | SLNO: 106292 (M9) → 281-26  |  |          |                                |         |
| - 4    | SL: 106311 (M7) → 281-28  |  |          |                                |         |
| 1.8.13 | KMSCL<br>IN. N. 9310000735 dated 8.12.12<br>order no: #432/12/281-31 dated 8.12.12<br>order no. 1802835672<br>#10200N Dialog + Blood pump |  |          |                                |         |
| - 1    | SLNO: 104719 (M3) 281-10  |  |          |                                |         |
| - 2    | SLNO: 104722 (M2) 281-8   |  |          |                                |         |
| - 3    | SLNO: 104724 (M4) 281-12  |  |          |                                |         |
| - 4    | SLNO: 105145 (M1) 281-7   |  |          |                                |         |
| - 5    | SLNO: 104351 (M5) 281-11  |  |          |                                |         |
| - 6    | SLNO: 104354 (M6) 281-9   |  |          |                                |         |
|        |   |  |          | RS 5.37 lac<br>26850           |         |
|        |   |  |          | (Per each machine)             |         |
|        |   |  |          | Total amount                   |         |
|        | (Warranty for 2016) 3yrs  |  |          | (10)                           |         |



2. Personnel with Sr