



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 220244

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>GH Pala</u>	Call Registration Date : <u>25/7/2024</u>
Address <u>kottayam</u>	Caller ID : <u>125 450</u>
Ph : <u>9495992990</u>	Date of Visit : <u>29/7/2024</u>
	Asset No. : <u>0511767</u>
	EQPT Name : <u>OT Light</u>
	Manufacture <u>Hospitech</u> Model : <u>Hospilite 1001</u>
	S. No. <u>Nil</u> Dept. <u>LR</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the machine Bulb, Bulb holder, Transformer and Bulb reflector are fully damaged and handles/Arms fully rusted not properly function.

Completed Date : 29/7/2024 Time : 11:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			
Cyrix Engineer <u>Josekutty George</u>		Date <u>29/7/2024</u>	Start Time <u>11am</u>
		End Time <u>11:30am</u>	

Customer Remark Completed Pending

Service Engineer Name : Josekutty George
Signature : [Signature]
Date : 29/7/2024
Contact Number : 8848230407

Customer Name : SINE-VC SNO LABOUR ROOM
Signature : [Signature]
Date : 29/7/24
Contact Number : 9495992990
Designation : Senior nursing officer
Hospital Seal :




REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	GH PALA
3	Name of Equipment with Make, Model and Serial Number	Equipment :OT LIGHT Make: HOSPITECH Model :1007 SN:NIL
4	Equipment ID & Barcode	125450 & 0511767
5	Date of purchase / Year of manufacture /Installation Date	3/02/2006
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	25/07/2024 (Toll free)
9	Action taken	Checked and found out the problem with transformer, lamp, lamp holder and handles arm need to be repaint because of rust in condition
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	Rs. 84515/- <u>9432.5</u>
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	<u>N/A</u>
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with transformer, lamp, lamp holder and handles arm need to be repaint because of rust in condition Machine installed in 3/2/2006 and covered up to 18 + years. So recommending the machine for condemnation
17	Name & Signature of CYRIX Authority	Josekutty George 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Asset value = 9432.5 / as per stock (15460 + 12.5-D)
 cost of spare not mentioned - transformer
 Lamp, lamp holder and handles arm rusted!!
 Life period - 18 years - the machine
 is very old and ~~rusty~~.


 Signature of JC BM (NHM)
 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION
 KOTTAYAM



Date 3/17/24 Seal

ABINASH
3/17
 Signature of
 Superintendent / Medical Officer (i/c)
 General Hospital
 Pala, Kottayam.

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - 0511767

