



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 232295

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THQH</u></p> <p>Address <u>Fore</u></p> <p>..... <u>Thiruvananthapuram</u></p> <p>Ph : <u>9497782744</u></p>	<p>Call Registration Date : <u>24/07/2024</u></p> <p>Caller ID : <u>125223</u></p> <p>Date of Visit : <u>30/07/2024</u></p> <p>Asset No. : <u>0122114</u></p> <p>EQPT Name : <u>B.P Apparatus</u></p> <p>Manufacture <u>Dr. odin</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Op</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found mainboard and pump defective. Need to replace these spares for further checking and working condition of the equipments


Completed Date : 30/7/24 Time : 10:30 AM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	NA	—	—	—
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
Akul	30/7/24	10:00 AM	10:30 AM

Customer Remark Pending

<p>Service Engineer Name : <u>Akul Manohar</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>30/7/24</u></p> <p>Contact Number : <u>8920568751</u></p>	 <p>Signature : <u>[Signature]</u></p> <p>Date : <u>30/7/24</u></p> <p>Contact Number : <u>8920568751</u></p> <p>Designation : <u>Superintendent</u></p> <p>Hospital Seal : <u>Taluk Hospital, Fort Thiruvananthapuram</u></p>
---	---

9497782744

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH FORT THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Make : Dr.Odin Model : NA Serial No : NA
4	Equipment ID/ Barcode	125223/0122114
5	Date of purchase/ Year of manufacture/Installation Date	07/05/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	24/07/2024(Toll free)
9	Action taken	Checked and found Mainboard and pump defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	2200/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found Mainboard and pump defective. The unit is installed on 07/05/2021 and covered up to 3+ years. Quotation not attached since the spares are not available in the market. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	AKHIL MANOHAR <i>Akhil</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Main board and Pump defective.
So can be RBER as per tender guidelines*

MANEESHA MOHAN M.B.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram
14-08-2024
Signature of JC BM (NHM)

Date:




[Handwritten Signature]
Superintendent
Signature of Taluk Hospital, Fort
Superintendent Thiruvananthapuram

Digital BP Apparatus

Date	Part Rd-	Dd	L	B	
17/6/2020	Received from store using 10 OPD	3	4	3	
7/5/2021	Received from store (Cost Rs. 1265/ per each one) (Entered in Stock register page 51)	2		5	(2) Working Cond/ of Contn of OPD
10/05/21	Received from store	1		6	
12/10/2022	Received from store (10) (Rak 1, 201-76)	2	7	8	OP
3/4/2021	Received from store P/insured 23-24 ins 00005/24-25 Rs 990.00/- per unit	5	4	13	
3/4/2021	Tr: out to uninv sub. Reg payments Rd date (3) 17/6/2020		3	10	



Superintendent
Taluk Hospital, Fort
Thiruvananthapuram



[Signature]
Superintendent
Taluk Hospital, Fort
Thiruvananthapuram