



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : **191439**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC Sachiothamapuram</u></p> <p>Address <u>Sachiothamapuram</u> <u>Kattayam</u></p> <p>Ph : <u>9496408368</u></p>	<p>Call Registration Date : <u>17/7/2024</u></p> <p>Caller ID : <u>123559</u></p> <p>Date of Visit : <u>19/7/2024</u></p> <p>Asset No. : <u>0530856</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>EZ Life</u> Model : <u>1303</u></p> <p>S. No. <u>N/A</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Fumes not coming

Action Taken : Found the motor of compressor of the equipment damaged. Need to replace the motor with compressor in order to check the working condition of the equipment.

Completed Date : 19/7/2024 Time : 11:25am Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sreyu Sreeh</u>	<u>19/7/2024</u>	<u>10:25am</u>	<u>11:05am</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Sreyu Sreeh</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>19/7/2024</u></p> <p>Contact Number : <u>7902854749</u></p>	<p>Customer Name : <u>Ravi P. John</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>19/7/2024</u></p> <p>Contact Number : <u>9496408368</u></p> <p>Designation : <u>SWO</u></p> <p>Hospital Seal :</p>
--	--




Medical Officer - in Charge
 Community Health Centre
 Sachiothamapuram
 Kottayam Dist. Pin - 686 532
 Kottayam Dist.

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	CHC SACHIVOTHAMAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Nebulizer Make : EZ Life Model :1303 SN:nil
4	Equipment ID & Barcode	123559 & 0530856
5	Date of purchase / Year of manufacture /Installation Date	29/1/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	17/07/2024 (Toll free)
9	Action taken	Checked and found out the problem .with compressor and motor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	Rs. 1800/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem .with compressor and motor. Machine installed in 29/01/2018 and covered up to 6+ years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Sreeju suresh 


*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Asset value - 1800
 cost of spare - Not submitted
 Life period - 6 years
 As per the BER guideline and spare not available
 recommended as BER


 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION
 Signature of JC BM (NHM)

Date


 Seal


 Signature of
 Superintendent / Medical Officer (Ue) - in- Charge

Community Health Centre
 Sachivothamapuram
 Kottayam Dist. Pin- 686 532

Ey-life™

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
HEALTH CARE PRIVATE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0530856**



S. P. S. Pillai

Medical Office
Community Health Centre
Sachintha
Kottayam Dist. 686 532

EZ LIFE Compressor Nebulizer.
Model: EzNeb-1303
Input: 220V/50HZ
SR No.:
Month of Import: Apr. 2017
Mode of operation: Continuous
30 Min. ON - 30 Min. OFF
Imported by:
Swami Ortho Aids





(414)68(254)05(21)3918

(E₃-life)™

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0530856



Handwritten signature
Medical Officer in Charge
Community Health Centre
Sreevallabapuram
Kottayam Dist. Ph: 586 532