



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234198

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : CHC, chunakkara	Call Registration Date : 13/07/2024
Address : Alappuzha Kerala	Caller ID : 122941
Ph : 8943935368	Date of Visit : 15/07/2024
	Asset No. : 0431223
	EQPT Name : Nebuliser
	Manufacture NuLife global Model : Mandynab
	S. No. NGM22-1939 Dept. Store

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working, Motor is not getting on

Action Taken : checked the machine and found that the compressor motor with piston is defective. Needs to replace these spare for further checking and working condition of the equipment

Completed Date : 15/07/24 Time : 2:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Maidula G	15/07/24	2:00pm	3:00pm

Customer Remark Completed Pending

Service Engineer Name : Maidula G
Signature : *Maidula G*
Date : 15/07/2024
Contact Number : 8590207135

Customer Name : BHAGYA ALEX HMI-S
Signature : *[Signature]*
Date : 15/07/2024
Contact Number : 8943935368
Designation : Nursing Officer
Hospital Seal :



[Handwritten Signature]

Dr. ANIL KUMAR S. MSBS, D.Ortho
Civil Surgeon Reg. No 24850
Medical Officer - Charge
Community Health Centre, Chunakkara
Charummal P.O., Alappuzha PIN-686001

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	CHC CHUNAKKARA
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :- NULIFE GLOBAL MEDICAL MODEL :- HANDYNEB SN:- NGM 227939
4	Equipment ID & Barcode	122941 & 0431223
5	Date of purchase / Year of manufacture /Installation Date	28/06/2014
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	13/07/2024
9	Action taken	CHECK THE MACHINE AND FOUND THAT THE COMPRESSOR MOTOR WITH PISTON IS DEFECTIVE.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDED FOR REPAIR
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS:- 1353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	CHECK THE MACHINE AND FOUND THAT THE COMPRESSOR MOTOR WITH PISTON IS DEFECTIVE. THIS MACHINE WAS INSTALLED ON 28/06/2014 AGED UP TO 10+ YEARS. QUOTATION IS NOT SUBMITTED SINCE SPARE IS NOT AVAILABLE IN THE MARKET. SO WE RECOMMENDING THE EQUIPMENT AS BER.
17	Name & Signature of CYRIX Authority	MRIDULA G <i>Mridula G</i>

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

Jyothish Thomas
21/08/24
JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM ALAPPUZHA

Signature of JC BM (NHM)



Date

16/8/24

Seal

Dr. ANIL KUMAR. S, MBBS, D Ortho
Civil Surgeon, Reg. No 24850
Medical Officer - In - Charge
Community Health Centre, Chunakkara
Chunakkara P.O., Alappuzha, PIN-690505

Anil Kumar S

Signature of
Superintendent / Medical Officer (i/c)



ALBULLZER

28/6/14	DOWNMOUNT Rtd from store [110/18 " 60"]	1	-	1
28/6/14	Jtd to IP	-	1	Nil
22/5/08	Rtd from store	1	-	1
22/5/08	Jtd to OP	-	1	Nil
31/03/2018	Rtd from store	1	-	1
04/04/18	Jtd to IP	-	1	Nil
29/4/2020	Recd from store	2	-	2
	(MEDI TECH HANDY ALD) (page no 37)		2	Nil
29/4/2020	Issued to OP		2	Nil



Handwritten signature in green ink.

Dr. ANIL KUMAR, S, MBBS, D.Ortho
 Civil Surgeon, Reg. No. 24850
 Medical Officer - In - Charge
 Community Health Centre, Chunnakkara
 Charummood P.O., Alappuzha, PIN-690505



Handwritten signature in green ink.

Dr. ANIL KUMAR, S, MBBS, D Ortho
Civil Surgeon, Reg No. 24850
Medical Officer - In - Charge
Community Health Centre, Chunakkara
Charummood P.O., Alappuzha, PIN-690505