

**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION**



**CYRIX**  
HEALTHCARE PVT LTD

No. : 230143

**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency  
30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

**Service Report**

Health Facility : CHC  
Address : Moothakunnam  
EKM  
Ph : 6282465935

Call Registration Date : 12/7/24  
Caller ID : 122716  
Date of Visit : 15/7/24  
Asset No. : 0721827  
EQPT Name : BP Apparatus  
Manufacture Life care Model : NA  
S. No. NA Dept. Observation Room

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Bp apparatus not getting ON.

Action Taken : Checked the equipment and found main board and charging socket defective. Need to replace these parts for further checking and working condition of the equipment.

Completed  Date : 15/07/24 Time : 1:30pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sebio Siby</u>	<u>15/7/24</u>	<u>12:00pm</u>	<u>1:30pm</u>


Customer Remark

Completed

Pending

Service Engineer Name : Sebio Siby  
Signature : [Signature]  
Date : 15/7/24  
Contact Number : 9746751735

Customer Name : Sure Alex  
Signature : [Signature]  
Date : 15/7/24  
Contact Number : 157769  
Designation : NO  
Hospital Seal : 6282465935






**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	ERNAKULAM
2	Name of Hospital	CHC MOOTHAKUNNAM ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Bp Apparatus Make : Life care Model : Na SN :Na
4	Equipment ID & Barcode	122716 & 0731827
5	Date of purchase / Year of manufacture /Installation Date	28-03-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	12-07-2024 (Toll free)
9	Action taken	Checked and found main board and charging socket defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	Rs. 1730
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found main board and charging socket defective. Machine installed on 28-03-2019 and covered up to 5+ years. Quotation not attached since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Sebin. 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

BER document not submitted in single page.  
 yearwise BER criteria met.  
 Quotation part Enquiry details not attached  
 Spares not available supported by Cyrix.

  
 9/8/24  
 Signature of JC BM (NHM)

Date

Seal

  
 മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്  
 സാമൂഹികാരോഗ്യം & പാരമ്പര്യം  
 Superintendent / Medical Officer (i/c)





Name of Article BP Apparatus (Mercury)


Date	Receipt kg / gm / ml	Issue kg / gm / ml	Balance kg / gm / ml	Prize on RD time	Designation & Sign of Receiver	Sign of Pharmacist	Sign of Hospital Superintendent
6.03.2010	3		3	Rs 195/-			
7.03.2013	1		4		H. J. J.		
20/12/2015	Sphygmomanometer, with 2 bracelets of two sizes - Child / Adult. (Round Model) (Rd date 8/9/2009) (Tr. in from NRHM KIT)						
	1		5				
19/2/18	Rtd from store (Supply taken from store) (Rs 200/-)						
			6		R		
5/10/18	M.O. a/c che M/cm (Supply taken from store) (Rs 200/-)						
12/9/18	Condensed & written off as per order NO 60/18 M.O a/c che M/cm dated 20/7/2018 [① 8/9/2009, ③ 6/2/2010] <u>Rs 195/-</u>						
			Balance → 2				
5/9/18	Rtd from the store (Rs 3080.00) [Electronic B.P Apparatus]						
			3				
28/3/19	Rtd from the store (Rs 1750/-)						
			2				
			Balance - 5		R. Rethakumar		
2/10/20	Tr out to Valsaruvayal Road (Rd 1/3/2013, 11/2/18, 5/9/18) (2)						
			3				
4/10/20	Condensed as per order no 328/2021 dt - 4/10/2021 3 NOS. Rd. 7/3/13 ① 19/2/18						
			Balance - 2				

Medical Officer in Charge  
Community Health Centre  
Moothakunnam

COMMUNITY HEALTH CENTRE  
MOOTHAKUNNAM  
Date.....  
5/9/18

Balance - 2  
cut 25/3/19





**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
HEALTH CARE PVT LTD



**TOLL FREE NO: 1800 - 425-7669**

**BAR CODE - (8004-890615225)**



**0731827**



മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്