



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 221339

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Phe</u>	Call Registration Date : <u>9/7/2024</u>
Address : <u>Peruvva</u>	Caller ID : <u>122040</u>
<u>Kottayam</u>	Date of Visit : <u>10/7/2024</u>
Ph : <u>9605259418</u>	Asset No. : <u>0543676</u>
	EQPT Name : <u>Autoclave</u>
	Manufacture : <u>NA</u> Model : <u>NA</u>
	S. No. : <u>NA</u> Dept. : <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked and found out the problem with top door, gasket, safety valve, steam valve and pressure gauge. Enquire the spare with OEM

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Na</u>	<u>Na</u>	<u>Na</u>	<u>Na</u>
2. <u>Na</u>	<u>Na</u>	<u>Na</u>	<u>Na</u>
3. <u>Na</u>	<u>Na</u>	<u>Na</u>	<u>Na</u>

Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>10/7/2024</u>	<u>1:00pm</u>	<u>2:00pm</u>

Customer Remark : Not working Deva Mathew Deva
Completed Pending

Service Engineer Name : AMAL BABU Customer Name : Alcent
Signature : [Signature] Signature : Dr Meenu Induchoodan
Date : 10/7/2024 Date : 10/7/2024
Contact Number : 956757598 Contact Number : contacted certified manufacturer
Hospital Seal : [Seal] Hospital Seal : found agency name, email




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC PERUVA
3	Name of Equipment with Make, Model and Serial Number	Equipment : Sterilizer Make :NIL Model :NIL SN:NIL
4	Equipment ID & Barcode	122040 & 0543676✓
5	Date of purchase / Year of manufacture /Installation Date	15/12/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	09-07-2024 (Toll free)
9	Action taken	Checked and found out the problem with top door ,gasket ,safety valve, steam valve and pressure gauge
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	gasket - 649/- pressure gauge -767/- safety valve - 649/- steam release valve - 413/- top door - 3835/- TOTAL - 6313/-

13	Asset Value	Rs .9435 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	67%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and quotation are attached
16	Reasons for recommending the equipment as BER	Checked and find out the problem with top door ,gasket ,safety valve, steam valve and pressure gauge Machine installed in 15/12/2017 and covered up to 6+ years. The repairing cost 67%. Both criteria met So recommending the machine for condemnation as per the tender clause 5.3.14.1
17	Name & Signature of CYRIX Authority	Amal babu 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Asset value . 9435 -
 Cost of spae = 6313 /- (67%)
 Life period - - 6 years 5 months
 As per the RBER guidelines
 recommended as RBER

CONSULTANT BIOMEDICAL
 NATIONAL HEALTH MIS
 KOTTAYAM


 Signature of JC BM (NHM)

Date




 Signature of
 Medical Officer (i/c)
 ഹെൽത്ത് (മെഡി) ഓഫീസ്: 686 615

AUTOCLAVE ALUMINIUM 512 12 X 12

15/12/2017

Purchased from J.K Mehta & Sons, Surgical & Scientific Dealers, Manorama Jn, Kottayam.

Invoice No : A1767, dated 15/Dec/2017.

HSN : 8419

Rate : 7995.76

Total : 9435.00

1 No.

1 No.

15/12/2017

Issue to s/n

1 No.

Nil

Recd
14.5.18
14/5/18



മെഡിക്കൽ ഓഫീസർ
പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം, പെരൂർ
കോട്ടയം ജില്ല
മെഡിക്കൽ ഓഫീസർ
പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം, പെരൂർ
കോട്ടയം ജില്ല
ഫോൺ നമ്പർ: 686 615



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO: : UDYAM-KL-07--0030714
DL NO: : KL/KTM/MD42/2023/000015

Estimate

Estimate For:

ARUN G C

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 8921992272

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-39

Date: 22/07/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	Amount
1	GASKET -		1	Nos	₹ 649.00	₹ 649.00
2	PRESSURE GAUGE	90262000	1	Nos	₹ 767.00	₹ 767.00
3	SAFETY VALVE		1	Nos	₹ 649.00	₹ 649.00
4	STEAM RELEASE VALVE	8481	1	Nos	₹ 413.00	₹ 413.00
5	TOP DOOR	8419	1	-	₹ 3,835.00	₹ 3,835.00
Total			5			₹ 6,313.00

Sub Total	₹ 6,313.00
Total	₹ 6,313.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH

Description

PHC PERUVA
EQUIPMENT: STERILIZER
MAKE : NIL
MODEL: NIL
SN: NIL

Estimate Amount In Words

Six Thousand Three Hundred Thirteen Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



മുൻ

അഡ്വക്കേറ്റ് ഓഫീസ് പ്രൊഫർട്ടർ
പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം, പെരുമ്പ
കോട്ടയം (Dist) പിൻ: 686 617

