



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 232708

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Thiruvananthapuram</u> <u>Kerala</u></p> <p>Ph : <u>9495469434</u></p>	<p>Call Registration Date : <u>6/7/24</u></p> <p>Caller ID : <u>121473</u></p> <p>Date of Visit : <u>5/8/24</u></p> <p>Asset No. : <u>0110298</u></p> <p>EQPT Name : <u>Ultrasound therapy Unit</u></p> <p>Manufacture <u>Steyker</u> Model : <u>Biotech</u></p> <p>S. No. : <u>NA</u> Dept. : <u>PMR</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** Checked and found that the high voltage cord and 1MHz ultrasound probe defective. Need to replace the spares for further checking and working condition of the equipment.

Completed  Date : 5/8/24 Time : 1:00 PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Soorya Sudheep</u>	<u>5/8/24</u>	<u>12:50 PM</u>	<u>1:00 PM</u>

**Customer Remark** Pending



**Service Engineer Name :** Soorya Sudheep  
**Signature :** [Signature]  
**Date :** 5/8/24  
**Contact Number :** 7907406244

**Customer Name :** \_\_\_\_\_  
**Signature :** \_\_\_\_\_  
**Date :** \_\_\_\_\_  
**Contact Number :** \_\_\_\_\_  
**Designation :** \_\_\_\_\_  
**Hospital Seal :** \_\_\_\_\_


Deepu B  
Physiotherapist  
General Hospital  
Thiruvananthapuram  
7907403923

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Ultrasound Therapy Unit Make : Stryker Model : Biotech SL: NA
4	Equipment ID/ Barcode	121473/0110298
5	Date of purchase/ Year of manufacture/Installation Date	19/09/2018
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	06/07/2024
9	Action taken	Checked and found low voltage PCB and 1Mhz Applicator defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spares from OEM.
10	Present status of the equipment(Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	1Mhz Applicator 5cm.sq with crystal - 2968/- Low voltage PCB - 7670/- Freight cost - 236/- TOTAL = 10874/-

13	Asset Value	14600/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	74.4%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND VENDOR QUOTATION ATTACHED
16	Reasons for recommending the equipments as BER	Checked and found low voltage PCB and 1Mhz applicator defective. The equipment was installed on 19/09/2018 and covered upto 5+ years. The repairing cost is 74.4%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unknit for condemnation.
17	Name & Signature of CYRIX Authority	SOORIYA SUDHEEP 


\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Low voltage PCB and 1MHz applicator defective. Estimate for repair cost verified. So can be RBER as per guidelines

**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
2.8.2024.  
Signature of JC BM (NHM)

Date:



  
Signature of Superintendent/Medical Officer (i/c)  
Superintendent  
General Hospital  
Thiruvananthapuram

# U.S.T (Ultrasound therapy)

DATE	Name and address of supplier with supply order No	Receipt	Balance	Warranty	Remarks
(E)	Technomads Electronics plot No. 41, funthomab Industrial Estate, perungudi, chennai 600096. SL No. TUI-8047 model S-14 (Electroson - 608)	1	1		✓
7/09/15	Bhanadha medical systems. KMCL Invoice No. 162 dt. 25/08/2015. Supply order No. 460/2015/130 17/08/15. P. No. - 274 Vol. - 8A Model - Digi Sound dual (Rs. 22000 + Tax 1100 - 23100) MSR page No. 279 Vol. 8A) Bhadra - 9995052532 Balasubramanyam 9995052532. SL No. (8338)			7/9/15 do 6/09/17	do do
(F)	19/09/18 Biotec sinter. kmcl order no. 101/323/2017-18/130 dt. 11/01/2018 ultra processor controlled ultrasonic therapy unit serial no. 999566839 SL No. 04539	1	3	11/09/18 do 18/09/18	do do do

### Estimate

#### Estimate For

Cyrix Healthcare Pvt Ltd

FIRST FLOOR, 30/641B

PETTA JUNCTION,

POONITHURA, ERNAKULAM-682038,

KERALA INDIA

GSTIN : 32AAFCC2499H2ZM

State: 32-Kerala

#### Estimate Details

Estimate No. : MT/QT/24-25/110

Date : 16-07-2024

Place of supply: 32-Kerala

#	Item name	Item Code	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	1Mhz Applicator 5cm.sq. with crystal	05100	90189099	1	Nos	₹ 2,650.00	₹ 318.00 (12%)	₹ 2,968.00
2	Services & Repairing Charges (Low Voltage pcb replacement for UST-0112287)	SRV-REP	995469	1	Nos	₹ 6,500.00	₹ 1,170.00 (18%)	₹ 7,670.00
3	freight			1	-	₹ 200.00	₹ 36.00 (18%)	₹ 236.00
	<b>Total</b>			<b>2</b>			<b>₹ 1,524.00</b>	<b>₹ 10,874.00</b>

Tax-type	Taxable amount	Rate	Tax amount	Amounts
SGST	₹ 2,650.00	6%	₹ 159.00	Sub Total
CGST	₹ 2,650.00	6%	₹ 159.00	
SGST	₹ 6,700.00	9%	₹ 603.00	<b>Total</b>
CGST	₹ 6,700.00	9%	₹ 603.00	<b>₹ 10,874.00</b>

### Estimate Amount In Words

Ten Thousand Eight Hundred Seventy Four Rupees only

### Terms and Conditions

Terms and conditions:

Payment Terms: 100% Advance

The Guarantee/Warranty would not cover: -

- No Warranty on Rubber, Plastic & Wooden Products / Parts.
- Any Mechanical damage/breakage.
- Any damages to vacuum Tube / Valve (if any), Meters.
- Any damages due to electrical fluctuation.
- Supply of consumables and accessories.
- Any Physical damage to the Device due to Intentional act or wilful neglect.
- Any Equipment's Serial number that has been altered, defaced or removed, or has been modified to alter its functionality or capability without the written permission of Hamb Meditech LLP.

### Bank Details



Name : STATE BANK OF INDIA

Account No. : 41310832759

IFSC code : SBIN0003106

Account holder's name : Meibot Technologies

For : MEIBOT TECHNOLOGIES

Authorized Signatory





---

## Quotation for the service of UST Machine

3 messages

---

**MURALIKRISHNA S** <muralikrishnaalpy@gmail.com>  
To: kasyepcyrix@gmail.com

Wed, 17 Jul 2024 at 6:54 pm

Dear Sir/ Madam,

Kindly find the attached quotation.

Thanks and Regards

Murali Krishna  
Sales & Service Engineer

Whatspp: <https://wa.me/919074126766>

---

**MURALIKRISHNA S** <muralikrishnaalpy@gmail.com>  
To: kasyepcyrix@gmail.com

Sat, 27 Jul 2024 at 10:12 am

Dear Sir / Madam,

Kindly find the attached quote.

-

Thanks and Regards

**Muralikrishna Suresh**  
**Sales & Service Manager**

Whatspp: <https://wa.me/919074126766>

[Quoted text hidden]

---

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: MURALIKRISHNA S <muralikrishnaalpy@gmail.com>

Mon, 29 Jul 2024 at 1:56 pm

Sir,  
Recieved the estimate, will do the payment on 30/7/2024

Thanks & Regards,

KASYEP PV  
District Incharge ,Trivandrum  
CYRIX HEALTHCARE Pvt.Ltd  
KLBEMP  
Ph : +91 7593847134  
kasyepcyrix@gmail.com

[Quoted text hidden]

**CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE**  
PROGRAMME (BEMF)



TEL: 011 4310-415-7669

BAR CODE: 0110298





Micro-Processor Controlled  
Ultrasound Therapy Unit



SW1



SW2

Biotech

Striker  
get's you going

