



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 220254

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|--------------------------------------|--|
| Health Facility <u>GH Pala</u> | Call Registration Date : <u>4/7/2024</u> |
| Address <u>Kottayam</u> | Caller ID : <u>121021</u> |
| Ph : <u>9605841191</u> | Date of Visit : <u>6/7/2024</u> |
| | Asset No. : <u>0511205</u> |
| | EQPT Name : <u>OT Light</u> |
| | Manufacture <u>Hospitech</u> Model : <u>Hospilite 1004</u> |
| | S. No. <u>Nil</u> Dept. <u>OT</u> |

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked the machine Bulb, Bulb holder, Transformer and Bulb reflector are fully damaged and handles/Arms fully rusted not working properly.

Completed Date : 6/7/2024 Time : 1pm Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |



ABHINAV
01/07/24
SUPERINTENDENT
Govt. General Hospital
Pala, Kottayam.

| | | | |
|-------------------------|-----------------|-------------|------------|
| Cyrix Engineer | Date | Start Time | End Time |
| <u>Josekutty George</u> | <u>6/7/2024</u> | <u>12pm</u> | <u>1pm</u> |

Customer Remark Completed Pending

Not Working

Service Engineer Name : Josekutty George
Signature : [Signature]
Date : 6/7/2024
Contact Number : 8848230407


Customer Name : Balukutty KR
Signature : [Signature]
Date : 6/7/2024
Contact Number : 9605841191
Designation : [Signature]
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|---|---|
| 1 | Name of District | KOTTAYAM |
| 2 | Name of Hospital | GH PALA |
| 3 | Name of Equipment with Make, Model and Serial Number | Equipment :OT LIGHT Make: HOSPITECH Model :1007 SN:NIL |
| 4 | Equipment ID & Barcode | 121021 & 0511205 |
| 5 | Date of purchase / Year of manufacture /Installation Date | 13/9/2005 |
| 6 | Warranty details (Yes/No) | No warranty |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | No AMC/CMC |
| 8 | Date of breakdown(Date of registration of complaint through email/ Toll free) | 04/07/2024 (Toll free) |
| 9 | Action taken | Checked and found out the problem with transformer, lamp, lamp holder and handles arm need to be repaint because of rust in condition |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged |
| 11 | Recommendations for repair (required service details) | Not recommending for repair. |
| 12 | Cost of spares (specify parts and cost) | N/A |


| | | |
|----|--|---|
| 13 | Asset Value | Rs . 46000/- |
| 14 | # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | N/A |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Cyrix service report are attached |
| 16 | Reasons for recommending the equipment as BER | Checked and found out the problem with transformer, lamp, lamp holder and handles arm need to be <u>repaint</u> because of rust in condition Machine installed in 13/09/2005 and covered up to 18 + years. So recommending the machine for condemnation |
| 17 | Name & Signature of CYRIX Authority | Josekutty George  |

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 46000, Life period - 18 years
 Cost of spare = Not mentioned = "transformer
 Lamp, Lamp holder and handles, Bulb reflector are
 damaged." The machine is 18 years older


 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH AUTHORITY
 KOTTAYAM
 Signature of JC BM (NHM)

Date 26/7/24

JANUARY 26/7/24
 Signature of
 Superintendent / Medical Officer (i/c)



SHADOWLESS LAMP

050

| dt of RT | dt of work | Qty | Worky | Net worky cost - copy | Name of mtr |
|----------|------------|-----|-------|-----------------------|-------------------------|
| 10-9-05 | 13-9-05 | 1 | ✓ | 46000/- | Diff's supply |
| 3-2-06 | 3-2-06 | 4 | ✓ | 75460+125/- | P.P. Vol X Page 57 |
| 3-2-06 | 3-2-06 | 1 | ✓ | 16836+125/- | L.P. Trl (P.P. 57) X |



PORTABLE + ROOF MOUNTED

051


| Model | Quantity | A.m.c | Issued to Dept | S.I.S Page no | Material no |
|---------------|----------|-------|----------------|---------------|-------------|
| Hospital 1007 | - | - | Issued to OT | 122 | 247/IV |
| - | - | - | Issued to OT | 122 | 247/IV |
| - | - | - | Issued to OT | 122 | 252/IV |

~~REQUISIT~~
P.P. 57

SUPERINTENDENT
Govt. General Hospital
Palakkad, Kerala



BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMIP)
FREE NO: 1800 - 425-7669
CODE - (8004-890615225)
0511205



SUPERINTENDENT
 General Hospital
 Peta, Kottayam.

GENERAL HOSPITAL
 PETA, KOTTAYAM