


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	CHC KODUVAYUR
3	Name of Equipment with Make, Model and Serial Number	NEBULISER Make : LIFE CARE Model: CNB69009 SN: LIFE092009230940
4	Equipment ID/ Barcode	0930066 # 127212
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date = Nil (actual installation date as per stock book 13-08-2018 ,stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	03-08-2024 (toll free)
9	Action taken	Checked the machine found not switching on. Found nebuliser motor burned , piston broken. Need to replace nebuliser motor with piston for further checking of machine
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS. 1353 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the equipment found not switching on. Found nebuliser motor burned and piston broken. As per stock book equipment installed on 13-08-2018 and aged up to 5 years 10 months. Quotation not submitted since Spare parts are not available on market. So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory

*Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Equipment- installed on 13.8.2018 and aged up to 5 years. Quotation not submitted.


09/08/24
Signature of Junior Consultant (Biomedical Engg)
NHM (Programme Officer), Palakkad


Date:




Signature of
Superintendent/Medical Officer (Uc)
Superintendent
Community Health Centre
Koduvayur

MEMORIALIZER.

B/F	From	To	Page no	R	I	Blnc
11-12-2008	Received from	Store (L)		1		(Receipt)
14/12/08	Issued to	Op			1	
14/3/18	Received from	Store (Kmscl)		2		(Receipt)
15/3/18	Issued to	Op			2	
14/8/18	Received from	Store	Rs. 140/1000	2		Cash
14/8/18	Issued to	Op			2	
15/10/18	Rel from	Store		1		for cash
24/1/2010	Rel from	Store		1		
(IX 1560)						
op-2 Ip				2	Rep 2.	
21/1/2021	Rel from	Store	Kmscl	1		
vide case Rs. 1232.						
2/9/21	Rel from	store	(Kmscl Supply)	2		2
3/9/21	Issued	to	Op		1	1
13/9/21	Issued	to	Ip		1	N.L
26/12/21	Rel from	store	(Kmscl - lifecare)	4		4
9/10/23	Cancellation done as per order no: H3/475/2023					10.
	9/10/2023.				1	13


 Superintendent
 Community Health Centre
 Koduvayur





Jay
Superintendent
Community Health Centre
Koduvayur