


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**


Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	CHC NEMMARA
3	Name of Equipment with Make, Model and Serial Number	NEBULISER Make : TRUCHEK Model: NA SN: NA
4	Equipment ID/ Barcode	0950035 # 124653
5	Date of purchase/ Year of manufacture/Installation Date	09/06/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	22/07/2024 (toll free)
9	Action taken	Checked the machine found not switching on. Found nebuliser motor burned , piston broken. Need to replace nebuliser motor with piston for further checking of machine
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS. 1353 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Assel Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipments as BER	Checked the equipment found not switching on. Found nebuliser motor burned and piston broken. Equipment installed on 09/06/2018 and aged up to 6 years 1 month. Quotation not submitted since Spare parts are not available on market. So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 


\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:  
 Nebuliser installed on 9/6/2018, aged up to 6 years. Quotation not submitted.

  
 Signature of Junior Consultant (Biomedical Engineering) NHM (Aranyakal) Palakkad  
 6/08/24

Date: \_\_\_\_\_

  
 Signature of Superintendent Medical Officer (M)



SUPERINTENDENT  
 Community Health Centre  
 Nemmara, Palakkad.





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. :

231631

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.km@cyrix.in

Health Facility ... <u>CHC NEMMARA</u>	Call Registration Date : <u>22/07/2024</u>
Address ... <u>NEMMARA</u>	Caller ID : <u>124653</u>
... <u>PALAKKAD</u>	Date of Visit : <u>24/07/2024</u>
Ph : <u>9446150288</u>	Asset No. : <u>0950035</u>
	EQPT Name : <u>NEBULIZER</u>
	Manufacture <u>TRU CHEK</u> Model : <u>NIL</u>
	S. No. <u>NIL</u> Dept. <u>MM WARD</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Machine not switching ON.

Action Taken : Checked the machine. Found that machine switching ON. Found nebulizer motor burned and piston broken. Need to replace nebulizer motor with piston for further checking of machine

Completed  Date : 24/07/24 Time : 10:00 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>DIVIN'S</u>	<u>24/07/24</u>	<u>9:30 AM</u>	<u>10:00 AM</u>

Customer Remark Completed  Pending

*Handwritten signature and notes:*  
 Pending  
 for Supervisor  
 SUPERINTENDENT  
 Community Health Center  
 Nemmara



Service Engineer Name : DIVIN'S  
 Signature : [Signature]  
 Date : 24/07/24  
 Contact Number : 7592847155

Customer Name : Susitha Varghese  
 Signature : [Signature]  
 Date : 24/7/24  
 Contact Number : 9446150288  
 Designation : S/O  
 Hospital Seal :




# NEBULISER.

		Rd	Used in ward	Bal. In Sub acc.	Balanc in main stock
	B.F Old book page No.				
9.6.18 (11/7/2014)	Rd from store (L.P. page 35) 18301	2	2	-	2
10.3.2020	Rd from store Air compressing nebuliser from Kanungu Medical Depot Kalam Rd 1562.65	1	1	nil.	3
10/2/2021	Dr in house & c Nemmara P.M. 57	2			5
5.1.22	Rd from store	2			7
6.1.22	Issued to ward		7	nil.	7
23.4.23	Tr-out to Unservice page No. 29		2	(nil)	5
10/12/23	Tr-out to unservice page No. 29		2	(nil)	3
23/1/24	Rd from the store (LP)	2			5
24/1/24	Went to ward for use		5	nil.	5
	Rate Rs. (only, for. man) (1686.72 x 2 = 3373.44/-)				



*Handwritten signature*





Sunitha Venghese  
Sunitha  
SND