



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :

221332

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>chc</u></p> <p>Address <u>edayathan</u> <u>Kattayam</u></p> <p>Ph : <u>9447558757</u></p>	<p>Call Registration Date : <u>24/6/2024</u></p> <p>Caller ID : <u>119048</u></p> <p>Date of Visit : <u>26/6/2024</u></p> <p>Asset No. : <u>0530416</u></p> <p>EQPT Name : <u>Refrigerator</u></p> <p>Manufacture <u>Godrej</u> Model : <u>EQM</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
No Cooling

Action Taken :
checked and found out the problem with compressor, gas leak. Thermostat. Inquired the spare with OEM

Completed Date : Time : Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. /	/	/	/
2. /	/	/	/
3. /	/	/	/

Cyrix Engineer	Date	Start Time	End Time
<u>ANIL BABU</u>	<u>26/6/2024</u>	<u>11:00am</u>	<u>11:30am</u>

Customer Remark Completed Pending
Equipment not working

Service Engineer Name : ANIL BABU
Signature :
Date : 26/6/2024
Contact Number : 956757598




Customer Name :
Signature : [Signature]
Date :
Contact Number :
Designation :
Hospital Seal :
മെഡിക്കൽ ഓഫീസർ (ഇൻ ചാർജ്ജ്)
സാമൂഹികാരോഗ്യ കേന്ദ്രം
ഇടയാഴം, കുടവെച്ചൂർ-626 144

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	CHC EDAYAZHAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Pharmacy Refrigerator Make : Godrej Model :Eon SN:Nil
4	Equipment ID & Barcode	119048 & 0530416 ✓
5	Date of purchase / Year of manufacture /Installation Date	20/10/2000 ✓
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24-06-2024 (TOLL FREE)
9	Action taken	Checked and found out the problem with compressor, gas leak, Thermostat. Enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Compressor -4000/- Gas charge -3500/- Thermostat 950 /- Total-8450/-

13	Asset Value	Rs . 12520/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	67%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM Quotation attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with compressor, gas leak and Thermostat. Machine installed in 20/10/2000 and covered up to 23 +.The repairing cost 8450/- both criteria met. So recommending the machine for condemnation as per the tender clause 5.3.14.1
17	Name & Signature of CYRIX Authority	Amal babu 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Assesd value - 12520
 cost of spare - Rs 8450 (67%)
 life period - 23 years
 As per the RBER guidelines recommended as RBER

CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION
 KOTTAYAM


 Signature of JC BM (NHM)

Date _____

Seal _____


 Signature of Superintendent / Medical Officer (i/c)

വെഡിംഗ് ഓഫീസർ (ഇൻ ചാർജ്ജ്)
 സാമൂഹികാരോഗ്യ കേന്ദ്രം
 ഇടയാഴം, കൂടവെച്ചൂർ - 626 144



REFRIGERATOR [FRIDGE]

	Qty	Issued Balance
20/10/19 From the Store	1	1
Issued to the Nurses staff Rs. 12500/-		1
BIF to Stock Reg. Vol II P.No. 107		1



[Handwritten Signature]
 റെഡിക്കൽ ഓഫീസർ (ഇൻ ചാർജ്ജ്)
 സാമൂഹികാരോഗ്യ കേന്ദ്രം
 ഇടയാഴം, കൂടവെച്ചൂർ - 686 144



ചെറുപുഴ ഓഫീസർ (ഇൻ ചാർജ്ജ്)
 സാമൂഹികാഭരണ കേന്ദ്രം
 ഇടയാഴം, കുടവെച്ചൂർ - 626 144