



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 229455

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

Health Facility <u>G.H. Kanjirappally</u>	Call Registration Date : <u>24-06-24</u>
Address <u>Kolthayam</u>	Caller ID : <u>118963</u>
Ph : <u>9497730055</u>	Date of Visit : <u>26-06-24</u>
	Asset No. : <u>0514088</u>
	EQPT Name : <u>Suction Apparatus</u>
	Manufacture <u>Monarch</u> Model : <u>MS1A602</u>
	S. No. <u>2020275</u> Dept. <u>Payward</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Not working.

Action Taken : checked and found that problem with suction motor, suction jar, regulator. Enquire the spare with OEM.

Completed Date : 26/6/24 Time : 1:00 P.M. Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Ajmal Shajahan</u>	<u>26/6/24</u>	<u>10:00</u>	<u>1:00</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Ajmal Shajahan</u> Signature : <u>[Signature]</u> Date : <u>26/06/24</u> Contact Number : <u>9497275612</u>	Customer Name : <u>Vamunakrishnan</u> Signature : <u>[Signature]</u> Date : <u>26/06/24</u> Contact Number : <u>7558047701</u> Designation : <u>STABB NCOE</u> Hospital Seal : <u>[Seal]</u>
-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------




[Signature]
Superintendent
General Hospital
Kanjirappally

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	GH KANJIRAPALLY
3	Name of Equipment with Make, Model and Serial Number	Equipment : Suction Apparatus Make : Monarch Model :MSI-RG02 SN: 2020275
4	Equipment ID & Barcode	118963 & 0514088 ✓
5	Date of purchase / Year of manufacture /Installation Date	21/10/2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registrationof complaint through email/ Toll free)	24/06/2024 ✓(Toll free)
9	Action taken	Checked and found out the problem with suction motor, jar and pressure gauge enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Suction motor - 4658/- Suction jar – 3155/- Pressure gauge- 1125/- TOTAL- 8938/-


13	Asset Value	Rs . 11682/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	76.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and quotation are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with suction motor, jar and pressure gauge . Machine installed in 21/10/2020 and covered up to 3 + years. The repairing cost 76.5 % Both criteria met. So recommended the machine for condemnation as per the tender close 5.3.14.1
17	Name & Signature of CYRIX Authority	Ajmal shajahan 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Asset value - 11682 / -
 cost of spares = 76.5 % (Rs 8938.)
 Life period - 3 year 10 months -
 As per the RBER guidelines Life period criteria not Satisfical


 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH SERVICES
 KANJIRAPPALLY
 Signature of JC BM (NHM)



Date 26/7/24

Seal


 Superintendent
 General Hospital
 Kanjirappally
 Superintendent / Medical Officer (I/c)

Section Appointments

- 77 -

Rd from store (Covid fund) 1

Serial no: 2020/75.

Warrantly 21/10/2020 - 20/10/2021
Project name (Covid-19)

Make - Marathi Surgicals.

Model No. - 02.

Rm. 9900 + Tax.

Dept. No. Dental. OPO (10)

(How kept in 2017 OPO, 1/18)

Not working

For Incharge.

Nto Sunetha



[Signature]
Socernya - Em.
Nto
**Superintendent
General Hospital
Kanjirappally**



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO: : UDYAM-KL-07--0030714

Estimate

Estimate For:

ARUN G C

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 8921992272

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-35

Date: 10/07/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,947.46	₹ 710.54 (18%)	₹ 4,658.00
2	SUCTION JAR	90189099	2	Nos	₹ 1,408.48	₹ 338.04 (12%)	₹ 3,155.00
3	PRESSURE GAUAGE		1	Nos	₹ 1,125.00	₹ 0.00 (0%)	₹ 1,125.00
	Total		4			₹ 1,048.58	₹ 8,938.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 7,889.42
SGST@6%	₹ 169.02
CGST@6%	₹ 169.02
SGST@9%	₹ 355.27
CGST@9%	₹ 355.27
Total	₹ 8,938.00

Description

GH KANJIRAPPALLY
BAR CODE- 0514088
ID- 118963

Estimate Amount In Words

Eight Thousand Nine Hundred Thirty Eight Rupees only



Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



Not working

Sarath
N/A

For Incharge
Suresha.



Sarath
Superintendent
General Hospital
Kanjirappally