


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	THQH PERUMBAVOOR ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Sterlizer Make : Paras Model :Na SN Na
4	Equipment ID & Barcode	118347 & 0721842
5	Date of purchase / Year of manufacture /Installation Date	25-05-2006
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20-06-2024 (Toll free)
9	Action taken	Checked and found coil burned and leaking from bottom of the body and drain valve side.And the joints are rusted and machine is in not repairable condition.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 3600 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found coil burned and leaking from bottom of the body and drain valve side. And the joints rusted and machine is in not repairable condition .Machine installed on 25-05-2006 and covered up to 18+ years and very old machine. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	YedhuKrishna. 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

machine is completely damaged reported
coil issue, leaking, rusted
more than 18 years old unit. (25/5/2006)
Submitted BER on 23/7/24.

*Deputy
Consultant BME*
24/7/24
Signature of JC BM (NHM)



SUPERINTENDENT
Taluk Head Quarters Hospital
Perumbavoor
Ernakulam Dist

16/08/24
Signature of
Superintendent / Medical Officer (i/c)

Date

Seal



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 230822

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>G.H.Q.H</u></p> <p>Address <u>G.H.Q.H Perumbavoor</u></p> <p style="padding-left: 40px;"><u>Ernakulam, Kerala</u></p> <p>Ph : <u>9549791835</u></p>	<p>Call Registration Date : <u>20.06.2024</u></p> <p>Caller ID : <u>118347</u></p> <p>Date of Visit : <u>26.06.2024</u></p> <p>Asset No. : <u>0721842</u></p> <p>EQPT Name : <u>sterilizer</u></p> <p>Manufacture <u>Paras</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>ward 7</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

Not working and water leakage.

Action Taken : checked and found coil burned and
leaking from bottom of the body and drain valve
side. And the joint are rusted and machine
is in not repairable condition.


Completed Date : 26.06.24 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. _____	—	_____	_____
2. _____	—	_____	_____
3. _____	—	_____	_____

Cyrix Engineer	Date	Start Time	End Time
<u>Yedhukorshna Cy</u>	<u>26/6/24</u>	<u>11:40am</u>	<u>12:00pm</u>


Customer Remark Completed Pending



SUPERINTENDENT
Taluk Head Quarters Hospital

<p>Service Engineer Name : <u>Yedhukorshna Cy</u></p> <p>Signature : _____</p> <p>Date : <u>26/6/2024</u></p> <p>Contact Number : <u>8086147256</u></p>	<p>Customer Name : <u>Soya. K. P</u></p> <p>Signature : _____</p> <p>Date : <u>26/6/24</u></p> <p>Contact Number : <u>9495559456</u></p> <p>Designation : <u>Nursing Officer</u></p> <p>Hospital Seal : _____</p>
---	---

16/6/24



16/6/2014
SUPERINTENDENT
Taluk Head Quarters Hospital
Perumbavoor
Ernakulam Dist