



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 230314

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>THQH</u></p> <p>Address : <u>THQH KUNDARA</u> <u>KOLLAM</u></p> <p>Ph : <u>9497620947</u></p>	<p>Call Registration Date : <u>20-06-2024</u></p> <p>Caller ID : <u>118197</u></p> <p>Date of Visit : <u>21-06-2024</u></p> <p>Asset No. : <u>0221178</u></p> <p>EQPT Name : <u>NST MACHINE</u></p> <p>Manufacture <u>BPL</u> Model : <u>FM9852</u></p> <p>S. No. <u>EPTA4H1253</u> Dept. <u>LABOUR ROOM</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not getting ON.

Action Taken : Checked the NST Machine found that machine not getting on, its Display and main board defective Enquire spare with OEM.

Completed Date : 21-06-2024 Time : 12:30pm Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number

Cyrix Engineer	Date	Start Time	End Time
<u>Wilfert Sunil William</u>	<u>21-06-2024</u>	<u>11:30am</u>	<u>12:30pm</u>

Customer Remark Completed Pending

Service Engineer Name : Wilfert Sunil William
Signature : [Signature]
Date : 21-06-2024
Contact Number : 9778526652

Customer Name : J.R.
Signature : [Signature]
Date : 21-06-2024
Contact Number : 6065576
Designation : EE nurse GAI
Hospital Seal : [Seal]

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOLLAM
2	Name of Hospital	THQH KUNDARA
3	Name of Equipment with Make, Model and Serial Number	Equipment : NST MACHINE Make: BPL Model: FM9852 SN : EPTA4H1253
4	Equipment ID & Barcode	Ticket id: 118197 / Barcode : 0221178
5	Date of purchase / Year of manufacture / Installation Date	15-09-2015
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	20-06-2024
9	Action taken	Regarding inspection of the equipment Display & Mother board are defective. Enquired spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	FM9852 Mother Board = 47600 FM Display LCD sharp FM9852 = 10640 Total = 58240/-

13	Asset Value	Rs . 75000/-
14	# Percentage value of the cost of spareswith respect to Cost of Purchase/ Asset Value	77.65%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report ,Quotation attached
16	Reasons for recommending the equipment as BER	Checked and found Diplay & Mother Board are defective, Equipment is installed on 15-09-2015 and aged up to 8 + years. Repairing cost of equipment is 77.65%. As per tender clause 5.3.14.1, both criteria met. so recommending to condemnation.
17	Name & Signature of CYRIX Authority	Wilfret sunil

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Consultant (Biomedical) NHM:

As per the quotation received from Biomedix solutions and as per the criteria. Recommended for Beyond economic repair.

[Signature]
23/07/24
Signature of ABMO (NHM)
NHM, KOLLAM.



Date

Seal

Signature of
Superintendent / Medical Officer (I/c)

BIOMEDIX SOLUTIONS

VII/437B KATHALANGAL PADI CHIRAKKADAVU

Phone no.: 8289991634,

Email: biomedixsales@gmail.com,biomedixservice@gmail



Estimate

Estimate For

CYRIX HEALTH CARE PVT LTD

FIRST FLOOR,PETTA JUNCTION 30/641B

POONITHURA,ERNAKULAM-682038

MAILID:Sales.KI2@cyrix.in Ph:9744555073

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate Details

Estimate No.: 2203202444

Date: 20-07-2024

Time: 04:15 PM

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Price/ Unit	GST	Amount
1	FM9852 Mother Board (single)		1	₹ 42,500.00	₹ 5,100.00 (12.0%)	₹ 47,600.00
2	FM display LCD sharp FM 9852		1	₹ 9,500.00	₹ 1,140.00 (12.0%)	₹ 10,640.00
Total			2		₹ 6,240.00	₹ 58,240.00

Estimate Amount In Words

Fifty Eight Thousand Two Hundred and Forty Rupees only

Terms And Conditions

Thanks for doing business with us!

Sub Total	₹ 52,000.00
CGST@6.0%	₹ 3,120.00
SGST@6.0%	₹ 3,120.00
Total	₹ 58,240.00

For: BIOMEDIX SOLUTIONS

Authorized Signatory



192 FOETAL MONITOR

Receipt issue Balas

15.9.15 Receipt from Stock.
Colour T. FT
[Main Stock Page No. 146
BPL Medical Technology
Rit 5,000/- 3 year Warranty]

1

5.

1.10.19 TR. out to Labour Room
Page No. 183

1

NR

PA
1.



