



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 223698

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>D.H. Nedumangad</u></p> <p>Address <u>Trivandrum</u></p> <p>Ph : <u>8129509201</u></p>	<p>Call Registration Date : <u>04/06/2024</u></p> <p>Caller ID : <u>114956</u></p> <p>Date of Visit : <u>10/06/2024</u></p> <p>Asset No. : <u>0112464</u></p> <p>EQPT Name : <u>Suction Apparatus</u></p> <p>Manufacture <u>Supreme plus</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Casualty</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** checked and found Suction jar and motor defective need to replaced the spares for further checking and working condition of the equipment.

Completed  Date : 10/06/2024 Time : 11:00 AM Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>midhun - BM</u>	<u>10/06/24</u>	<u>9:50 AM</u>	<u>11:00 AM</u>

**Customer Remark** Completed  Pending

<p>Service Engineer Name : <u>midhun</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>10/06/24</u></p> <p>Contact Number : <u>9074167987</u></p>	<p>Customer Name : <u>Achila R-S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>10/06/24</u></p> <p>Contact Number : <u>8129509201</u></p> <p>Designation : <u>N/A</u></p> <p>Hospital Seal : <u>N/A</u></p>
--	--



സൂപ്രണ്ട്  
ജില്ലാ ആശുപത്രി  
തൃശ്ശൂർ

[Signature]


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	DH NEDUMANGAD
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Suction Apparatus Make : Supreme Enterprises Model : Supreme Plus Serial No : NA
4	Equipment ID/ Barcode	114956/0112464
5	Date of purchase/ Year of manufacture/Installation Date	16/04/2014
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	04/06/2024(Toll free)
9	Action taken	Checked and found suction motor and jars defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Suction Motor – 4650/- Suction Jars–3600/- Total – 8250/-



13	Asset Value	10200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	80.8%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT VENDOR QUOTATION ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found suction motor and suction jars defective. The unit is installed on 16/04/2014 and covered upto 10+ years. The repairing cost is 80.8%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	MIDHUN BM 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Suction motor and suction jars defective.  
So as per tender conditions can be RBER.

MANEESHA MOHAN M.R.  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
20-07-2024  
Signature of JC BM (NHM)

Date:



  
Signature of Superintendent/Medical Officer (i/c)

സൂപ്രണ്ട്  
ജില്ലാ ആശുപത്രി  
തൃശ്ശൂർ







97447603, 97447624

cloud99biotech@gmail.com

TMC XIII-1333  
Room No: 22  
CSI Golden Jubilee Complex  
Baker Hill  
Shastri Road  
Kottayam

# CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1ZD  
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

## Estimate

Estimate For:

### KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD  
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AA=CC2499H2ZM

State: 32-Kerala

Estimate No.: C998-QTN-28

Date: 09/07/2024

Place of Supply: 32-Kerala

#	Item Name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
<b>Total</b>						<b>₹ 1,095.03</b>	<b>₹ 8,250.00</b>

Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
<b>Total</b>	<b>₹ 8,250.00</b>
You Saved	₹ 0.01



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBI0N0536121



Account holder's name : CLOUD99 BIOTECH

### Description

DH NEDUMANGAD

SUCTION APPARATUS

Barcode: 0112464

Ticket id: 114956

Make : Supreme Enterprises

### Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

### Terms And Conditions


Thanks for doing business with us!

For: CLOUD99 BIOTECH




Authorized Signatory

*Handwritten Signature*  
 സുപ്രണ്ട്  
 ജില്ലാ ആശുപത്രി  
 നെടുമംഗാട്







DONE ON: 27/11/23  
 PMS DUE: 21/12/23  
 PERFORMED BY: [Signature]

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800-425-7669

BAR CODE - 0112464




*[Signature]*  
 സൂപ്രണ്ട്  
 കേരള ആരോഗ്യ സേവന കമ്പനി  
 കൊച്ചി





---

## Quotation suction Machine - DH nedumangad (114956/0112464)

3 messages

---

**Kasyep P V** <kasyepcyrix@gmail.com>

Tue, 9 Jul 2024 at 5:28 pm

To: CLOUD 99 <cloud99biotech@gmail.com>

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;  
Make : Supreme Enterprises

Quotation required;  
Suction motor  
Suction jar

Thanks & Regards,

KASYEP PV  
District Incharge ,Trivandrum  
CYRIX HEALTHCARE Pvt.Ltd  
KLBEMP  
Ph : +91 7593847134  
[kasyepcyrix@gmail.com](mailto:kasyepcyrix@gmail.com)

---

**Kasyep P V** <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

To: CLOUD 99 <cloud99biotech@gmail.com>

Dear sir,  
Consider this mail as gentle Reminder

[Quoted text hidden]

---

**CLOUD 99** <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

To: Kasyep P V <kasyepcyrix@gmail.com>

[Quoted text hidden]