



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 228265

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>T.H.Q.H</u></p> <p>Address : <u>Cherthala</u> <u>Alappuzha</u></p> <p>Ph : <u>9995511497</u></p>	<p>Call Registration Date : <u>29/5/2024</u></p> <p>Caller ID : <u>112915</u></p> <p>Date of Visit : <u>30/5/24</u></p> <p>Asset No. : <u>0420311</u></p> <p>EQPT Name : <u>Deep Freezer</u></p> <p>Manufacture <u>Cel Frost</u> Model : <u>CF-100</u></p> <p>S. No. <u>NA</u> Dept. <u>LAB</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Deep freezer cooling issue

Action Taken : checked the machine and found that the compressor, freezer cooling coil are defective. Need to replace this spares for further checking and working condition of equipment

Completed Date : 30/5/24 Time : 12:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Naseef</u>	<u>30/5/24</u>	<u>10:00am</u>	<u>12:00 PM</u>

Customer Remark : not working

Completed Pending

<p>Service Engineer Name : <u>Naseef</u></p> <p>Signature : <u>Naseef</u></p> <p>Date : <u>30/5/24</u></p> <p>Contact Number : <u>9526744524</u></p>	<p>Customer Name : <u>Syisha S</u></p> <p>Signature : <u>Syisha S</u></p> <p>Date : <u>30/5/24</u></p> <p>Contact Number : <u>8281421922</u></p> <p>Designation : <u>Lab Technician (RNIV)</u></p> <p>Hospital Seal : <u>Lab</u></p>
--	--

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH CHERTHALA
3	Name of Equipment with Make, Model and Serial Number	DEEP FREEZER MAKE :- CELFROST MODEL :- CF-100 SN :- 131241044
4	Equipment ID & Barcode	113915 & 0420344
5	Date of purchase / Year of manufacture /Installation Date	11/12/2014
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29/05/2024
9	Action taken	Checked the machine and found that the Compressor, Freezer cooling coil are defective. Enquired spare with the Vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	Compressor = 6000 Gas Charging = 4500 Freezer Cooling Coil = 2200 <u>Total = 12700</u>

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH CHERTHALA
3	Name of Equipment with Make, Model and Serial Number	DEEP FREEZER MAKE :- CELFROST MODEL :- CF-100 SN :- 131241044
4	Equipment ID & Barcode	113915 & 0420344
5	Date of purchase / Year of manufacture /Installation Date	11/12/2014
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29/05/2024
9	Action taken	Checked the machine and found that the Compressor, Freezer cooling coil are defective. Enquired spare with the Vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	Compressor = 6000 Gas Charging = 4500 Freezer Cooling Coil = 2200 <u>Total = 12700</u>

STOCK REGISTER

Name 79 CHEST TYPE FREEZER

Date	Received from or issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	Ps.				
11-12-14	Received from Calgon Scientific Company Edappully Kochi. In No: 11594 dtd 10.12.2014 Model - CF-100 Make - Cellfront Supplies with Digital Temp Controller and Stabilizer (Eq Code 429-25) Supply Order No 4182014/133 dtd 14.10.2014		20530	-	1 No	-	1 No	✓
11-12-14	Issued to RNTCP Lab					1 No	NIC ✓	

Page No (57)

Received by
S. S. Srinivasan
15/12/14

DA
pbb

Srinivas S
Lab Technician (RNTCP)
15/12/14

Name MO

Date	Remarks
15/12/2014	Received from Scientific Company No: 193 Model - Ke Make - Ke Supply O. dtd 20 Static 1810 x 6 Mobile 1810 x 4 Mobile 1810 x 9

15/12/14 Issued



[Handwritten signature]
15/12

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA

