



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

**No. :** 232707

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Thiruvananthapuram</u> <u>Kerala</u></p> <p>Ph : <u>9495469434</u></p>	<p>Call Registration Date : <u>4/6/24</u></p> <p>Caller ID : <u>115020</u></p> <p>Date of Visit : <u>5/8/24</u></p> <p>Asset No. : <u>0110375</u></p> <p>EQPT Name : <u>Vibrasound therapy Unit</u></p> <p>Manufacture <u>Stzyku</u> Model : <u>Bio tech</u></p> <p>S. No. <u>4963</u> Dept. _____</p>
---	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** Checked and found that the high voltage card and 1Mhz ultrasound probe defective. Need to replace the spare for further checking and working condition of the equipment


Completed  Date : 5/8/24 Time : 12:50 PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Soomeya Sudheep</u>	<u>5/8/24</u>	<u>12:15PM</u>	<u>12:50PM</u>

**Customer Remark** Pending



<p>Service Engineer Name : <u>Soomeya Sudheep</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>5/8/24</u></p> <p>Contact Number : <u>74107406244</u></p>	<p>Customer Name : <u>Deepu B</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>5/8/24</u></p> <p>Contact Number : _____</p> <p>Designation : _____</p> <p>Hospital Seal : _____</p>
--	---


General Hospital  
Thiruvananthapuram

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name: Ultrasound Therapy Unit Make : Stryker Model : Biotech Serial No : 4963
4	Equipment ID/ Barcode	115020/0110375
5	Date of purchase/ Year of manufacture/Installation Date	19/09/2018
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	04/06/2024
9	Action taken	Checked and found that the high voltage card and 1 Mhz ultrasound probe defective. Need to replace the spares for further checking and working condition of the equipment. Enquired spares from vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	1 Mhz Applicator - 2968/- High voltage PCB - 8850/- Freight cost - 236/- Total - 12054/-

13	Asset Value	14600/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	82.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND OEM QUOTATION ATTACHED
16	Reasons for recommending the equipments as BER	Checked and found that the high voltage PCB and 1 Mhz ultrasound probe defective. The unit was installed on 19/09/2018 and covered upto 6 + years. The repairing cost is 82.5%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

High voltage PCB and 1 MHz ultrasound probe defective. Estimate verified and as per tender conditions this can be BER.

Call date : 4-6-2024

Submitted date : 6-8-2024

MANEESHA MOHAN M.P.  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram  
6-8-2024  
Signature of JC BM (NHM)

Date:



Signature of  
Superintendent/Medical Officer (i/c)

Superintendent  
General Hospital  
Thiruvananthapuram

# UST - Ultrasound Therapy

35

DATE	Name & Address of Supplier with Supply orders No.	Receipt	Balance	Inventory	Period
13-4-16	Bharat Medical Systems M-1074 Main Road Thiruvalluvar Nagar Chennai - 41 Cost: 22000 + 5% Vat Installed on 11-4-16 Service Em: 9444052532 Sl. No - 9213 Page No - 73 Vol: 1	1	1		11-4-16 to 10-4-2018
13/4/16	Bharat Medical Systems M-1074 Main Road Thiruvalluvar Nagar Chennai - 41 Cost: 22000 + 5% Vat Installed on 11-4-2016 Service Em: 9444052532 Sl. No - 9217 Page No - 73 Vol: 1	1	2		11-4-16 to 10-4-2018
19/9/18	Biotech Micro Control Orders No - 101/223/2017-18/130 dated 11-1-2018 Model - Striker SR. No. 4963 Cost - Rs. 14600 + 12% Tax Service Em: Sajith Mob: 9995665834 Page no - 73 Vol - 1	1	1		19-9-18 to 18/9/2020



## MEIBOT TECHNOLOGIES

ARA 44, 2nd Floor, Thyparambil Building,  
Ammankovil Street Mullackal,  
Alappuzha Kerala- 688011

Phone no.: 9447038191 Email: info@meibot.com

GSTIN: 32BVOPM0744D1ZT, State: 32-Kerala

### Estimate

#### Estimate For

Cyrix Healthcare Pvt Ltd

FIRST FLOOR, 30/641B  
PETTA JUNCTION,  
POONITHURA, ERNAKULAM-682038,  
KERALA INDIA

GSTIN : 32AAFCC2499H2ZM

State: 32-Kerala

#### Estimate Details

Estimate No. : MT/QT/24-25/111

Date : 16-07-2024

Place of supply: 32-Kerala

#	Item name	Item Code	HSN/SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	1Mhz Applicator 5cm.sq. with crystal	05100	90189099	1	Nos	₹ 2,650.00	₹ 318.00 (12%)	₹ 2,968.00
2	Services & Repairing Charges (Low Voltage pcb replacement for UST-0110298)	SRV-REP	995469	1	Nos	₹ 6,500.00	₹ 1,170.00 (18%)	₹ 7,670.00
3	freight			1	-	₹ 200.00	₹ 36.00 (18%)	₹ 236.00
	Total			2			₹ 1,524.00	₹ 10,874.00

Tax type	Taxable amount	Rate	Tax amount	Amounts
				Sub Total
SGST	₹ 2,650.00	6%	₹ 159.00	₹ 10,874.00
CGST	₹ 2,650.00	6%	₹ 159.00	
SGST	₹ 6,700.00	9%	₹ 603.00	
CGST	₹ 6,700.00	9%	₹ 603.00	
				<b>Total</b>
				<b>₹ 10,874.00</b>

#### Estimate Amount In Words

Ten Thousand Eight Hundred Seventy Four Rupees only

#### Terms and Conditions

Terms and conditions:

Payment Terms: 100% Advance

The Guarantee/Warranty would not cover: -

- No Warranty on Rubber, Plastic & Wooden Products / Parts.
- Any Mechanical damage/breakage.
- Any damages to vacuum Tube / Valve (if any), Meters.
- Any damages due to electrical fluctuation.
- Supply of consumables and accessories.
- Any Physical damage to the Device due to Intentional act or wilful neglect.
- Any Equipment's Serial number that has been altered, defaced or removed, or has been modified to alter its functionality or capability without the written permission of Hamb Meditech LLP.

#### Bank Details



Name : STATE BANK OF INDIA

Account No. : 41310832759

IFSC code : SBIN0003106

Account holder's name : Meibot Technologies



For : MEIBOT TECHNOLOGIES

**Authorized Signatory**



---

## Quotation for the service of UST Machine

3 messages

---

**MURALIKRISHNA S** <muralikrishnaalpy@gmail.com>  
To: kasyepcyrix@gmail.com

Wed, 17 Jul 2024 at 6:54 pm

Dear Sir/ Madam,

Kindly find the attached quotation.

Thanks and Regards

Murali Krishna  
Sales & Service Engineer

Whatspp: <https://wa.me/919074126766>

---

**MURALIKRISHNA S** <muralikrishnaalpy@gmail.com>  
To: kasyepcyrix@gmail.com

Sat, 27 Jul 2024 at 10:12 am

Dear Sir / Madam,

Kindly find the attached quote.

—  
Thanks and Regards

**Muralikrishna Suresh**  
**Sales & Service Manager**

Whatspp: <https://wa.me/919074126766>

[Quoted text hidden]

---

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: MURALIKRISHNA S <muralikrishnaalpy@gmail.com>

Mon, 29 Jul 2024 at 1:56 pm

Sir,  
Recieved the estimate, will do the payment on 30/7/2024

Thanks & Regards,

KASYEP PV  
District Incharge ,Trivandrum  
CYRIX HEALTHCARE Pvt.Ltd  
KLBEMP  
Ph : +91 7593847134  
kasyepcyrix@gmail.com

[Quoted text hidden]

IWS

Micro-Processor Controlled  
Therapy Unit  
Ultrasound



BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM

BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAM (BEMP)

TOLL FREE NO: 1-800-425-7169

CALL SIDE - 0110975



SCL

CYRIX HEALTH  
BEMP SERVICE  
PREVENTIVE MAINT

PMS DONE ON: 3/3

NEXT PMS DUE: 2/5

PERFORMED BY: J...



15

Micro-Processor Controlled  
Ultrasound Therapy Unit

SW1

SW2

Striker  
GET IT FOR GOOD

Biotech

Product  
No. 04