



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 228262

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : DHC	Call Registration Date : 24/5/2024
Address : Vayalax Alappuzha	Caller ID : 113008
Ph : 9074152061	Date of Visit : 25/5/2024
	Asset No. : 0442676
	EQPT Name : UPS
	Manufacture BPE Model : MF1101B3
	S. No. M024700 Dept. LAB

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : UPS is not getting ON

Action Taken : Checked the machine and found that the control card, Driver card and Batteries are defective need to replace this spares for further checking and working condition of Equipment.

Completed Date : 25/5/24 Time : 12:00 PM Spare Required
Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Nasuf	25/5/24	10:30am	12:00 PM

Customer Remark : Not working

Completed Pending

Service Engineer Name : Nasuf	Customer Name : Rejith PC
Signature : <i>Nasuf</i>	Signature : <i>Rejith PC</i>
Date : 25/5/24	Date : 25/5/24
Contact Number : 9526744524	Contact Number : 9995759501
	Designation : Lab Technician
	Hospital Seal :





**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 228300

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>PHC</u>	Call Registration Date : <u>24/5/2024</u>
Address : <u>Vayalar</u>	Caller ID : <u>113002</u>
<u>Alappuzha</u>	Date of Visit : <u>25/5/2024</u>
Ph : <u>9074152061</u>	Asset No. : <u>0442676</u>
	EQPT Name : <u>UPS</u>
	Manufacture <u>BPE</u> Model : <u>MF1101B3</u>
	S. No. <u>MDALI7C0056</u> Dept. <u>LAB</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : UPS is not getting ON

Action Taken : checked the machine and found that the control card, Driver card and Batteries are defective. Equipped Spare with the vendor.

Completed Date : 29/6/24 Time : 3:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Naseef</u>	<u>29/6/24</u>	<u>1:30 PM</u>	<u>3:00 PM</u>

Customer Remark Completed Pending

Not working
Rejith P.C
Sub Technician

Service Engineer Name : Naseef
Signature : Naseef
Date : 29/6/24
Contact Number : 9526744524

Customer Name : PH Centre, Vayalar
Signature : [Signature]
Date : [Date]
Contact Number : [Number]
Designation : [Designation]
Hospital Seal : [Seal]




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	PHC VAYALAR
3	Name of Equipment with Make, Model and Serial Number	UPS 1KVA MAKE :- BPE MODEL :- MF1101B3 SN :- M02L17C00862
4	Equipment ID & Barcode	113008 & 0442676
5	Date of purchase / Year of manufacture /Installation Date	13/10/2017
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24/05/2024
9	Action taken	Checked the machine and found that the Control Card, Driver Card and Batteries are defective. Enquired spare with the Vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	Control Card = 2000 Driver Card = 2000 Battery 12v,7Ah (2nos:) = 2400 Total = 6400

13	Asset Value	RS :- 4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	142%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report And Quotation attached.
16	Reasons for recommending the equipment as BER	Checked the machine and found that the Control Card, Driver Card and Batteries are defective. The equipment was installed on 13/10/2017 and aged up to 6 years and 7 months. Spare cost is 142% both criteria met. So we are recommending the equipment for condemnation as per the tender clause 5.3.14.1.
17	Name & Signature of CYRIX Authority	NASEEF K N 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


15/07/24
Signature of JG BM (NHM)
JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM Alappuzha



Date

Seal

for condemnation or approval
Signature of
Superintendent / Medical Officer
PH Centre, Vayalar
Alappuzha (Dist.)



Dr. Anitha
Medical Officer
PH Centre, Vayalar
Alappuzha (Dist)

CALORIMETER

Date	particulars	Recepd	Issue	Balance
29/7/09	Received from Store	1		1
29/7/09	Issue to use		1	NIL

SEMI AUTO ANALYSER

30/3/16	Received from store	1		1
30/3/16	Issue to use (Instal on 19-5-16)		1	NIL

Fully Auto Hematology Analyser

13/10/17	Received from store	1		1
13/10/17	Issue to		1	
	UPS with back up model		1 no	
	Driluent 20 L for PE 6		2 no	
	lyse 500 ml for PE 6		2 no	
	probe (clean) 100 ml for cell count		1 no	
	E2 (cleaner) 100 ml for cell counter		1 no	



Dr. D. Chithra

Medical Officer
PH Centre, Vayalar
Alappuzha (Dist)

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 29/06/2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN : 32CGUPM9450L1ZN

1KVA BPE ONLINE UPS SERVICE PROPOSAL

TO,
CYRIX HEALTHCARE PVT. LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

MAKE : BPE
MODEL: MF1101B3

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	CONTROL PCB	1	2000.00	2000.00	18	2360.00
2	DRIVER PCB	1	2000.00	2000.00	18	2360.00
3	12V 7Ah BATTERY	2	1200.00	2400.00	28	3072.00
PAYABLE AMOUNT						7792.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20%After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares, 28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN

