



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223254

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THRH</u></p> <p>Address <u>Parassala</u></p> <p><u>Tiruvallur</u></p> <p>Ph : <u>9495431928</u></p>	<p>Call Registration Date : <u>23/5/24</u></p> <p>Caller ID : <u>112821</u></p> <p>Date of Visit : <u>24/5/24</u></p> <p>Asset No. : <u>0120487</u></p> <p>EQPT Name : <u>Suction Apparatus</u></p> <p>Manufacture <u>Supreme</u> Model : <u>instance</u></p> <p>S. No. <u>na</u> Dept. <u>OT</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working

Action Taken : Checked the machine and found that suction motor and Jar damaged. asked to replace the spares for further checking and working condition of the machine. Enquired spares with OEM.

Completed Date : 24/5/24 Time : 1pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Asith George</u>	<u>24/5/24</u>	<u>12pm</u>	<u>1pm</u>

Customer Remark Completed Pending



Service Engineer Name : Asith George

Signature : [Signature]

Date : 24/5/24

Contact Number : 9593897111

Customer Name : valsaldamavu

Signature : [Signature]

Date : 9/5/24

Contact Number : 9495431928

Designation : SND


Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No.	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH PARASSALA
3	Name of Equipment with Make, Model and Serial Number	Equipment : Suction machine Make : Supreme Model : Instavac Sl : NA
4	Equipment ID/ Barcode	112831 / 0120487
5	Date of purchase/ Year of manufacture/Installation Date	30/09/2020
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	23/05/2024
9	Action taken	Checked and found the suction motor and the suction jars defective. Need to replace the spares for further checking and working condition of the equipment. Enquired spare with vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Suction Motor - Rs.4650/- Suction Jars - Rs.3600/- Total - Rs.8250/-

13	Asset Value	Rs.9900/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	83.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and vendor quotation attached.
16	Reasons for recommending the equipment as BER	Checked and found suction motor and the jars defective. The equipment was installed on 30/09/2020 and aged up to 3+ years. Repairing cost of the equipment is 83.3%. As per the tender clause 5.3.14.1 both criteria met, So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Suction motor and jars defective.
So can be BER

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


10.7.2024
Signature of JC BM (NHM)

Date:





Signature of
Superintendent/Medical Officer (i/c)
പ്രസിഡൻ്റ്

സംയുക്ത ആസ്ഥാന ആശുപത്രി
പാലക്കാട്

Section Apparatus

	R	I	B	
5/11 Received from store	1	✓	1	
8/1/14 Sent to HER for repairing			all	
4/14 Received after repair	1		1	
30/7/14 (Vol 1) BF from old stock register Page No. 118	3		4	Rel on
5/12/14 Sent to HER for repair. 1 No.			5	
15/10/14 Received from LR (P.No. 50)	1		4	
3/15 1 No transferred to unserviceable register (page 11)			6	
6/6/15 Recd from stock	2		5	
15/10/15 1 No transferred to unserviceable register (page 14)	1			
① Condemnation done as per order no. c/1068/15 ② Condemnation done as per order no. c/2350/2015 alt on 26/2/16				
Recd from store			5	
Instalar - 1025 Rel from stock			1	
Rel from stock (model R5-02)			5	
17/7/17 Received from female const. ward				



5/11/14
 30/9/2020
 RECEIVED FROM STORE
 1025



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q120
State: 32 - Kerala

MSME NO: : UDYAM-KL-07--0030714

Estimate

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT
LTD KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C998-QTN-17

Date: 27/05/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
2	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
Total						₹ 1,095.03	₹ 8,250.00



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH



Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
You Saved	₹ 0.01

Description

EQUIPMENT: SUCTION APARATUS
MAKE: INSTAVAC
TICKET ID: 112831
BARCODE: 0120487

Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



[Handwritten Signature]
മുഖ്യൻ
മാലൂക്ക് ആസ്ഥാന ആരോഗ്യപരി
2024





Repairing Quotation suction ID 112831

1 message

CLOUD 99 <cloud99biotech@gmail.com>
To: kasyepcyrix@gmail.com

Tue, 28 May 2024 at 11:36 am

SUSAN JOHN
Cloud99 Biotech
Medical Engineering Works
Room No: 22, CSI Golden Jubilee Complex
Kottayam 686 001
Kerala, India
Ph: 9744473043
email: cloud99biotech@gmail.com




വൃപണ്ട്
താലൂക്ക് ആസ്ഥാന ആഭ്യുപരി
പാറശ്ശാല.