

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)


PROFORMA


| Sl. No | Particulars   | Details   |
|--------|---|---|
| 1      | Name of District  | Ernakulam   |
| 2      | Name of Hospital  | 24x7 KUTTAMPUZHA  |
| 3      | Name of Equipment with Make, Model and Serial Number                          | Equipment: LABOPRATORY MICROSCOPE<br>Make :LABOVISION<br>Model : KXI 2000<br>Sr.no : 3185554 <span style="float: right;">29962</span>         |
| 4      | Equipment ID & Barcode  | 112336 & 0746580  |
| 5      | Date of purchase / Year of manufacture /Installation Date                     | 18-01-2018  |
| 6      | Warranty details (Yes/No)   | No warranty   |
| 7      | *AMC/ CAMC Period agreed at the time of purchase                              | No AMC/CAMC   |
| 8      | Date of breakdown(Date of registration of complaint through email/ Toll free) | 21-05-24 (Toll Free)  |
| 9      | Action taken  | Checked the machine and found that it is a flood affected equipment, eyepiece, lens ,stage are faulty also the machine is in rusty condition. |
| 10     | Present status of the equipment (Fully damaged / partially damaged)           | Fully damaged   |
| 11     | Recommendations for repair (required service details)                         | Not recommending for repair.  |
| 12     | Cost of spares (specify parts and cost)                                       | NA  |




# STOCK REGISTER

| Date     | Name of Article           | Received from or Issued to       | Inv. No. or V. No. | Amount | Receipts | Issues | Balance       | Remarks |
|----------|---------------------------|----------------------------------|--------------------|--------|----------|--------|---------------|---------|
|          | 4 BINOBU A.P. Microscope  |                                  |                    |        |          |        |               |         |
| 2        | Received from store       |                                  |                    |        | 1        |        | 1             |         |
|          | Model                     | Vision                           | 2000               |        |          |        |               |         |
|          | Make                      | N.K. Jain                        |                    |        |          |        | 2             |         |
| 18/11/18 | Received from store       |                                  |                    |        |          |        | Rate 21,288/- |         |
|          | Model                     | Kx1                              | 2000               |        |          |        |               |         |
|          | Make                      | labor vision                     |                    |        |          |        |               |         |
|          | Invoice No                | 0091                             | Serial No          |        | 3195554  |        |               |         |
|          | Warranty                  | 22/1/19                          | 22.1.2020          |        |          |        |               |         |
|          | Konnect supply order No   |                                  | 50/452/2017-18/13c |        |          |        |               |         |
|          | Note                      | working due to flood on Aug 2018 |                    |        |          |        |               |         |
| 21/1/19  | Received from store       |                                  |                    |        |          |        |               |         |
|          | Model                     | Vision 2000                      |                    |        |          |        |               |         |
|          | Invoice No                | 44 277/2018                      |                    |        |          |        |               |         |
|          | Warranty                  | 2 year                           |                    |        |          |        |               |         |
| 20/1/19  | Received from store       |                                  |                    |        |          |        |               |         |
|          | Model                     | Make KR L-5B                     |                    |        |          |        |               |         |
|          | Make                      | RADICAL INDIA                    |                    |        |          |        |               |         |
|          | Konnect supply Invoice No |                                  | D 657/3584         |        |          |        |               |         |
|          | Order No                  |                                  | 50/762/2017-20     |        |          |        |               |         |
|          | Warranty                  | 2 year                           |                    |        |          |        |               |         |
|          | Project name              | N.M. 420                         |                    |        |          |        |               |         |
|          | management equipment      |                                  |                    |        |          |        |               |         |

  
 Dr. Anoop



MEDICAL OFFICER  
 Family Health Centre  
 Kuttampuzha - 686 601



Dr. Anoop Thulas

**MEDICAL OFFICER**  
Family Health Centre  
Kuttampuzha - 686 681



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMIP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890615225)  
0746580



LABOVISION  
T/F  
EGLASS  
GERMANY

INPUT : 200V-240V, AC-50Hz  
LED LAMP 3W

D. Anoop Tulas

MEDICAL OFFICER  
Family Health Centre  
Kuttampuzha - 686 681






*[Handwritten signature]*

*D. Anoop Thulas*

**MEDICAL OFFICER**  
Family Health Centre  
Kuttamouzha - 686 681





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 171591

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ..... phc  
Address ..... Kuttamparba  
..... Ernakulam  
Ph : ..... 9495289358

Call Registration Date : ..... 21/5/24  
Caller ID : ..... 112336  
Date of Visit : ..... 21/5/24  
Asset No. : ..... 0746580  
EQPT Name : laboratory microscope  
Manufacture labomision Model : NA  
S. No. 3185554 Dept. Laboratory

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... unit not working.

**Action Taken :** ..... checked the unit, check working found that during the flood, the entire lab was flooded. all the machinery was destroyed and the asset was incorrectly labeled Barcoded. while check unit totally damaged, mirror, lens, Eye piece, adaptor, etc.

Completed  Date : ..... 21/5/24 Time : ..... Spare Required

Spare Replaced  Requested


| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1.          |      |             |           |
| 2.          |      |             |           |
| 3.          |      |             |           |

| Cyrix Engineer         | Date           | Start Time | End Time |
|------------------------|----------------|------------|----------|
| <u>Akshil T. Lenin</u> | <u>21/5/24</u> |            |          |

Customer Remark .....  Completed  Pending

**Service Engineer Name :** Akshil T. Lenin  
**Signature :**   
**Date :** 21/5/24  
**Contact Number :** 9605474326

**Customer Name :** Resmi S. R.  
**Signature :**   
**Date :** 21/5/24  
**Contact Number :** 9495289358  
**Designation :** Laboratory technician  
**Hospital Seal :**



From,

Medical officer  
FHC Kuttampuzha

To,


CYRIX

Sir,

3 microscopes received from Kmscl were damaged in 2018 Flood. But CYRIX barcoded them and found them useless. In 2019 got a new microscope from Kmscl. It is currently being used in the lab. So these three microscopes are to be removed from the lab.

27/6/24  
Kuttampuzha



  
Medical officer  
Dr. Anoop Thomas  
MEDICAL OFFICER  
Family Health Centre  
Kuttampuzha - 686 681