


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	24x7 KUTTAMPUZHA
3	Name of Equipment with Make, Model and Serial Number	Equipment: LABORATORY MICROSCOPE Make: LABOMED Model: VISION 2000 Sr.no : NA 15400
4	Equipment ID & Barcode	112332 & 0746583
5	Date of purchase / Year of manufacture /Installation Date	24-01-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	21-05-24 (Toll Free)
9	Action taken	Checked the machine and found that it is a flood affected equipment, eyepiece, lens ,stage are faulty also the machine is in rusty condition.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs.15,400
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that it is a flood affected equipment, eyepiece, lens, stage are faulty also the machine is in rusty condition. Equipment installed on 24-01-2019 and the machine outlived 5+years.As it is an old machine and in rusty condition, the customer not prefer this equipment for daily use in laboratory. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 This equipment is kept idle noted during my institute visit and registered the call. It is observed that the same is flood affected and damaged on 2018. By mistake cyrix harassed the unit. Registered the call and supported complete damage from cyrix side. This institution is completely affected by flood and almost all lab equipments were flooded and damaged during that time. Not Submitted any Quot. Asset value is Rs. 15400/-

Signature of JC.BM (NHM)

24/6/24
 Bio-medical Consultant
 National Health Mission (NHM)
 Arogyakaram Ernakulam


 Dr. Anoop Thulasi
 Superintendent / Medical Officer (i/c) Signature of seal Date

MEDICAL OFFICER
 Family Health Centre
 Kuttampuzha - 686 681




STOCK REGISTER

4

MICROSCOPE


Name of Article BINOCULAR

Date	Received from or Issued to	Inv. No. or V. No.	Amount Rs.	Receipts Rs.	ISSUES	Balance	Remarks
	Received from store				1	1	
	Model: VISION	2000					
	Make: N.K. Jain				1	2	
19/11/18	Received from store					Rate 21,288/-	
	Model: KX1	2000					
	Make: LABORVISION						
	Invoice No: 0091				3195554		
	Warranty: 22/1/18 - 22.1.2020						
	Kansar Supply order No: 50/452/2017-18/13c						
	Note: working due to flood on Aug 2018						
24/1/19	Received from store				1	2	
	Model: VISION	2000					
	Make: N.K. Jain						
	Invoice No: 4A 2777/2018				28/1/19	Rate 19712/-	
	Warranty: 2 year						
20/9/19	Received from store				1	3	
	Model: Make: KX1-5B					Rate 22116/-	
	Make: RADICAL INDIA						
	Kansar supply Invoice No: D GST/3564						
	Order No: 50/762/2017-20					31/8/2019	
	Warranty: 2 year						
	Project: N.A.M. 420						
	management equipment						Replanning of food


 Dr. Anoop Kulan

MEDICAL OFFICER
 Family Health Centre
 Kuttampuzha - 686 681





SYGFA
TOLL FREE NO: 1800 - 452 7446
www.sygfa.org (0824 891013234)
0746583


Dr. Anoop Chulani
MEDICAL OFFICER
Family Health Centre
Kuttampuzha - 686 681

Centre
Dist.
Date.....
KUTTAMPUZHA

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)
0746583



MEDICAL OFFICER
Family Health Centre
Kuttampuzha - 686 681

[Handwritten Signature]
Dr. Anoop Thomas



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 171592

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency
30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility phc	Call Registration Date : 21/5/24
Address Kuttampuzha	Caller ID : 112332
..... Ernakulam	Date of Visit : 21/5/24
Ph : 9495289358	Asset No. : 0746583
	EQPT Name : Laborety microscope
	Manufacture Labomed Model : uspon2000
	S. No. Na Dept. laboratory

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : unit not working.

Action Taken : check the unit working, found that during the flood, the entire lab was flooded all the machinery was destroyed and the asset was incorrectly Barcoded. while checking the unit total damage, adapter, lens, mirror's, Body corroded, etc.


Completed Date : 21/5/24 Time : Spare Replaced Requested Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Akmal T. Lenin	21/5/24		

Customer Remark Completed Pending

Service Engineer Name : Akmal T. Lenin	Customer Name : Resmi. S. P.
Signature :	Signature :
Date : 21/5/24	Date : 21/5/24
Contact Number : 9605474320	Contact Number : 9495289358
	Designation : Laboratory technicians U R I
	Hospital Seal :



From,

Medical officer
FHC Kuttampuzha


To,


CYRIX

Sir,

3 microscopes received from KMSCL were damaged in 2018 Flood. But CYRIX barcoded them and found them useless. In 2019 got a new microscope from KMSCL. It is currently being used in the lab. So these three microscopes are to be removed from the lab.

27/6/24
Kuttampuzha




Medical officer
Dr. Anoop Thulasi
MEDICAL OFFICER
Family Health Centre
Kuttampuzha - 686 681