



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223867

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility : THQH	Call Registration Date : 18/05/2024
Address : Attingal	Caller ID : 11914
Thiruvananthapuram	Date of Visit : 18/07/2024
Ph : 9766407566	Asset No. : 0120831
	EQPT Name : Sterilizer
	Manufacture : Steris Model : Steris
	S. No. : NA Dept. : DT

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT WORKING

Action Taken : Checked and found the heater coil, door gasket, pressure gauge, door alignment defective and the unit is rusted so that the unit is not suitable for the safety use

Completed Date : 18/7/24 Time : 11am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Deepak. E.M	18/7/24	10.30	11am

Customer Remark Completed Pending

Service Engineer Name : Deepak. E.M
Signature :
Date : 18/7/24
Contact Number : 7902881038


Customer Name : Sekh.K
Signature :
Date : 18/7/24
Contact Number : 9600000000
Designation : Superintendent
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH ATTINGAL
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Sterilizer Make : Steri Model : NA Serial No : NA
4	Equipment ID/ Barcode	111914/0120831
5	Date of purchase/ Year of manufacture/ Installation Date	21/06/2006
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	18/05/2024(Toll free)
9	Action taken	Checked and found the heater coils, door gasket, pressure guage, door alignment defective and the unit is rusted, so that the unit is not suitable for the safety use.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	49491/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND HOSPITAL SIDE LETTER ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found that the heater coils, door gasket, pressure gauges, door alignment defective and the body rusted. The unit was installed on 21/06/2006 and covered upto 18+ years. The unit is too old and not suitable for daily use by considering customer safety. Hospital side letter requesting condemnation attached. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	DEEPAK EM 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Heater coil, door gasket, pressure gauge etc defective - So can be BER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


20.07.2024
Signature of JC BM (NHM)

Date: _____





Signature of
Superintendent/Medical Officer (i/c)
Superintendent
Taluk Hospital
Atingai

Autoclave


Date	Description	Cost	Received	Issues	Balance	Signature
040						
12.1.2021	B/E form Vol 1 page No: 59		4		4	<i>[Signature]</i>
	D/R					
	25.6.2002 - 1 No.s					
	21.6.2006 - 1 No.s					
	1.10.2015 - 1 No.s <i>(for (medical) treatment)</i>		59623/			
	2.11.2018 - 1 No.s <i>(Foreign)</i>					
	condemnation done as per order		1 No		3 250-	
	No. 02/2021/2017/112/112					
	dated on 20/1/2021					

[Signature]
for Peter

**PHARMACY STORE KEEPER
TALUK HOSPITAL ATINGAL
PIN**



[Signature]
Superintendent
Taluk Hospital
Atingal



[Handwritten signature]
Superintendent
Taluk Hospital
Atingai

From

09 Incharge

Taluk Hospital

Attingal

To

The medical officer

Taluk hospital

Attingal.

Respected madam,

09 മിറാക്സ് autoclave (vertical) Bar code
 0120831 ലഭ്യമാക്കുന്ന രജിസ്ട്രേഷൻ. ഇത്
 മെഡിക്കൽ സർവ്വീസുകളിനുള്ള ഓട്ടോമേറ്റഡ്
 റെഗിസ്ട്രേഷൻ സമ്പ്രദായത്തിൽ ഉൾപ്പെടുത്തണമെന്നും

✓
 26/6
 2024
 Supt

Superintendent
 Taluk Hospital
 Attingal

മേൽ

മുദ്ര

Sd/-

Selvi.K

Senior of officer

26.6.24

അസിസ്റ്റന്റ് സൂപ്പർവൈസർ

