



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 169989

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>P.N.C.</u>	Call Registration Date : <u>15/5/2024</u>
Address : <u>Kannikuzhy</u>	Caller ID : <u>111244</u>
<u>Alappuzha</u>	Date of Visit : <u>16/5/2024</u>
Ph : <u>9847217446</u>	Asset No. : <u>0440856</u>
	EQPT Name : <u>O₂ concentrator</u>
	Manufacture <u>EverFlow</u> Model : <u>NA</u>
	S. No. <u>0068145</u> Dept. <u>Nursing Station</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : O₂ purity is too low and Alarm issue machine is not running continuously, switching OFF with in 5 minutes

Action Taken : checked the machine and found that the main board and sieve beds are defective need to replace these spars for further checking and working condition of equipment.

Completed Date : 16/5/24 Time : 2:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nasret</u>	<u>16/5/24</u>	<u>12:00 PM</u>	<u>2:00 PM</u>

Customer Remark Completed Pending

Not working

Service Engineer Name : Nasret
Signature : [Signature]
Date : 15/5/2024
Contact Number : 9526744524

Customer Name : Ranjana
Signature : [Signature]
Date : 16/5/24
Contact Number : 8078555192
Designation : Nursing Officer
Hospital Seal :



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 231300

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC</u></p> <p>Address <u>Kanjikkuzhi</u></p> <p>..... <u>Alappuzha</u></p> <p>Ph : <u>9847217446</u></p>	<p>Call Registration Date : <u>15/5/2024</u></p> <p>Caller ID : <u>111244</u></p> <p>Date of Visit : <u>16/5/2024</u></p> <p>Asset No. : <u>0440856</u></p> <p>EQPT Name : <u>O2 Concentrator</u></p> <p>Manufacture <u>Respirationics</u> Model : <u>Everflow</u></p> <p>S. No. <u>0068195</u> Dept.</p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : O2 Purity is too low and Alarm issue machine is not running continuously, switching OFF with in 5 minutes

Action Taken : Checked the machine and found that the main board and servo beds are defective. Enquired quotation with the vendor.

Completed Date : 1/5/2024 Time : 2:00 PM Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nasut</u>	<u>1/5/24</u>	<u>12:00 PM</u>	<u>2:00 PM</u>

Customer Remark Not working Completed Pending


<p>Service Engineer Name : <u>Nasut</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>1/5/2024</u></p> <p>Contact Number : <u>9526744524</u></p>	<p>Customer Name :</p> <p>Signature : <u>[Signature]</u></p> <p>Date :</p> <p>Contact Number :</p> <p>Designation :</p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	PHC KANJIKKUZHI
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR MAKE :- RESPIRONICS MODEL :- EVERFLO SN:- 0068145
4	Equipment ID & Barcode	111244 & 0440856
5	Date of purchase / Year of manufacture /Installation Date	19/09/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	15/05/2024
9	Action taken	CHECKED THE MACHINE AND FOUND THAT THE MAIN BOARD AND SIEVE BEDS ARE DEFECTIVE. ENQUIRED QUOTATION WITH THE VENDOR.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDED FOR REPAIR
12	Cost of spares (specify parts and cost)	SIEVE BED (2) = 30000 MAIN BOARD = 12000 TOTAL = 42000


13	Asset Value	RS:- 26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	158%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND SPARE QUOTATION IS ATTACHED.
16	Reasons for recommending the equipment as BER	CHECKED THE MACHINE AND FOUND THAT THE MAIN BOARD AND SIEVE BEDS ARE DEFECTIVE. THE MACHINE WAS INSTALLED ON 19/09/2018 AGED UP TO 5+ YEARS. REPAIR COST OF THE EQUIPMENT IS 158%. BOTH CRITERIA MET AS PER THE TENDER CLAUSE 5.3.14.1 SO WE RECOMMENDING THE EQUIPMENT AS RBER.
17	Name & Signature of CYRIX Authority	NASEEF K N, 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


16/08/24
JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM ALAPPUZHA
Signature of JC BM (NHM)

for Continuation.

Date



Seal


Signature of
Superintendent / Medical Officer (i/c)




 ജഡിക്കൽ ആഫീസുകൾ
 മുടങ്ങാത്തതുകൊണ്ട്
 തിരിച്ചറിയുക



BCS Healthcare Pvt.Ltd
3136, IInd door, Mathewsons square
MetroPillar No:464 NearChangapuzha park
Cochin, Kerala – 682024,
Email : info@bcshealthcare.in
web: www.bcshealthcare.in



GST NO: 32AAKCB2766M1Z

DATE: 01/08/2024

Ref: BCS/105/24-25

QUOTATION

TO,
CYRIX HEALTHCARE PVT LTD

SL.NO	PRODUCT NAME	QTY	UNIT RATE	TAX 12%	TOTAL
01	Oxygen Concentrator Make :- Respironics Model :- Everflo				
	Sievebed (2)		26,785.71	3,214.29	30,000.00
	Main board	1	10,714.28	1,285.72	12,000.00
Grand Total					42,000.00

Terms & Conditions

- Barcode - 0440856
- Ticket no - 111244
- Delivery Within 8 Days After Confirmation Of Purchase Order
- Payment 50% Advance Balance 50% After Delivery Time



For, BCS HEALTH CARE PRIVATE LIMITED



Abhilash A <abhilashacyrix@gmail.com>

Quotation

1 message

BCS Healthcare <info@bcshealthcare.in>
To: Abhilash A <abhilashacyrix@gmail.com>

Fri, Aug 9, 2024 at 1:43 PM

Respected Sir/Madam
Please find the attached file

Regards,
BCS HEALTH CARE PRIVATE LIMITED

 **CYRIX HEALTHCARE PVT LTD.pdf**
215K