


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC PANOOR
3	Name of Equipment with Make, Model and Serial Number	X RAY MACHINE Make: ALLENGERS Model: Mars 15 SN: 2K140615253-X/HF
4	Equipment ID/ Barcode	1330640 #119447
5	Date of purchase/ Year of manufacture/Installation Date	18-05-2015
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 26-06-2024
9	Action taken	Checked the Machine found Equipment is not switching on, found that Generator Unit and Tube Head Is Defective Enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Generator Unit(full set)- 504000/- Tube head- 200000/- Total- 7,28,000/- Include GST

13	Asset Value	186000/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	391.397%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Machine found Equipment is not switching on , found that Generator Unit and Tube Head Is Defective Enquired OEM for quotation. Machine installed on 18-05-2015 aged up to 9 years 2 months. Repairing cost is 391.397% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that the machine not switching on, due to generator unit, Tube head one defective. Aged up to 9 years & above. The repairing cost economically not feasible. So recommended for BER as per BER Protocol.


Signature of JC BM (NHM)
22/07/24



22/07/24
Date:


Dr. Anitha

MEDICAL OFFICER
Community Health Centre
Signature of ANOOR
Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
225198

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC PANOOR</u></p> <p>Address <u>PANOOR</u> <u>KANNUR</u></p> <p>Ph : <u>8547251751</u></p>	<p>Call Registration Date : <u>26-6-24</u></p> <p>Caller ID : <u>119447</u></p> <p>Date of Visit : <u>28-6-24</u></p> <p>Asset No. : <u>12330640</u></p> <p>EQPT Name : <u>X ray machine</u></p> <p>Manufacture <u>Allengers</u> Model : _____</p> <p>S. No. <u>2K1410551</u> Dept. <u>X ray</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not switching ON.

Action Taken : checked the machine found Equipment is not switching ON, found that generator unit and tube head is defective enquired OEM for quotation.

Completed Date : 28-6-24 Time : 11:50 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasilraj</u>	<u>28-6-24</u>	<u>9:30 AM</u>	<u>11:50 AM</u>

Customer Remark _____ Completed Pending



Service Engineer Name : Rasilraj
Signature : Rasilraj
Date : 28-6-24
Contact Number : 9995521436

Customer Name : Viji Antony
Signature : [Signature]
Date : [Date]
Contact Number : 982043407
Designation : Medical Officer
Hospital Seal : Community Health Centre PANOOR

Date: 22/07/2024

To,
Cyrix Healthcare (P) Ltd
30/641B, Vandipetta Rd, near SBI Bank,,
Poonithura, Maradu, Ernakulam,
Kerala-682038

Sub: Estimate for Spares of Mars 15-Xray (THQH Panoor)

Dear Sir,

As per the above mentioned subject please find the below estimate.
Kindly go through the same and we are expecting your work order at the earliest.

SI No.	Item description	QTY	Price	Taxes	Amount in Rs.
1	GENERATOR UNIT (FULL SET)	1	450000	12%	504000.00
2	TUBE HEAD (NEW)	1	200000	12%	224000.00
Grand Total					7,28,000.00
(Rupees seven lakh twenty eight thousand only)					

*Above information is not an invoice and only an estimate of services described above.If prices and tax change or additional parts, and labor are required, we will inform you prior to proceeding with the work. Kindly confirm your billing address (as mentioned above) and intimate us if any changes required. This is not included our service charges.

Kindly give an approval for the service in written (Work order) to following address/ Email id, so that we can perform the Job accordingly.

Payment: 100% in advance.

Our GSTIN number: 03AABCA2332H1ZE

For Allengers Medical Systems Ltd



Authorized Signatory
Allengers Medical Systems Ltd, 5B, 5th Floor, Hedge house
Edapally-Palarivattom Road-682025
9946669084/0484-4035626/ker.service@allengers.net

Fwd: Quotation - Generator Unit, X ray Tube

Blesson Jose <zm2.klbemp@cyrix.in>

Mon 2024-07-22 13:41

To: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; avinashtcyrix@gmail.com <avinashtcyrix@gmail.com>

📎 1 attachments (240 KB)

Cyrix Spare Estimate AMSL -Spares THQH Panoor.pdf;

Thanks& Regards,

Blessen Jose
Zonal Manager|North Zone
BEMMP- Kerala
Mobile: 7593847153
Email : zm2.klbemp@cyrix.in

CYRIX HEALTHCARE (P)LTD
30/641B, Petta, Poonithura, Ernakulam - 682038
www.cyrixhealthcare.com

From: RENJITH <ker.service@allengers.net>
Sent: Monday, July 22, 2024 11:30:14 AM
To: Aneesh P B <mklse@cyrix.in>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>
Subject: RE: Quotation - Generator Unit, X ray Tube

Dear Aneesh,

As discussed please refer the attached estimate.

Thanks and regards

RENJITH RAMESH C | Executive, Office Operations | Cochin Branch
Allengers Medical Systems Ltd
5-B, 5th Floor, Hedge House (Pearl Opulence), Edapally-Palarivattom Road, Ernakulam – 682 025
Tel (484) 4035626 | mobile +91 7087029374| Toll free 1800 266 8800
www.allengers.com | ker.service@allengers.net

Confidentiality Note: This e-mail and any attachments are confidential and may be protected by legal privilege. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this e-mail or any attachment is prohibited. If you have received this e-mail in error, Please notify us immediately by returning it to the sender and delete this copy from your system. Thank you for your cooperation.

 Please consider the environment before printing this e-mail!

From: Aneesh P B [mailto:mklse@cyrix.in]
Sent: Friday, July 19, 2024 11:23 AM
To: RENJITH
Cc: Blesson Jose
Subject: Quotation - Generator Unit, X ray Tube

Dear Renjith,

Please share quotation for Generator Unit and X ray Tube head to below site,

THQH Panoor

SN: 2K140615253-X/HF

Location : THQH panoor
Equipment : Allengers Mars 15
SN:

Kind Regards,
Aneesh P B
Deputy Manager - Service
Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

Ph: 9744555073 | M 9072522093 | W www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR PRADESH || TAMIL NADU || KUWAIT || KSA || UAE

INSTALLATION REPORT

AMSL/F/SRV/04
Rev '5' Eff. date 01/02/13

allengers

MEDICAL SYSTEMS LTD.

SCO 212-213-214, Sec. 34-A, Chandigarh- 160 022 (INDIA)
+91-172-3258435, 3012280-84, Fax : +91-172-2621912, 2621913
Email : service@allengers.net
Website : www.allengers.com

Job Card No. _____ Date 18-5-15

Customer Name : Medical officers
Address : Community health centre
Kannur (Kannur)
City : Panoor State : Kerala
E-mail Address : _____
Phone No. (With STD Code) : _____

Branch : Bangalore
Tele No. : 080-32939804
Engineer's Name : Prakash A.
Emp. Code : 2349

M/C Combination	Sr. Nos. of Assemblies	Earthing Condition (Volts) :
<u>Mals - 30 Allengers</u> <u>x-ray machine.</u> M/C Sr. No.	Control <u>2K140615253</u>	PN _____, NE _____, EP _____ Phase to Phase Voltage : RY _____, YB _____, RB _____
	Tube Head <u>400154</u>	
	I.I. <u>-</u>	
	Memory <u>-</u>	
<u>2K140615253-XHF</u>	Humidity Permissible : 80% Rh Available : <u>-</u>	
Voltage Drop During Exposure : _____ (at 75 KVP & 50/200mA)		

Job Performed : Mals 30 Allengers x-ray machine mechanical installation & delivery table kept x-ray room mechanical installation completed

Pending Job, if any & Reason thereof : _____

Breakage/Short shipment, if any : _____

Installation Started on : 12-5-15 Physically Installed on : 18-5-15 Commissioning & Handed over on : _____

Customer's Comments (if any):

Spares : Accessories Handed Over

- Lad Apron Skull Cone Thyroid Guard User Manual
 Dental Cone Mastoid Cone Cassette Holder Other

Spares Used

Sr.No.	Part Description	P. No.	Old ID	New ID	Amount
1					
2					
3					

Job Completed to my Satisfaction & m/c received in Working Order

Date : _____

Time : _____

[Signature]

MEDICAL OFFICER
Community Health Centre

Customer's Sign. & Seal

[Signature]

Engr' gn.

Br. Service Incharge

Service Co-ordinator

Blue = Head Office Copy, Green = Customer Copy, Pink = Accounts Copy, Yellow = Branch Office Copy



	MANUFACTURED DATE	06/2014	 IS (PART-1) CM/L-9129174
	MODEL	MARS 15	
	SR. NO.	SN 2K140615253-X/HF	
	ASSEMBLY REF.	<HF-R/15-1/7239/M3>	
	POWER SUPPLY	400V AC ± 10%, 50/60Hz	
	AERB NO.	12/25	
	RATING	15 KW	
	CATALOGUE NO.	REF 19	
	"Class I Equipment with B Type HV Generator"		
	Mode of Operation - Intermittent		
		IS: 7620	
Manufactured by: Allengers MEDICAL SYSTEMS LIMITED WORKS: BHANKARPUR, MUBARAKPUR ROAD, DERABASSI, DISTT. MOHALE, PUNJAB (INDIA) H.O. S.C.O. 212-214, SEC.34-A, CHANDIGARH (INDIA) TEL:+91-172-3012280-84, FAX: +91-172-2621913 Website: www.allengers.com			
	Our EU Authorised Representative: JMC Medical Spri., Belgium Tel.: +32.2.672.56.44, Fax: +32.2.662.29.92		

TOLL FREE NO: 1800 - 425-7088
 BAR CODE - (8004-800615225) 1330641

MEDICAL OFFICER
 Community Health Centre
 PANOOR

