



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234115

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.ki@cyrix.in

<p>Health Facility : PHC PAMPURUTHY</p> <p>Address : Pampuruthy KANNUR</p> <p>Ph : 9495 296 526</p>	<p>Call Registration Date : 19/07/2024</p> <p>Caller ID : 124210</p> <p>Date of Visit : 20/07/2024</p> <p>Asset No. : 1342342</p> <p>EQPT Name : BP APPARATUS</p> <p>Manufacture : ELKO Model : EL 510</p> <p>S. No. : NA Dept. : OP</p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mainboard, Pressure Sensor and Pump are defective

Action Taken : Checked the equipment and found equipment is not switching ON. Found that Mainboard, Pressure Sensor and pump are defective. Need to replace these spares for further checking.

Completed **Date** : 20/07/2024 **Time** : 11:00 AM **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Sayanth	20/07/2024	9:30 AM	11:00 AM

Customer Remark



Completed Pending
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
PAMPURUTHY
KANNUR - 670601
Dr. Dishu Palach

Service Engineer Name : Sayanth
Signature : *Sayanth*
Date : 20/07/2024
Contact Number : 8606345195


Customer Name : DREPA.P
Signature : *Drepa.P*
Date : 20/7/2024
Contact Number : 9495296526
Designation :
Hospital Seal : NURSING OFFICER U.T.1

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC PAMPURUTHY
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make: ELKO Model: EL 150 SN: NA
4	Equipment ID/ Barcode	1342342 #124210
5	Date of purchase/ Year of manufacture/Installation Date	26-09-2019 (Actual Installation date is 06.11.2019 Stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 19-07-2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking. In software Equipment installed on 26.09.2019. As per the stock book installation date is 06.11.2019 aged up to 04 years 08 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that machine not switching ON, due to Main board, Pressure sensor and pump defective. Need to replace these spares. & also aged upto 4 years & above. The service cost economically not feasible. So recommended for BER as per BER protocol.

Signature of JC BM (NHM)


8/11/2024
CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002


DSEPA.P


NURSING OFFICER 47.1

Dyp

Date: 21/8/2024




Signature of
Superintendent/Medical Officer (i/c)



ELKO®

3

2

1

N

SYS


mmHg

DIA

mmHg

PUL.

/min.



START

STOP

MEM

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)

1342142



Dr. Latha Palanisamy
2/8/2024



PRIMARY HEALTH CENTRE
PAMPURUTHY
P.O. NARATH, KANNUR-670801