



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
231573

ISO 13405 : 2012 & ISO 9001:2000 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 D, Pottu Junction, Poonithur, Kochi - 682 030, Kerala
Ph : 08472 09500 Website : www.cyrilx.com E-mail : bomp.kl@cyrilx.in

<p>Health Facility : <u>PHC KUMARAMPUTHUR</u></p> <p>Address : <u>KUMARAMPUTHUR</u> <u>PALAKKAD</u></p> <p>Ph : <u>9846528955</u></p>	<p>Call Registration Date : <u>23/7/2024</u></p> <p>Caller ID : <u>124956</u></p> <p>Date of Visit : <u>24/7/2024</u></p> <p>Asset No. : <u>0940859</u></p> <p>EQPT Name : <u>Weighing machine</u></p> <p>Manufacture : <u>Samsa electronic</u> Model : <u>NA</u></p> <p>S. No. : <u>NA</u> Dept. : <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Machine not working

Action Taken : checked the machine and found not switching on. Found leadcell and display damaged. Need to replace these spare for further checking

Completed Date : 24/7/2024 Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>SREETHI C</u>	<u>24/7/2024</u>	<u>9:30 AM</u>	<u>11:00 AM</u>

Customer Remark Completed Pending OFFICER
PRIMARY HEALTH CENTRE
KUMARMPUTHUR
PAYYANEDAM P.O
PALAKKAD DISTRICT

<p>Service Engineer Name : <u>SREETHI C</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>24/7/2024</u></p> <p>Contact Number : <u>9072522058</u></p>	<p>Customer Name : <u>SHERA</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>24/7/24</u></p> <p>Contact Number : <u>9645199765</u></p> <p>Designation : <u>NURSING OFFICER</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
--	---




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHIC KUMARAMPUTHUR
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE MAKE : SAMSO MODEL : NA SN : NA
4	Equipment ID/ Barcode	0940859 # 124956
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil (As per stock book installation date 13/02/2018, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	23/07/2024 (toll free)
9	Action taken	Checked the machine found not switching on. Found load cell and Display damaged. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	RS.850/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Assot Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found not switching on. Found load cell and display damaged. As per stock book equipment installed on 13/02/2018. aged up to 6 year 7 months. Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Machine installed on 13/02/2018, aged up to 6 years. Quotation not submitted.


 Signature of Dsika
 06/08/24
 Consultant (Biomedical Engg)
 NHM (Aranyakeralam), Palakkad


 Date: 09/08/2024


 MEDICAL OFFICER
 Signature of Medical Officer
 Superintendent
 PRIMARY HEALTH CENTRE
 KUMARCAMPUTHUR
 PAYYANEDAM. P.O
 PALAKKAD - 678 502

INVENTORY MACHINE


15

Date	Machine Received	To whom Issued	Received	Issued	Balance	Signature
8/1/17	Found on physical verification				1	<i>[Signature]</i>
10/4/18	Transfer to	Unserviceable register		1	NIL	<i>[Signature]</i>
13/2/18	From pharmacy store (Electronic)		1			
15/2/18		To Pre-check room		1	NIL	<i>[Signature]</i>
30/1/18	From pharmacy store (Electronic)		1		1	<i>[Signature]</i>
30/1/18		To NCD Clinic		1	NIL	<i>[Signature]</i>
5/3/22	From pharmacy store. 200kg personal weighing machine Metro enterprises. Kodathi patti Rs - 3,899.99. Cstm No. 32BMPN8572 R 225		1		1 (1)	<i>[Signature]</i>
16/3/22		To NCD Clinic		1	NIL	<i>[Signature]</i>

[Signature]
09/08/24



MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KUMARAMPUTHUR
PAYYANEDAM P.O
PALAKKAD - 678 583



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0940859



OMES CALIBRATION CENTRE

AT: Dept of Engineering & Services
Bioscience - 24/10/2017/2018/2019

MSR No: 09-02-21
18-02-2018
18-02-2018



09/08/2018

MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KUMARAMPUTHUR
PAYYANEDAM PO
PALAKKAD - 678 544