


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	WOMEN AND CHILD HOSPITAL PALAKKAD
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS MAKE: ANITA INDUSTRIES MODEL: ELKOMETER SN: 224957
4	Equipment ID/ Barcode	0912502 #126107
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date is 10/04/2017 As per stock book equipment installed on 14/06/2017 (stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/07/2024 (toll free)
9	Action taken	Checked the BP Apparatus and found that mercury spilled and glass tube broken. Need to replace glass tube and mercury for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Mercury 3 ml = 991/- Glass tube = 112/- Total = 1103/-


13	Asset Value	RS.1323/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	83.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the BP Apparatus and found that mercury spilled and glass tube broken. In ticket master installation date is 10/04/2017, as per stock book Equipment installed on 14/06/2017. aged up to 7 years 1 months. Spare cost is 83.3% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Bp apparatus installed on 10/4/2017, aged up to 7 years. Quotation submitted. Spare cost is 83.3% of asset value. Both criteria of RBER met.

Signature of JC BM (NHM) 

22/08/2024
Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad

Date: 21/8/2024



Signature: 
SUPERINTENDENT
Govt. W&C Hospital
Palakkad
Superintendent Medical Officer (I/c)

11
itai



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. :
231690

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ... <u>Women & children</u>	Call Registration Date : ... <u>29/07/2024</u>
Address ... <u>Hospital</u>	Caller ID : ... <u>126107</u>
... <u>Palakkad</u>	Date of Visit : ... <u>30/07/2024</u>
Ph : ... <u>7306033001</u>	Asset No. : ... <u>0912502</u>
	EQPT Name : ... <u>BP APPARATUS</u>
	Manufacture <u>Anita Industries</u> Model : <u>Elkometer</u>
	S. No. <u>224957</u> Dept. <u>CH Ward</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mercury leak, glass tube broken

Action Taken : Checked the BP Apparatus and found that mercury spilled and glass tube broken. Need to replace glass tube and mercury for further checking.

Completed Date : 30/7/2024 Time : 11:00AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>LIBIL LALIL</u>	<u>30/7/24</u>	<u>10:00AM</u>	<u>11:00AM</u>

Customer Remark Completed Pending

Service Engineer Name : LIBIL LALIL
Signature : [Signature]
Date : 30/07/2024
Contact Number : 8891382686



Customer Name : SHIJA BABY
Signature : [Signature]
Date : 30/7/24
Contact Number : 9846439896
Designation : NURSING OFFICER
Hospital Seal :

SUBMITTED
Govt. W&C Hospr
Palakkad

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL:armedicare07@gmail.com
MOB: 9446057350, 7994654108,



AR MEDICARE

QUOTATION

REF: AR/QTN-113/24-25

Date: 17-08-2024

To,
Cyril Health Care Private Limited
30/641B, Vandipetta Rd, near SBI Bank, Poonithura, Maradu
Ernakulam, - 682038
Kerala 682038

No.	Description	Qty	Unit Price(INR)	TAX	Total(INR)
1.	BP APPARATUS ACCESSORIES MERCURY 3ML	1	835.00	151.2	991.20
2.	GLASS TUBE	1	100.00	12.00	112.00
				Grand Total	1103.20 INR
				Round off	- 0.20 INR
				Net Total	1103 INR

INR One thousand one hundred three Only

Terms and Conditions for sale :

- Validity:** Six months from the date of quotation.
- Payment:** Within 30 days from the date of supply.
- Delivery:** Within 7 days after receipt of the signed purchase order.
- If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- Bank Details:
Account holder's name : AR MEDICARE
Bank Name : State Bank Of India, Thazhechovva
Account No. : 39648455024
IFSC code : SBIN0016861



For AR MEDICARE,
KANNUR



REQUIRED FOR PRICE ENQUIRY BP APPARATUS

2 messages

Fri, 16 Aug, 2024 at 8:23 pm

Akhilesh P <akhileshcyrix@gmail.com>
To: ar medicare <armedicare07@gmail.com>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir

Please share the price and stock availability of bp glass tube and mercury for bp apparatus.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

ar medicare <armedicare07@gmail.com>
To: Akhilesh P <akhileshcyrix@gmail.com>

Sat, 17 Aug, 2024 at 10:58 am

Sir,
PFA



AR MEDICARE

Thanks & Regards

AR MEDICARE
Room No:865,2nd floor
AL-RAHA Arcade
Thazhe Chovva, Kannur
Mob: 9446057350, 7994654108, 8138091108
Email: armedicare07@gmail.com
GSTIN :32BDYPV5030P1ZP

[Quoted text hidden]

BP APPARATUS
ADULT


Qty. Qty. Balance Remarks Old book page no.

1-03-2016	Balance [8-10-09]		1		29
05-06-17	Conderned and reduced from stock as per order NO: B 5196/2016 dt 13-01-17 of Supdt. W & C Palakkad.		1	NIL	
14-06-17	Rd from the Store [From Karunya Pharmacy] Batch NO: 0983369 - Exp. date 5/18 Rs 1650/- [As Local Purchase]		1		
25/09/17	Rd from the store (LP) Digital	7	1		2

[Signature]
SINDA BABY
Nursing OFFICER W & C

[Signature]
SUPERINTENDENT
Govt. W & C Hospital
Palakkad





Shona Babu
Nursing Officer
21/8/2024

21/8/2024
SUPERINTENDENT
Govt. W&C Hospital
Palakkad