


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	THQH CHITTUR
3	Name of Equipment with Make, Model and Serial Number	SUCTION APPARATUS MAKE: SUPREME ENTERPRISE MODEL: SUPREME PLUS SN: NA
4	Equipment ID/ Barcode	0920077 #126119
5	Date of purchase/ Year of manufacture/Installation Date	28/04/2014
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/07/2024 (toll free)
9	Action taken	Checked the machine and found not switching on. Found Problem with motor for suction apparatus, vacuum pressure gauge and tubing set. Requested local vendor for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Motor for suction apparatus = 6000/- Vacuum pressure gauge = 950/- Internal Tubing = 600/- GST 18% = 1359/- Total = 8909/-



13	Asset Value	RS.10200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	87.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached & Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found not switching on. Found Problem with motor for suction apparatus, vacuum pressure gauge and tubing set. Requested local vendor for quotation. Equipment installed on 28/04/2014 aged up to 10 years 3 months. spare cost is 87.3 % both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Suction apparatus installed on 28/04/24 aged up to 10 years old. spare cost is 87.3% of asset value. Quotation ~~is~~ submitted. Both criteria of RBER met. Hence recommended.

Signature of JC BM (NHM) 

21/08/24
Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad

Date:



Signature of 
Superintendent
Taluk Head Quarters Hospital
Chittur-678 101, Palakkad



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. :
231693

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : THQH	Call Registration Date : 29/07/2024
Address : Chittuv Palakkad	Caller ID : 126119
Ph : 9895813180	Date of Visit : 30/07/2024
	Asset No. : 0920077
	EQPT Name : SUCTION APPARATUS
	Manufacture: Supreme Enterprise Model: Supreme plus
	S. No. : NA Dept. : LABOUR OT

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT working
Suction motor, Pressure gauge and tubings damaged

Action Taken : checked the machine and problem found with suction motor, vacuum pressure gauge and tubings set. Need to replace these faulty spares for further checking

Completed Date : 30/7/24 Time : 11:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
LIBIL LALIL	30/7/24	12:00pm	1:00pm

Customer Remark Completed Pending

Service Engineer Name : LIBIL LALIL
Signature :
Date : 30/07/2024
Contact Number : 8891382686

Customer Name : Taluk Head Quarters Hospital
Signature :
Date : 30/7/24
Contact Number : 9895813180
Designation : Superintendent
Hospital Seal :

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL: armedicare07@gmail.com
MOB: 9446057350, 7994654108,



AR MEDICARE

QUOTATION

REF: AR/QTN-112/24-25

Date: 17-08-2024

To,
Cyrix Health Care Private Limited
30/641B, Vandipetta Rd, near SBI Bank, Poonithura, Maradu
Emakulam, - 682038
Kerala 682038

No.	Description	Qty	Unit Price(INR)	Total(INR)
1.	SUCTION MOTOR,	1	6,000.00	6,000.00
2.	VACCUM PREASSURE GUJAGE	1	950.00	950.00
3.	INTERNAL TUBINGS,	1	600.00	600.00
4.	SUCTION BOTTLE	1	2,000.00	2,000.00
			Grand Total	9,550.00 INR
			GST(18%)	1,719.00 INR
			Net Total	11,269.00 INR
INR Eleven Thousand Two Hundred And Sixty Nine Only				

Terms and Conditions for sale :

- 1 **Validity:** Six months from the date of quotation.
- 2 **Payment:** Within 30 days from the date of supply.
- 3 **Delivery:** Within 7 days after receipt of the signed purchase order.
- 4 If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- 5 **Bank Details:**

Account holder's name : AR MEDICARE
Bank Name : State Bank Of India, Thazhechovva
Account No. : 39648455024
IFSC code : SBIN0016861



For AR MEDICARE,
KANNUR



REQUIRED QUOTATION 0920077 # 126119 SUCTION APPARATUS @ THQH CHITTUR, PALAKKAD

2 messages

Akhilesh P <akhileshcyrix@gmail.com>
To: ar medicare <armedicare07@gmail.com>
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Fri, 16 Aug, 2024 at 8:22 pm

Dear sir
Please share the price and stock availability of suction motor, Guage, internal tubings,suction bottle for suction apparatus @ THQH CHITTUR

Details are furnished below

Equipment Name: suction apparatus

Hospital Name: THQH CHITTUR

Make : SUPREME

Barcode : 0920077

Ticket id: 126119

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

ar medicare <armedicare07@gmail.com>
To: Akhilesh P <akhileshcyrix@gmail.com>

Sat, 17 Aug, 2024 at 10:50 am

Sir,
PFA



AR MEDICARE

Thanks & Regards

AR MEDICARE
Room No:865,2nd floor
AL-RAHA Arcade
Thazhe Chovva, Kannur
Mob: 9446057350, 7994654108, 8138091108
Email: armedicare07@gmail.com
GSTIN :32BDYPV5030P1ZP

[Quoted text hidden]

SUCTION APPARATUS

	Rd	Issued	Used	Condn	Balance
28/4/14 Rd from Stone	2				2
28/4/14 ISSUED TO USE		2			NIL
14/1/19 BF from Reg code AN Vol code 1 Rd date 24/7/13 page no: 143 2nd pressing Rs: 8019/- page no: 93 (extra)	1				
14/1/19 ISSUED TO USE					
01.08.2019 Rd from store model RC 02-151 monarch surgical Industries Serial No: 201907A Installation dt 01.08.19 to 31/07/2020 warranty period 31/07/2020 to 30/7/2023	1				
11/8/19 ISSUED TO USE		1			
26/2/20 portable suction apparatus T.E.A (Neonatal) Rd from Stone Cost Rs 6500/-	1				1
07/2/20 ISSUED TO USE		1			NIL
10/3/22 Shifted to unrepairable Regy - (1-24/7/13 2-24/4/14)			3		
20/6/22 Rd from Stone Mini portable Suction Apparatus B No 1 Purchased from Sun Surgicals and Lth Tiruvandram. invoice 20-23/9/15 dated 8/6/22 invoice amount 56656	1				



[Signature]
SUPERINTENDENT
Taluk Head Quarters Hospital
Chittur-678 101, Palakkad



AP
11/08/24
SUPERINTENDENT
Taluk Head Quarters Hospital
Chittur-678 101, Palakkad