


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	DH KANHANGAD
3	Name of Equipment with Make, Model and Serial Number	MULTIPARA MONITOR MAKE: MINDRAY MODEL: MEC 1000 S/N: AQ11149282
4	Equipment ID & Barcode	#124811 -1410398
5	Date of purchase / Year of manufacture /Installation Date	21-03-2011
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registrationof complaint through email/ Toll free)	23-07-2024 (Toll free)
9	Action taken	Checked the machine found that monitor not switching on. found problem with Main board, Power board and rear cover assembly. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	fully damaged.
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Power board (Lead-Acid) : 20160/- Main board: 39000/- Rear cover assembly: 7500/- Grand Total: 66660/- GST18% :11998.80/- Net Total: 78659/-

13	Asset Value	₹70000
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	112.37%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine found that monitor not switching on. Found that Main board, power board and rear cover assembly are defective. Machine installed on 21-03-2011 and aged up to 13 years and 4 months. Spare cost is 112.37% of asset value. Both criteria for RBER met. As per the tender clause 5.3.14.1, We recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JAYESH KRISHNAN 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Checked the machine and found Problem with main board, power board, and rear cover assembly. Equipment installed on 21/03/2011; spare cost is 112.37% of asset value. cyrix service report & OEM quotation attached. so recommending the equipment for BER.


 Signature of JC BM (NHM) Jr. CONSULTANT
 Biomedical Engineering
 Arogyakeralam (NHM)
 Kasaragod/Kanhangad

Machine very old
 Repairing charge is very high since this machine

Date 23/8/24


 Seal


 Signature of
 Superintendent / Medical Officer (V)

സുപ്രണ്ടു
 ജിനറൽ മെഡിക്കൽ ഓഫീസർ (വ)
 കാസറഗോഡ് ജില്ല



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 198320

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>DH Kanhangad</u></p> <p>Address : <u>Kanhangad</u> <u>Kasaragod</u></p> <p>Ph : <u>9961088830</u></p>	<p>Call Registration Date : <u>23/07/24</u></p> <p>Caller ID : <u>124811</u></p> <p>Date of Visit : <u>25/07/24</u></p> <p>Asset No. : <u>1410398</u></p> <p>EQPT Name : <u>Multipara monitor</u></p> <p>Manufacture : <u>mindray</u> Model : <u>mec 1000</u></p> <p>S. No. <u>AQ11149282</u> Dept. <u>mmw</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not getting on.

Action Taken : Checked the unit, found problems with main board, power board and rear cover assembly. Need to replace these spares for further servicing of the equipment.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrrix Engineer	Date	Start Time	End Time
<u>Jayesh Krishnan</u>	<u>25/07/24</u>	<u>2.30pm</u>	<u>3.00pm</u>

Customer Remark Completed Pending



<p>Service Engineer Name : <u>Jayesh Krishnan</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>25/07/24</u></p> <p>Contact Number : <u>7593847156</u></p>	<p>Customer Name : <u>Rebhaile</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>25/7/24</u></p> <p>Contact Number : <u>9961088830</u></p> <p>Designation : <u>Sno</u></p> <p>Hospital Seal :</p>
---	--

സമൂഹാരോഗ
കമ്മ്യൂണിറ്റി കാന്തോപ്പാട്
കമ്മ്യൂണിറ്റി കാന്തോപ്പാട്

Sree Gokulam Healthcare Pvt. Ltd
Dhanya Complex, Sastha Temple Road
Ernakulam North , Kochi - 682018, Kerala. Ph: 0484 -
2402408, Fax: 0484 - 2402418



Quotation

REF:SGH/2024-25/2474446

Date: 16-Aug-24

To,
Cyrix Health Care Private Limited
FIRST FLOOR,PETTA JUNCTION 30/641B
POONITHURA
ernakulam - 682038
Kerala, India

No.	Item Code	Description	Qty	Unit Price(INR)	Total(INR)
1		Power board(Lead-Acid)	1	20,160.00	20,160.00
2		Main Board	1	39,000.00	39,000.00
3		Rear cover assembly	1	7,500.00	7,500.00
				Grand Total	66,660.00 INR
				GST(18%)	11,998.80 INR
				Net Total	78,659.00 INR


INR Seventy Eight thousands Six hundred and Fifty Nine Only

Terms and Conditions

- Validity:** One month from the date of quotation
- Payment:** 100% in advance along with the purchase order
- Delivery:** Within 60 days after receipt of the signed purchase order and payment
- If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- Bank Details**
Name of Organisation : Sree Gokulam Healthcare Pvt Ltd.
Name of Bank : Federal Bank
Branch : Kaloor,Ernakulam
Account Number : 14215500001738
IFSC Code : FDRL0001421

Dineshan C
Managing Director





Jayesh Krishnan <jayeshcyrix@gmail.com>

PRICE ENQUIRY 1410398

14 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Fri, Jul 26, 2024 at 9:23 AM

To: service gokulamhealthcare <service@gokulamhealthcare.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Please share the price and availability of SMPS, main board and outer cover for Multipara monitor.

Make: Mindray

model: Mec 1000

S/N: AQ11149282

—
Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sat, Jul 27, 2024 at 12:04 PM

To: service gokulamhealthcare <service@gokulamhealthcare.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>, Anjitha T L

<purchasecyrix2@gmail.com>

Dear sir,

Kindly share the quotation asp

Thanks & Regards

Sarang.k.m
Divisional Manager | North Zone
Bemp-Kerala
Mobile : +91 7593847150
Email : dm1.klbemp@cyrix.in

CYRIX HEALTHCARE (P) LTD**30/641B | Petta | Punithura | Kochi | Kerala- 682038**www.cyrixhealthcare.com

From: Jayesh Krishnan <jayeshcyrix@gmail.com>

Sent: 26 July 2024 09:23

To: service gokulamhealthcare <service@gokulamhealthcare.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>
Subject: PRICE ENQUIRY 1410398

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, Jul 30, 2024 at 11:25 PM

To: service gokulamhealthcare <service@gokulamhealthcare.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, Anjitha T L <purchasecyrix2@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Kindly update urgently.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Thu, Aug 1, 2024 at 11:57 AM

To: service gokulamhealthcare <service@gokulamhealthcare.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, Anjitha T L <purchasecyrix2@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,

Kindly update.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Fri, Aug 2, 2024 at 3:07 PM

To: Anjitha T L <purchasecyrix2@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Anjitha,

Kindly collect quotation for the spares in the trailing mail.

[Quoted text hidden]

Anjitha T L <purchasecyrix2@gmail.com>

Fri, Aug 2, 2024 at 3:09 PM

To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

As per the discussion with the vendor they will share quotations by tomorrow.

[Quoted text hidden]

-

Thanks & Regards

Anjitha T L

Sr. Purchase Executive

PH:- 9947719111

purchasecyrix2@gmail.com

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Aug 5, 2024 at 9:56 AM

To: Anjitha T L <purchasecyrix2@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Anjitha,

Kindly update.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, Aug 6, 2024 at 9:32 AM

To: Anjitha T L <purchasecyrix2@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Reminder...


[Quoted text hidden]

service gokulamhealthcare <service@gokulamhealthcare.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Tue, Aug 6, 2024 at 10:03 AM

Dear Sir,
Kindly see the attached quotation.

[Quoted text hidden]

 **Quot_Cyrix Health Care Private Limited_REF-SGH-2024-25-2474423.pdf**
98K

Jayesh Krishnan <jayeshcyrix@gmail.com>
To: service gokulamhealthcare <service@gokulamhealthcare.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Anjitha T L <anjithacyrix@gmail.com>

Wed, Aug 7, 2024 at 8:57 AM

Dear sir,
We have requested quotation for power board, main board and outer cover for Multipara monitor, only received power board quotation. Kindly send updated quotation at the earliest.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>
To: service gokulamhealthcare <service@gokulamhealthcare.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Anjitha T L <anjithacyrix@gmail.com>

Fri, Aug 9, 2024 at 8:38 AM

Dear sir,
Kindly update

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>
To: service gokulamhealthcare <service@gokulamhealthcare.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Anjitha T L <anjithacyrix@gmail.com>

Mon, Aug 12, 2024 at 9:40 AM

Sir,
Kindly update urgently

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>
To: service gokulamhealthcare <service@gokulamhealthcare.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Anjitha T L <anjithacyrix@gmail.com>

Fri, Aug 16, 2024 at 10:45 AM

Reminder


[Quoted text hidden]


service gokulamhealthcare <service@gokulamhealthcare.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>, Anjitha T L <anjithacyrix@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>


Fri, Aug 16, 2024 at 4:18 PM

Dear Sir,
Kindly see the attached quotation.

On Fri, Jul 26, 2024 at 9:23 AM Jayesh Krishnan <jayeshcyrix@gmail.com> wrote:
[Quoted text hidden]

 **Quot_Cyrix Health Care Private Limited_REF-SGH-2024-25-2474446.pdf**
113K





 സൂപ്രണ്ട്
 ജില്ലാ ആരോഗ്യ
 കാസറഗോഡ് ജില്ല
 കാസറഗോഡ് ജില്ല

Government District Hospital, Kasaragod - Kanhangad

STOCK REGISTER

59

MULTIPARAMETER

Date	Name of articles	Received Qty.	Issued Qty.	Balance Qty.	Signature of receiver	Remarks
28/11	Received from store Issued to use Page No: NRHM Book - Page 40 - in store 10 - Mindray Model - mee 1000 S. No: 11149282	1	1	1	<u>Chintu</u>	
30/3/14	Transfer in from FSU C-bck Register page no. 60 (C Mindray)	1		2	<u>Ch</u>	സൂപ്പർ ജില്ലാ ആശുപത്രി കാസർഗോട് കാനറതോട് ജില്ല
18/6/14	Multiparameter i Capnograph Accessories: - Reusable adult S lead - 2 no + 2 ECG cable set 2 no + 2 NIBP hose 1 + 1 no. obese adult cuff 1 + 1 no. adult NBP cuff 1 + 1 no. child NBP cuff 1 no + 1 no. Infant NBP cuff 1 no + 1 no. SpO2 Extremity cable 1 no + 1 no. Reusable adult SpO2 finger probe - 1 no + 1 no. Disposable non-invasive SpO2 probe - 1 no + 1 no. side stream ECG Disp kit adult 25ml 2 no. side stream ECG Disp kit Pediatric + neonatal 2 no + 2 no. Rs - 5,43,152/- (2) Warranty - 10/6/2014 to 9/6/2017 - (3 Yrs) contact no. 099995004523 - (Bijoy P) [Instrument / Equipment Vol. 2 - P. 53].	2		4	<u>Ch</u>	
29/12/19	Multiparameter (Comen) Storage system & Services (E firm) cost 110500/- MRP no - 108	0/1		5	<u>Ch</u>	Ph. no 94960189.

