


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	DH PALAKKAD
3	Name of Equipment with Make, Model and Serial Number	VENTILATOR MAKE : MAQUET MODEL : SERVO S SN : 22232
4	Equipment ID/ Barcode	0911531 #123146
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date -18/05/2012 As per stock book installation date is 26/05/2012 (stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	15/07/2024 (toll free)
9	Action taken	Checked the machine and found restart ventilation error showing. Found Problem with control cable, battery module, circuit board pc 1772, compressor with motor damaged. Enquire oem for quotation.
10	Present status of the equipment(Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Control cable = 41206/- Battery module = 82574/- Circuit board PC 1772 = 286810/- Compressor with motor = 312547/- Total tax = 94204.24/- Total = 817341.24/-

13	Asset Value	RS.940000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	86.95%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found restart ventilation error showing. Found Problem with control cable, battery module, circuit board pc 1772, compressor with motor. In Ticket master installation date 18/05/2012 as per stock book installation date is 26/05/2012 aged up to 12 year 2 months. Spare cost is 86.95% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Ventilator installed on 26/05/2012 aged up to 12 years. spare cost is Rs 86.95% of asset value. Quotation submitted.


 21/08/24
 Signature of JC BMNH
 Consultant (Biomedical Engg)
 NHM (Arogyakeralam), Palakkad

Date:




 Deputy Director of Health Services
 & Superintendent.
 Signature of District Hospital, Palakkad
 Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 173659

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>District Hospital</u></p> <p>Address <u>Palakkad</u> <u>Palakkad</u></p> <p>Ph : <u>9074868652</u></p>	<p>Call Registration Date : <u>15/07/2024</u></p> <p>Caller ID : <u># 123146</u></p> <p>Date of Visit : <u>16/07/2024</u></p> <p>Asset No. : <u>0911531</u></p> <p>EQPT Name : <u>Ventilator</u></p> <p>Manufacture <u>Maguel</u> Model : <u>Servo S</u></p> <p>S. No. <u>22232</u> Dept. <u>ICU</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Restart Ventilation Error

Action Taken : Checked the Ventilator Found Restart Ventilation Error Showing. Found problem with Control Cable, Battery module, PC1772 Cir.cwt board, Compressor with motor damaged. Engive OEM For quotation.

Completed Date : 16/07/24 Time : 10.00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Akhilesh.P</u>	<u>16/07/24</u>	<u>10.00 AM</u>	<u>10.30 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Akhilesh.P
Signature : [Signature]
Date : 16/07/24
Contact Number : 7593897154



Customer Name : [Signature]
Signature : [Signature]
Date : 16/07/24
Designation : Deputy Director of Health Service & Superintendent, District Hospital, Palakkad.

Principal Place of Business:

Getinge Medical India Pvt. Ltd.

Airport Road

204, Fulcrum "B" Wing

204, 2nd Floor

Near Hyatt Hotel

Mumbai-Andheri East 400099

State Code & Name: 27 MAHARASHTRA**GSTIN / Unique ID:** 27AADCM8545J1Z2

Cyrix Healthcare Pvt Ltd

1st Floor, Petta Junction 30/641B, Poonithur,

Ernakulam 682038

GST No: 32AAFCC2499H2ZM

Quotation CRM**8200091544****2024/08/16****Customer:** 2226400**Purchase order no.:** Quotation submitted**Purchase order date:** 2024/08/16**Person in charge:** Prasada Reddy**Phone:** 9849042409**Email:** prasada.reddy@getinge.com**Sales employee:** Prasada Reddy

Validity date: 2024/08/16 - 2024/09/30

Terms of payment: 100% Prepayment

Terms of delivery: DAP Ernakulam

Subject:

Estimate for Spare/Consumables of Getinge Ventilator unit.

(Servo S/I No:22232)

in Favour of M/s Cyrix Healthcare Pvt. Ltd, Ernakulam, Kerala (C/f. DH

PALAKKAD).

Payment Terms:100% Advance Payment along with PO.

Note: Tax Collection at Source (TCS) on certain goods introduced vide

Finance Act, 2020

Special Note: Purchase Order and payment to be made in favour of "Getinge Medical India Pvt Ltd" only

Item	Material Description	Quantity	Unit price INR	Total price INR
10	6651298 Control cable *CSA* HSN Code: 85444999	1 EA		
			41,206.00	41,206.00
	I_GST	18%		7,417.08
20	6487180 Battery Module HSN Code: 85075000	2 EA		
			41,287.00	82,574.00
	I_GST	18%		14,863.32

Quotation CRM/Date:
8200091544 / 2024/08/16

GETINGE 

Item	Material Description	Quantity	Unit price INR	Total price INR
30	6696957 Circuit board PC1772 incl. 6467620 HSN Code: 90192090	1 EA		
			286,810.00	286,810.00
	I_GST	12%		34,417.20
40	06545367 Compressor with Motor 230V/60Hz HSN Code: 90194578	1 EA		
			312,547.00	312,547.00
	I_GST	12%		37506.64
	Item sum			723,137.00
	Total Tax			94,204.24
	Total price			817,341.24

Total Amount	INR	SEVEN LACK TWENTY-THREE THOUSAND ONE HUNDRED THIRTY SEVEN
Tax Total	INR	NINETY-FOUR THOUSAND TWO HUNDRED - FOUR TWENTY- FOUR
Total amount with tax	INR	EIGHT LACK SEVENTEEN THOUSAND THREE HUNDRED FORTY-ONE - TWENTY- FOUR

Delivery Period: 8-12 weeks from the date of receipt of your confirmed order.

Warranty terms:

- (a) 3 months warranty on spare parts against manufacturing defects.
- (b) No warranty on consumables & disposals.
- (c) Warranty void if equipment is handles by unauthorised personnel.

Duties & Taxes:

The Duties & Taxes mentioned in the quote are at the current GST rates. Any change in Custom Duty rates, Statutory Taxes added etc. recovery whereof from customers is not prohibited under the provisions of the relevant law shall be to purchaser's account, as applicable as the time of supply.

Computer generated hence no signature required.



Purchase Order -KL 24 25 101527 -Getinge Medical India Pvt Ltd - 0911531 -DH PALAKKAD

5 messages

Anjitha T L <purchasecyrix2@gmail.com>

Tue, 6 Aug, 2024 at 11:50 am

To: Medicuits Service <medicuits.service@gmail.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Sir,

Please find the attached Purchase Order and do the needful. Kindly send a confirmation mail with details of material delivery (Docket details) along with the invoice Copy mentioning our PO No. (For reference) with Seal and Signature.

EDD as per payment term: 12/08/24

For Queries and Clarifications

1. Technical: dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in
2. Accounts: paymentcyrix@gmail.com
3. Logistics: logistics@cyrix.in
4. PO Clarification: ampurchase@cyrix.in/purchasemgr@cyrix.in
5. Delivery address: **Cyrix Healthcare PVT LTD**

Methanam Road
Door No: 312/a Eloor North,
Kerala - 683501- India
Nikesh V
Mob: 9072522014

@ Cyrix Accounts

Payment Term: 100% ADV

-

Thanks & Regards

Anjitha T L

Sr. Purchase Executive
PH:- 9947719111
purchasecyrix2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Cyrix Payments <paymentcyrix@gmail.com> Thu, 8 Aug, 2024 at 11:11 am
To: Medicuits Service <medicuits.service@gmail.com>
Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Sir/Mam,

Please note that we had transferred Rs.23200 /- as advance against PO 101527 after deducting TDS Rs.400/-

Getinge Medical India Pvt Ltd	72833018	23,200.00	AXISP00526296404	PAID	07-AUG-24
-------------------------------	----------	-----------	------------------	------	-----------

[Quoted text hidden]

Thanks & Regards,

Tincy T Jose
Accounts Department.
Cyrix Healthcare Pvt Ltd
Office : 30/641 B, First Floor, Petta Junction, Poonithura, Cochin, Kerala - 682 038
Website:- www.cyrixhealthcare.com
Mob : 7593847123

Izrath K K <purchasecyrix5@gmail.com> Fri, 9 Aug, 2024 at 4:55 pm
To: Medicuits Service <medicuits.service@gmail.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Akhilesh P <akhileshcyrix@gmail.com>, Anjitha T L <purchasecyrix2@gmail.com>

Hi,

Kindly arrange the service as soon as possible.

Thanks & Regards
Izrath KK
Jr Purchase Executive
PH: 8848002238

On Tue, Aug 6, 2024 at 11:50 AM Anjitha T L <purchasecyrix2@gmail.com> wrote:
[Quoted text hidden]

Medicuits Service <medicuits.service@gmail.com> Fri, 16 Aug, 2024 at 4:29 pm
To: Izrath K K <purchasecyrix5@gmail.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix

Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Akhilesh P <akhileshcyrix@gmail.com>, Anjitha T L <purchasecyrix2@gmail.com>

Dear Sir/Madam,

With reference to the above subject, please find the enclosed Service quotation and Service Call Report of Servo S ventilator/ SI No: 22232 at District Hospital, Palakkad.

With regards
[Quoted text hidden]

ANITHA .V .R

Bio Medical Engineer

* Service Help Line : 93 4 93 24 365 *

MEDICUITS Inc.,

Ponnambalam Complex , Kumarapuram ,

Medical College P.O , Thiruvanthapuram - 695 011

Phone : 0 471 2 55 28 73

Izrath K K <purchasecyrix5@gmail.com>

Fri, 16 Aug, 2024 at 4:55 pm

To: Medicuits Service <medicuits.service@gmail.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Akhilesh P <akhileshcyrix@gmail.com>, Anjitha T L <purchasecyrix2@gmail.com>

Hi,

Kindly share the tax invoice against the po.

Thanks & Regards

Izrath KK

Jr Purchase Executive

PH: 8848002238


[Quoted text hidden]

one Ventilator 1
 multi use

B, P: 179
 Ventilator

Palakkad Rd
 26/5/12
 State
 Ventilator C maguet medical india

servo-s - serial no: 2222
 94,650/-



Vayusu

Deputy Director of Health Services
 & Superintendent,
 District Hospital, Palakkad

1) maguet ventilator's accessories.

- (1) NIV mask RS-4350/-
- (2) Re-usable skin ventilator circuit - double water trap RS-4650/-

Issued to use in ICU

Continue usage and repair.

Palakkad Rd
 State

Model & make: Entrol AT

2) Air Liquide medical system

Sl. No: 42627 23-15-12

All accessories - see list

Chk list enclosed

Mr. 90192090.

Rate - 9,63,250.00


GST - 2.12%

= 41,85,90/-

₹ = 10,78,840/-

9847049907

21/12/12 issued to use in ICU 1 NIL 2



Vijayakumar

**Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad.**