


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	PHC NARIPATTA
3	Name of Equipment with Make, Model and Serial Number	Equipment: WEIGHING MACHINE Make: SAMSO Model: NA SN : NA
4	Equipment ID & Barcode	127042 & 1141649
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date is 02/05/2018 (As per stock book installation date is 14/01/2021, stock book attached for reference)
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	02-08-2024
9	Action taken	Checked and found main board, load cell and battery are defective. Need to replace these spares for further checking of working condition of the equipment
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	1600/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found main board, load cell and battery are defective. Need to replace these spares for further checking of working condition of the equipment. Ticket master installation date is 02/05/2018 (As per stock book installation date is 14/01/2021, stock book attached for reference) aged up to 03 years 6 months. Quotation not submitted since spares are not available in the market So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JINEESH PP 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked and found that main board, load cell and battery are defective. Machine installed on 2/5/2018. Quotation is not submitted. Spares are not available in the market. So recommended for condemnation.

Signature of JC BM (NHM)

Consultant Bio Medical Engineer
National Health Mission
Kozhikode

Date 22/8/21



Signature of  Dr. Pradeesh

Superintendent / Medical Officer (I/C)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234962

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

<p>Health Facility <u>PHC</u></p> <p>Address <u>NARIPPATTA</u> <u>KOZHIKODE</u></p> <p>Ph : <u>9072278692</u></p>	<p>Call Registration Date : <u>218/24</u></p> <p>Caller ID : <u>127042</u></p> <p>Date of Visit : <u>318/24</u></p> <p>Asset No. : <u>1141649</u></p> <p>EQPT Name : <u>Weighing Machine</u></p> <p>Manufacture <u>SANSO</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Error showing on display.

Action Taken : Checked and found the issue man board, load cell and batteries are defective need to replace these spare for further checking of working condition of equipment.

Completed Date : 318/24 Time : 1:00pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>ARUN KK</u>	<u>318/24</u>	<u>11:00am</u>	<u>1:00pm</u>

Customer Remark Completed Pending



Service Engineer Name : ARUN KK

Signature : [Signature]

Date : 318/24

Contact Number : 7593847184.

Customer Name : Deepa C. Bharan

Signature : [Signature]

Date : 318/24


Contact Number : N/A

Designation : N/A


Hospital Seal : 9072278692

STOCK BOOK OF ABDRAM MISSION ITEMS


Name of article WEIGHING MACHINE DIGITAL ADULT-SCALE 150KG

Date	From whom received or to whom issued	Receipt	Issued	Balance	Batch No Exp. Date Mfg. Date	Price per Unit	Remarks
	Brought Forward						
24.1.19	IN No 4050515.3-19 KORUNYA MEDICINE 02			02	111 12/2028 11-2019	3411.00	
	DEPT ERNAKULAM						
	[size: 30.8x30.8x2.8cm Capacity 150kg x 100gm. Weight: 1.10kg. mfg. Doctor Belirem PUL LAK]						
					 MEDICAL OFFICER PRIMARY HEALTH CENTRE NAKUPATTA		
14.1.2022	Issued to Ranjya P.C Staff nurse [SSR page no. 39]		01	01 NIL			
12-12-22	issued to Ranjya's		01	NIL			
22/12/22	Dr Pradish kumar FHC Nampyall						




മെഡിക്കൽ ഓഫീസർ
കുടുംബാരോഗ്യകേന്ദ്രം
നക്പട്ട




 Dr. Perushan
 ഫാസിക്കൽ ഓഫീസർ
 കൃഷ്ണമേലൂർ
 നഗർ