


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	CHC THIRUVANGOOR
3	Name of Equipment with Make, Model and Serial Number	Equipment : NEBULIZER Make: COMFO CARE Model: N8002 SN : NA
4	Equipment ID & Barcode	128255 & 1130328
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date - Nil As per the stock book installation date is 15-07-2015
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	08-08-2024
9	Action taken	Checked and found compressor motor with piston defective. Need to replace these spares for further checking of working condition of the equipment
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	1353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Need to replace these spares for further checking of working condition of the equipment. In ticket master installation date is Nil, As per the stock book installation date is 15-07-2015 aged up to 09 years 1 month. Quotation not submitted since spares are not available in the market So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	JINEESH PP 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked and found compressor motor with piston defective. Machine installed on 15/7/2015. Quotation is not received. Spares are not available in the market. So recommended for condemnation.

Signature of JC BM (NHM) 

Consultant Bio Medical Engineer
National Health Mission
Kozhikode

Date



Seal



Signature of
സൂപ്രണ്ടൻ്റ് മെഡിക്കൽ ഓഫീസർ
ആരോഗ്യ മിഷൻ
കോഴിക്കോട് - 673 304
Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 213740

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.ki@cyrix.in

<p>Health Facility <u>CHC Thiruvangoor</u>.....</p> <p>Address <u>Thiruvangoor</u>..... <u>Kozhikode</u>.....</p> <p>Ph : <u>9074198489</u>.....</p>	<p>Call Registration Date : <u>8/8/24</u>.....</p> <p>Caller ID : <u>128255</u>.....</p> <p>Date of Visit : <u>9/8/24</u>.....</p> <p>Asset No. : <u>1130228</u>.....</p> <p>EQPT Name : <u>Nebulizer</u>.....</p> <p>Manufacture <u>Comfo care</u>..... Model : <u>N8002</u>.....</p> <p>S. No. <u>NA</u>..... Dept. <u>Nursing Station</u>.....</p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

Fumes not getting.

Action Taken : Checked & found compressor motor with piston defective.
..... Need to replace these spares for further checking ab working
..... condition ab the equipment......

Completed Date : 9/8/24..... Time : 11:30 AM..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
	<u>9/8/24</u>	<u>9:30 AM</u>	<u>11:30 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Akhil Raj

Signature :

Date : 9/8/24

Contact Number : 8943677676

Customer Name : Shreekrishna


Signature :

Date : 9/8/24

Contact Number : 9074198489

Designation : NA

Hospital Seal :




സാമൂഹിക ആരോഗ്യ കേന്ദ്രം
 തിരുവങ്ങൂർ - 673 304

Name of article Nebulizer

Date	No and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought Forward			1		(10 nurse rooms)
21/6/07							
21/6/10		Received from the store	1		2		(10 nurse rooms)
24/2/11		Received from the store	2		4		
24/2/11		Issued to OP	1	2	2		
23/2/13		Rd from the store	2		6		
21/8/13		Rd Issued to OP		2			
11/11/14		Rd from the store	1		7+1=8		
		Issued to OP		1			
11/1/15		Received from the store	1		8+1=9		
16/7/15		Issued to OP		1			
15/07/15		Received from the store	1		9+1=10		
		Issued to OP		1			
08/09/15		Received from the store	1		10+1=11		
21/2/20		Recd from the store		2 NO.	12+1=13		Genio Spre
		Per/O - 156250 (B.S. 00 - 2)					
		Pr. out to Unserviceable Register Page no. 82					
		Candent as per order no. 4162/22 dt 20/12/21 no. Genio					
28/8/20	3	Pr. out to CFHC Annultha		3 NO.			
23/9/20		Pr. out to CFHC Annultha		2	6 NO.		
15/6/22		CS Rd pack from CFHC Annultha. Per/O No.					
18/6/22		Pr. out to Unserviceable Register Page no. 82 (9 No.)					
		9 Candent as per order no. 498/2023 dt 19/5/2023					
					2 NO.		Genio




 MEDICAL OFFICER
 COMMUNITY HEALTH CENTRE
 THIRUVANGOOOR - 673 304

Compressor Nebulizer

Model No.: N8002

Input : AC 230V/50Hz

Lot: 12/2014

Mode of operation-continuous

No AP/PG IPX0



0120

CYRIX

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)
1130328



COMFO - CARE

Shobha

മെഡിക്കൽ സൗകര്യങ്ങൾ
മുഖ്യകാര്യാലയം
തിരുവനന്തപുരം - 695 004