


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	CHC MELADY
3	Name of Equipment with Make, Model and Serial Number	EQUIPMENT: OXYGEN CONCENTRATOR MAKE: SHENYANG CANTA MEDICAL TECH CO.LTD MODEL: V8-WN-NS SN: 2448902042721
4	Equipment ID & Barcode	123797 & 1130189
5	Date of purchase / Year of manufacture /Installation Date	In Ticket master installation date - Nil As per the stock entry installation date is 25/05/2021
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	22-06-2024
9	Action taken	Checked and found compressor, main board and sieve bed defective. Need to replace these spares for further checking working of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and NHM, KMSCL Bme Mail communication
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. As per the stock entry Installation date is 25-05-2021. Unfortunately, the supplier information is unavailable from end user as the equipment was donated during covid pandemic. Due to the lack of spare part information and supplier/ manufacture details (china make equipment) spares are not available, so we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	JINEESH PP 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked and found that Compressors, sieve bed and main board are damaged. Machine installed on 25/05/2021. Supplies <sup>manufactures</sup> details are not available at institution. Spare not available in the market. Recommended for ~~condemning~~ <sup>condemning</sup>


Signature of JC BM (NHM)

Consultant Bio Medical Engineer

National Health Mission  
Kozhikode

Date \_\_\_\_\_ Seal \_\_\_\_\_

Signature of  
Superintendent / Medical Officer (i/c)

  
MEDICAL OFFICER  
C.H CENTRE MELADY  
PO THIKYODI KOZHICKODE





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 213743

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>CHC melady</u> .....</p> <p>Address ..... <u>melady</u> .....</p> <p style="padding-left: 40px;"><u>Kozhikode</u></p> <p>Ph : ..... <u>9042 197130</u> .....</p>	<p>Call Registration Date : ..... <u>18/7/24</u> .....</p> <p>Caller ID : ..... <u>123797</u> .....</p> <p>Date of Visit : ..... <u>19/7/24</u> .....</p> <p>Asset No. : ..... <u>1130189</u> .....</p> <p>EQPT Name : <u>oxygen concentrator</u> .....</p> <p style="padding-left: 100px;"><u>medical Tech CO LTD</u></p> <p>Manufacture <u>Shenyang Consta</u> Model : <u>VS-WN-MS</u></p> <p>S. No. <u>244890204272</u> Dept. <u>OP</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** .....

oxygen not getting

**Action Taken :** checked & found that compressor, main board & sieve bed defective. need to replace these spares for further checking working at the equipment.

Completed  Date : 19/7/24 Time : 1:30pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Akhilraj</u>	<u>19/7/24</u>	<u>10:30AM</u>	<u>1:30pm</u>

**Customer Remark** Completed  Pending

**Service Engineer Name :** Akhilraj

**Signature :**

**Date :** 19/7/24

**Contact Number :** 8943677676

**Customer Name :** Dr Veena V.A.

**Signature :**

**Date :** 19/7/24

**Contact Number :** 9847299500

**Designation :** Medical Officer

**Hospital Seal :**





**Regarding Supplier details required for service purpose-Oxygen Concentrator -CHC MELADY,  
KOZHICODE - 123797- 1130189**

3 messages

Fri, 26 Jul, 2024 at 12:29 pm

**Jineesh P P** <jineeshcyrix@gmail.com>  
To: BME Kozhikode <bmekkd1@gmail.com>  
Cc: kmsclbmenorth1@gmail.com, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Madam,

Attended the break down call of o2 concentrator at chc melady and found spares are required for rectification and noted the supplier details not available from the site and so unable to rectify this issue. So here i am attaching available stock details and photos for your reference and requesting to provide supplier details as soon as possible

Barcode no : 1130189  
Ticket Id : 123797

Make : SHENYANG CANTA MEDICAL TECH CO.LTD  
Model : V8-WN-NS  
SN: 2448902042721

Thanks & Regards,

Jineesh PP  
District Incharge - Kozhikode  
Mob-7593847163  
jineeshcyrix@gmail.com  
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

**Varsha EK** <kmsclbmenorth1@gmail.com>  
To: Jineesh P P <jineeshcyrix@gmail.com>

Sat, 27 Jul, 2024 at 4:13 pm

The supply details for the above mentioned equipment were not available with us.  
rgrds

Varsha EK  
Regional Biomedical Engineer  
KMSCL  
[Quoted text hidden]

**Jineesh P P** <jineeshcyrix@gmail.com>  
To: BME Kozhikode <bmekkd1@gmail.com>  
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Wed, 31 Jul, 2024 at 12:11 pm

Dear Madam,

Kindly update the status  
[Quoted text hidden]

**Jineesh P P** <jineeshcyrix@gmail.com>  
To: Remya manikandan <remyavaigam@gmail.com>, BME Kozhikode <bmekkd1@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 1 Aug, 2024 at 12:35 pm

Dear Madam,

Kindly update

Thanks & Regards,

Jineesh PP  
District Incharge - Kozhikode  
Mob-7593847163  
[jineeshcyrix@gmail.com](mailto:jineeshcyrix@gmail.com)  
[www.cyrix.healthcare.com](http://www.cyrix.healthcare.com)

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

On Fri, Jul 26, 2024, 12:29 PM Jineesh P P <[jineeshcyrix@gmail.com](mailto:jineeshcyrix@gmail.com)> wrote:  
[Quoted text hidden]

Fri, 2 Aug, 2024 at 11:18 am

**BiomedicalEngineer Kozhikode** <[bmekkd1@gmail.com](mailto:bmekkd1@gmail.com)>  
To: Jineesh P P <[jineeshcyrix@gmail.com](mailto:jineeshcyrix@gmail.com)>

Sir/Madam

This is to inform you that there are no more details available at the institution..

[Quoted text hidden]

Wed, 7 Aug, 2024 at 2:34 pm

**BiomedicalEngineer Kozhikode** <[bmekkd1@gmail.com](mailto:bmekkd1@gmail.com)>  
To: Jineesh P P <[jineeshcyrix@gmail.com](mailto:jineeshcyrix@gmail.com)>

Sir/Madam

This is to inform you that there are no more details available at the institution. Supplier details are not available. Equipment was supplied through donation

[Quoted text hidden]

**Stock Entry**

Goods Receipt No : KL01251 FS 0000121 22  
 Purchase Order No :

Received Date : 25.05.2021

Order Date :

Invoice No : KL01181 FS 0000091 22

Invoice Date :

Discount Amount : 0.00

Handling/Other Charges : 0.00

Supplier Name : DEWANNED

Receipt No : KL01251 FS 0000121 22, Received Date : 25-May-2021, Scheme: COVID

SL No	Manufacturer Name	Batch No	Memo Date	Exp Date	Rate	IGST (%)	CGST (%)	SGST (%)	Quantity (Meters)	Unit	Name
1	SHENYANG GANTA MEDICAL TECH CO.LTD	VRUNMS	01/04/2021	31/03/2021	0.00	0.00	0.00	0.00	0.00	1	0.00
Total Qty & Value											
Total Value:											
Handling / Other Charges:											
Discount Amount:											
Total Invoice Amount:											



*[Signature]*  
 OFFICER  
 HEALTH CENTRE MELADY  
 PO THIRU-ODI KOZHICODE

Consignment Supplied by  
 Generated by Melady CHC (PEN No : N/A)

Remarks: RED AND ENTERED IN STOCK REG FILE PEG NO 61  
 (24-05-2021)

Consignment Collected by

*[Signature]*  
**MEDICAL OFFICER**  
 CH CENTRE MELADY  
 PO THIRU-ODI KOZHICODE

Kozhikode District

Drug Warehouse,

Manjeri,

Manjeri (Post),

Kozhikode-673614

Covid Donations - Issue Details

Print Date: 24/05/2021

Issue No & Date: KLD1 LFD/00083/21-22 & 24-MAY-21

Issued To: C.H.C. MELADI

Sr. No.	Item Name	Manufacturer	Batch No.	Mfg Date	Exp Date	Stock No.	Quantity
1	OXYGEN CONCENTRATOR	SHENYANG CANIA MEDICAL TECHNOLOGY	VA W2175 <b>2448080320</b>	01 Apr-2021	31 Mar-2031	G-12	1

Drug (with  
DRG)  
Page No. 6/  
FIE Reg

Let and entered in stock on  
25/5/21  
No release/stock/issue  
Covid scheme  
[Signature]

[Signature]  
[Circular Stamp: COMMUNITY HEALTH CENTRE MELADI, P.O. THIRYODI KOZHIKODE]

[Signature]  
MEDICAL OFFICER  
C.H. CENTRE MELADI  
PO THIRYODI KOZHIKODE

COMMUNITY HEALTH CENTRE  
MELADI, P.O. THIRYODI  
KOZHIKODE

[Signature]  
MEDICAL OFFICER  
C.H. CENTRE MELADI  
PO THIRYODI KOZHIKODE

COMMUNITY HEALTH CENTRE  
KOZHIKODE  
MELADI, P.O.  
Date: .....

# OXYGEN CONCENTRATOR FOR MEDICAL USE

VOLTAGE: ~ 230V±10%

POWER: 480W

FREQUENCY: 50HZ

NET WEIGHT: 24Kg

OXYGEN CONCENTRATION: 93±3%

SOUND LEVEL: ≤48dB

OXYGEN FLOW: 0-8L/min

OXYGEN DELIVERY PRESSURE: 0.04-0.08MPa

MODEL: V8-WN-NS

SN



2448902042721



2021.04.27

IP21



CE

0123

SHENYANG CANTA MEDICAL TECH.CO.,LTD.

NO.76-39 Shenbei Road, Daoyi Economic Development Zone,

Shenbei New District, 110136 Shenyang, People's Republic of China

EC REP

MedNet EC-REP GmbH

Address: Borkstrasse 10 · 48163 Muenster · Germany

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)

1130189

CYRIX HEALTH CARE PVT.L  
BEMP SERVICE PROVIDER  
EFFECTIVE MAINTENANCE SERVICE



MEDICAL OFFICER  
C.H. CENTRE MELADY  
P.O. THIKKODI KOZHICKODE

