


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC ALAPAD
3	Name of Equipment with Make, Model and Serial Number	HEAMATOLOGY ANALYZER Make: Prokan Model:PE-6000 Sr.No- A6006161200234EH
4	Equipment ID & Barcode	#126649 – 0831253
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date -17-03-2020 As per stock book installation date – 16-03-2017 (stock book attached for reference)
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	31-07-2024
9	Action taken	Checked the machine and found 3way valve,2way valve, CCU board, WBC chamber and RBC chamber are complaint. Need to replace these spares for further checking working condition of equipment and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	3way valve (1) – 18450/- 2way valve (2) – 36900/- CCU board – 59500/- WBC chamber- 35250/- RBC chamber – 35250/- TOTAL - 185350/- GST 18% - 33363/- TOTAL - 218713/-

13	Asset Value	Ticket master asset value is 199083/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	109.8%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report, and oem quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found 3way valve, 2way valve, CCU board, WBC chamber and RBC chamber are complaint, In ticket master installation date is 17.03.2020 As per stock book Equipment installed on 16/03/2017 and aged up to 7 years and 3 months. Spare cost is 109.8% of asset value and both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Spare cost is veeey high.


Jr. Consultant (Biomedical)
NHM, Thrissur



Signature of JC BM (NHM)

Date




Superintendent / Medical Officer (i/c)

Signature of seal


SUPERINTENDENT
Community Health Centre
Alapad, Thrissur



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 224442

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>CHC</u>	Call Registration Date : <u>31/7/24</u>
Address : <u>ALAPAD</u>	Caller ID : <u>126649</u>
<u>THRISSUR</u>	Date of Visit : <u>2/8/24</u>
Ph :	Asset No. : <u>0831253</u>
	EQPT Name : <u>Haematology Analysed</u>
	Manufacture : <u>Prokan</u> Model : <u>PE6000</u>
	S. No. <u>A600616120023</u> Dept. : <u>LABORATORY</u>
	<u>4 EH</u>

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : leakage inside. Chamber issue. Values not showing.

Action Taken : checked the machine and found 3 way valve, 2 way valve, ccu board, WBC Chamber, RBC chamber are complaint. Need to replace these spares for further checking working condition of equipment and spare quoted to bemp

Completed Date : 2/8/24 Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>2/8/24</u>		

Customer Remark Completed Pending



Service Engineer Name : Christy
Signature : [Signature]
Date : 2/8/24
Contact Number : 6238844347

Customer Name : FLAIBY C. JOHNSON
Signature : [Signature]
Date : 2/8/24
Contact Number : 8086365025
Designation : L.T.C.
Hospital Seal : SUPERINTENDENT Community Health Centre Alapad, Thrissur



Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



ISO 9001:2015
CERTIFIED COMPANY

Proforma Invoice No.: ADPL/S/RD/20240812/24-25

Date :12-08-2024

To,
Cyrux Healthcare Pvt. Ltd.
Ernakulam, First Floor, Petta Junction
30/641B, Poonithura, kerala

Ref. : By Mail.

Date :12-08-2024

CHC ALAPPAD

Sr. No	Description	Part No.	Price Each Ex- Work	Qty	Total (In Rs.)
	(Instrument Model: Hematology Analyzer) (Model : PE-6000) (S/No. A6006161200234EH)				
A	<u>Spares/Consumables Procurement Details:</u>				
(i)	3-Way Valve	PE00-006	18,450.00	1	18,450.00
(ii)	2-Way Valve	PE00-005	18,450.00	2	36,900.00
(iii)	CCU Board	PE60-010	59,500.00	1	59,500.00
(iv)	WBC Chamber	PE60-003	35,250.00	1	35,250.00
(v)	RBC Chamber	PE60-002	35,250.00	1	35,250.00
Total					1,85,350.00
(+ GST @ 18%)					33,363.00
Total: Two Lakh Eighteen Thousand Seven Hundred Thirteen Only.					2,18,713.00

Terms & Conditions : { Validity: 90 Days } { Delivery : 3-4 Weeks } { Payment Terms : 100% Advance }

Bank Account Details :-

Name of the Firm : Aspen Diagnostic Pvt. Ltd.

Aspen Diagnostics Pvt. Ltd.

Bank Name : ICICI Bank Limited

Account No. : 016005003239

Account Type : Current Account

MICR Code : 110229029

IFSC Code : ICIC0000160

12/08/24
Rajaram Krishnamurthy

National Service Manager



barcode - 0831253 caller id - 126649- chc alappad

messages

Mon, 12 Aug, 2024 at 12

From: Bijo T Joy <bijocyrix@gmail.com>

To: service support <servicesupport@rdgc.in>

Cc: binto@rdgc.in, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

Kindly provide the quotation for below mentioned Haematology Analyzer's spares.

hospital - CHC ALAPPAD

equipment - hematology analyzer

quotation need - ccu board

wbc chamber

rbc chamber

3-way valve (1nos)

2 way valve (2nos)

barcode - 0831253

caller id - 126649

sr no - A6006161200234EH

make - Prokan

model - PE-6000

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEALTHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

From: service support <servicesupport@rdgc.in>

Mon, 12 Aug, 2024 at 5

To: Bijo T Joy <bijocyrix@gmail.com>

Cc: binto@rdgc.in, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

Please find the attached P.I. & share the P.O. along with payment details for billing.

Thanks & Regards

Hari Ram

[Quoted text hidden]

HEMATOLOGY ANALYSER

S.No.	Name of the Item	Receipt Date	Name of Company	Sl. No.
1	Hematology Analyser (3. PARS) Thyeard 7m 14.	16/3/17	PROKAM Rapid Diagnostic Private Ltd.	96066161 20023464
2	Hematology Analyser (Part) of Erba - H510 (Tranviva) Rs -> 4,48,899/- Installed on -> 15/3/24	25/3/24	Tranviva (Erba H510)	Sl. No. 111048222888

810/500X
 Under BHS-PS-14-15
 KMS/EP-10755/2015
 Installed on 17/3/17.
 Purchase Order No 456/20-16-2017/55
 CHC/ALPD/LAB/CEP/19

8006-540615025
 0831253

Installed on 15/3/24

Remarks:
 Received from Blue Paritylabs Govt Project fund.
 Page no: 44 of Instrument Register.
 Serial No 12 B.

Purchased through Blue Health Grant - Project 2014



