


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	UPHC KANHANGAD
3	Name of Equipment with Make, Model and Serial Number	ECG MACHINE MAKE: BPL MODEL: CARDIART 6108 T S/N: AVMT6D26901
4	Equipment ID & Barcode	#124772, 1440898
5	Date of purchase / Year of manufacture /Installation Date	13-06-2016
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	22-07-2024 (Toll free)
9	Action taken	Checked the machine, found problem with Main board and battery. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Not working, fully damaged.
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	ECG 6108T MAIN PCB ASSY- 14000/- BM/OM NIMH BATTERY 9.6V 1.4AH- 2048/- Total: 16048/-


13	Asset Value	₹18656
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	86.02%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine, found Main board and battery are defective. Machine installed on 13-06-2016 and aged up to 8 years and 1 month. Repairing cost is above 86.02% of asset value. Both criteria for RBER met. As per the tender clause 5.3.14.1, We recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JAYESH KRISHNAN 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Checked the machine and found main board and battery defective. Machine installed on 13/06/2016. Repairing cost is above 86% of asset value. Cyrix service report and OEM quotation attached. So recommending the equipment for BER.

 14/08/24
Signature of JC BM (NHM) CONSULTANT
Biomedical Engineering
Arogyakeralam (NHM)
Kasaragod

Date 14/08/24



Seal


Signature of
Superintendent / Medical Officer (i/c)
Medical Officer

Urban Primary Health Centre
National Urban Health Mission
Kanhangad Kadappuram
Kasaragod Dist.



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 198398

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility	UPHC Kanhangad	Call Registration Date	22/07/24
Address	Punjari Kadurgod	Caller ID	124772
Ph	7902522046	Date of Visit	25/07/24
		Asset No.	11440898
		EQPT Name	ECG machine
		Manufacture	BPL
		Model	Cardiart 61087
		S. No.	AVMT6D25901
		Dept.	OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Lead Ie not getting, vibrations in ECG

Action Taken : checked the unit, found vibrations coming in ECG wave. Lead I not getting also. Need to replace main board and battery for further checking of the machine

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Jayesh Prithwan	25/07/24	4:10 PM	4:50 PM

Customer Remark

Completed
Urban Primary Health Centre
Kanhanged Kadappuram
Kasaragod Dist

Pending
Urban Primary Health Centre
National Urban Health Mission
Kanhanged Kadappuram
Kasaragod Dist

The
m. Anuetha
19/07/24

Service Engineer Name : Jayesh Prithwan	Customer Name : Sandhya K V
Signature : <i>[Signature]</i>	Signature : <i>[Signature]</i>
Date : 25/07/24	Date : 25/7/24
Contact Number : 7593847156	Contact Number : 9526641086
	Designation : Staff nurse
	Hospital Seal :



QTN/BISO/250724-030

25/07/2024

QUOTATION

To,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

Dear Sir,

Please find below our lowest quotation for the following items:-

SL NO.	ITEM DESCRIPTION	QUANTITY	UNIT PRICE	PRICE (incl of all taxes)
1	ECG 6108T MAIN PCB ASSY NEW (WITH LPC IC)	01	12500+12%TAX	14000/-
2	BM/OM NIMH BATTERY 9.6V 1.4 AH	01	1600+28%TAX	2048/-

Terms and conditions:

1. Payment 100% in advance along with Purchase order
2. Delivery within 15 days

Thanking you and assuring you our best attention always.

For Biomedix Solutions



Authorized Signatory

Mobile: 9846969564

Branches: Kozhikode, Kannur, Kottayam-Kollam.

For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com



BIOMEDIX SOLUTIONS
No: 7/437- B, Mannarakayam P.O,
Kathilankalpadi, Kanjirapally, Kottayam - 686506,
Mob:9846969564,



PRICE ENQUIRY 1440898

5 messages

Wed, 24 Jul, 2024 at 11:28 pm

Jayesh Krishnan <jayeshcyrix@gmail.com>
To: BPL Medical Equipment Service <biomedixservice@gmail.com>
Cc: BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Dear sir,

Please share the price and availability of battery and main board for ECG machine.

Make: BPL

model: Cardiart 6108T

S/N: AVMT6D26901

—
Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Wed, 24 Jul, 2024 at 11:28 pm

BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>
To: jayeshcyrix@gmail.com



Thu, 25 Jul, 2024 at 9:00 am

BPL Medical Equipment Service <biomedixservice@gmail.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>
Cc: BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

All 6108T ecg machine's spares are available in our stock.....
Rate already updated in our quotation to cyrix in previous months.... Please check.... If needed we can provide the same again.....

For Biomedix Solutions
9846949564
[Quoted text hidden]

Thu, 25 Jul, 2024 at 10:10 am

Jayesh Krishnan <jayeshcyrix@gmail.com>
To: BPL Medical Equipment Service <biomedixservice@gmail.com>
Cc: BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,
Kindly share the quote for Main board and battery only.
[Quoted text hidden]

Thu, 25 Jul, 2024 at 11:59 am

BPL Medical Equipment Service <biomedixservice@gmail.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>



Cc: BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Please find the attached quotation.....


For Biomedix Solutions

9846949564

[Quoted text hidden]



REDMI NOTE 8 PRO
JAVESH



REDMI NOTE 8 PRO
JAVESH

*For the
Dr. Anantha
19/08/24*

Medical Officer
Urban Primary Health Centre
National Urban Health Mission
Kadappuram
Kadappuram Dist.

