

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHC PERINGOTTUKURISSI
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS MAKE : MEDICARE EQUIPMENTS MODEL : MCP SN : 07C1709
4	Equipment ID/ Barcode	0942163 # 128535
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil As per stock book equipment installed on 30/12/2017 (stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	09/08/2024 (toll free)
9	Action taken	Checked the BP Apparatus found that mercury spilled and glass tube broken. Need to replace glass tube and mercury for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Mercury 3 ml = 991/- Glass tube = 112/- Total = 1103/-

13	Asset Value	RS.1323/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	83.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the BP Apparatus and found that mercury spilled and glass tube broken. In ticket master installation date is Nil, As per stock book Equipment installed on 30/12/2017. aged up to 6 years 7 months. Spare cost is 83.3% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P


*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 B.p apparatus installed on 30/12/2017
 aged up to 6 years 7 months.
 spare cost is Rs-83.3%. Quotations
 submitted. Both criteria of RBER met.
 Ash
 Signature of JC BM (NHM) 21/08/24.

**Consultant (Biomedical Engg)
 NHM (Aranyakeralam), Palakkad**

Date: 22.8.24



Signature of _____
 Superintendent/Medical Officer
 Medical Officer
 Family Health Center
 Peringottukurussi
 Paruthippuli (PO)-678 573
 Phone: 04922-217340



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1001206

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>PHC Peringottukurussery</u>	Call Registration Date: <u>9.08.2024</u>
Address: <u>Palakkad (H)</u>	Caller ID: <u>128535</u>
<u>Kerala</u>	Date of Visit: <u>12.08.2024</u>
Ph: <u>9072788140</u>	Asset No: <u>0942163</u>
	EQPT Name: <u>Bp Apparatus</u>
	Manufacture: <u>medicare</u> Model: <u>mcp</u>
	S. No: <u>07C1709</u> Dept: <u>Subcentre</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : mercury spilled, glass column broken

Action Taken : checked the equipment found that glass column broken and mercury spilled. also found that mercury tank leaked. need to replace glass column, mercury tank and mercury for further checking.

Completed Date: 12.8.24 Time: 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Nil</u>			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Dileep</u>	<u>12.8.24</u>	<u>10.00 AM</u>	<u>11.00 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Dileep
Signature: Dileep AF
Date: 12.08.2024
Contact Number: 9048751999



Customer Officer
Signature: [Signature]
Date: 12.08.2024
Peringottukurussery
Family Health Center
Peringottukurussery (PO)-678 038
Contact No: 98472 99500
Phone: 04998 2178107
Designation: Asst Surgeon
Hospital Seal

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL: armedicare07@gmail.com
MOB: 9446057350, 7994654108,



QUOTATION

REF: AR/QTN-113/24-25

Date: 17-08-2024

To,
Cyrux Health Care Private Limited
30/641B, Vandipetta Rd, near SBI Bank, Poonithura, Maradu
Ernakulam, - 682038
Kerala 682038

No.	Description	Qty	Unit Price(INR)	TAX	Total(INR)
1.	BP APPARATUS ACCESSORIES				
1.	MERCURY 3ML	1	835.00	151.2	991.20
2.	GLASS TUBE	1	100.00	12.00	112.00
				Grand Total	1103.20 INR
				Round off	- 0.20 INR
				Net Total	1103 INR

INR One thousand one hundred three Only

Terms and Conditions for sale :

- 1 **Validity:** Six months from the date of quotation.
- 2 **Payment:** Within 30 days from the date of supply.
- 3 **Delivery:** Within 7 days after receipt of the signed purchase order.
- 4 If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- 5 Bank Details:

Account holder's name : AR MEDICARE
Bank Name : State Bank Of India, Thazhechovva
Account No. : 39648455024
IFSC code : SBIN0016861



For AR MEDICARE,
KANNUR



REQUIRED FOR PRICE ENQUIRY BP APPARATUS

2 messages

Akhilesh P <akhileshcyrix@gmail.com>

Fri, 16 Aug, 2024 at 8:23 pm

To: ar medicare <armedicare07@gmail.com>

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir

Please share the price and stock availability of bp glass tube and mercury for bp apparatus.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

ar medicare <armedicare07@gmail.com>

Sat, 17 Aug, 2024 at 10:58 am

To: Akhilesh P <akhileshcyrix@gmail.com>

Sir,
PFA




AR MEDICARE

Thanks & Regards

AR MEDICARE
Room No:865,2nd floor
AL-RAHA Arcade
Thazhe Chovva, Kannur
Mob: 9446057350, 7994654108, 8138091108
Email: armedicare07@gmail.com
GSTIN :32BDYPV5030P1ZP

[Quoted text hidden]



MCP
Sphygmomanometer
No. 1000
5
MCP
No. 07C1721



H.A. to
22.8.2017
Medical Officer
Family Health Center
Perinikottukkurussiy
Paruthippalli (PO)-678 572
Phone: 04922-217340



[Handwritten Signature]
 Medical Officer
 Family Health Center
 Peringottukurussi
 Paruthipulli (PO)-678 573
 Phone: 04922-217340

134

BP Apparatus

Receipt

20/12/17

Received From PHE Peringol/Enkumma

1

LS



Handwritten signature

22-8-29

Medical Officer
Family Health Center
Peringottukur
Paruthippulli (PO)-678 673
Phone: 04922-217340