


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	THQH PERAVOOR
3	Name of Equipment with Make, Model and Serial Number	BABY WEIGHING MACHINE Make : SKNOL Model: NA SN:NA
4	Equipment ID/ Barcode	1320280 #129880
5	Date of purchase/ Year of manufacture/Installation Date	23-04-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 17-08-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found mainboard, Load cells damaged. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	4990 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 23-04-2020 Aged up to 04 years 3 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that machine not switching ON, due to main board & load cells are defective. Aged up to 4 years & above. The spares not available in the market. So recommended for BER as per BER protocol.



Signature of JC BM. (NHM)

KANNUR - 670002

PERSIS CYRIAC
SNO.

Date: 29/8/2024.



Signature 
SUPERINTENDENT
PERAVOOR
KANNUR
670673



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 232002

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.ki@cyrix.in

Health Facility : THOH PERAVOOR	Call Registration Date : 17/08/2024
Address : Peramboor	Caller ID : 129880
KANNUR	Date of Visit : 19/08/2024
Ph : 7012 376 493	Asset No. : 1320280
	EQPT Name : Weighing Machine
	Manufacture : SKNOC Model : Na
	S. No. : Na Dept. : Post-operative

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mainboard and load cells are damaged.

Action Taken : Checked the machine and found that machine is not switching ON. Found mainboard, Load cells damaged. Need to replace these spares for further checking.

Completed Date : 19/08/2024 Time : 10:30 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Sayanth	19/08/2024	9:00 AM	10:30 AM

Customer Remark : Completed Pending



SUPERINTENDENT
KALUK HOSPITAL
PERAVOOR
KANNUR
670673

Dr. PERSUS CYRIAC

Service Engineer Name : Sayanth
Signature :
Date : 19/08/2024
Contact Number : 8606 345 195

Customer Name :
Signature :
Date : 19/8/24
Contact Number : 7012376493
Designation :
Hospital Seal : GNO

STOCK BOOK OF.....

21

Name of Article..... Baby weighing machine.....

Date	No. and date of Voucher or Invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward					
5/4/17	To	is from PND					
		page no 17	1		1		
16/2/18		Rd from store	1		2		
		(Electronic) mechanical					
		currently working in Pharmacy					
		RD 4900 stock					
14/1/18		condemned by Supt T-11 Peravoor					
		dr order no 82/2015		1	1		
23/4/20	✓	Received from store (station)	1		2		
10/1/2021		Tr out to unserviceable Register page no: 3		1	1		
21.1.2021		Rd from store (electronic)	1		2		kmel Rs-5888 (warranty 1 year)
1/2/21		Transferred part of page no-15		1	1		



Dr. Sahni
 29/8/2024
 SUPERINTENDENT
 TALUK HOSPITAL
 PERAVOOR
 KANNUR
 670673



edshw
Dr. Sahani
29/8/2024

**SUPERINTENDENT
TALUK HOSPITAL
PERAVOOR
KANNUR
670673**