


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	PHC PAMBOOR
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus Make: OMRON HEALTH CARE Model: HBP 1300 SN:04010252LF
4	Equipment ID/ Barcode	0842528 #126073
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date – NA As per the stock book installation date – 21-12-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 29-07-2024
9	Action taken	Checked the equipment and found board and display are defective. Need to replace these spares for further service.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200 /-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked equipment found board and display are defective, As per the stock book equipment installed on 21-12-2020 and covered up to 3 year and 7 months, Quotation not submitted since spares are not available in the market. So we are recommending equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Space not available*

  
*M...*  
Signature of JC BM (NHM)

Date: 

*for 10/19/2021*  
Signature of   
PIN - 680013



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000255

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... P.H.C..... Address..... Pambour..... ..... THIRUVAR..... Ph : ..... 9495417441.....	Call Registration Date : ..... 29-07-2024..... Caller ID : ..... 126073..... Date of Visit : ..... 30-07-2024..... Asset No : ..... 0842528..... EQPT Name : ..... Bp. apparatus..... Manufacture <i>amman health care</i> Model : ..... 18BP.1300..... S. No. 0401025246..... Dept..... OP.....
---	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... Not functioning on.....  
 .....  
 .....

Action Taken : ..... checked the equipment and found board and display are defective need to replace these spare len further service.....  
 .....

Completed  Date : ..... 30-07-2024 Time : ..... 2:00pm..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Buo	30-07-2024	1:00pm	2:00pm

Customer Remark Completed  Pending

Service Engineer Name : *Byjo*  
 Signature: *[Signature]*  
 Date: 30-07-2024  
 Contact Number: 8921870010

Customer Name : ANUSHA C X  
 Signature: *[Signature]*  
 Date: 30/7/24  
 Contact Number: 9495417441  
 Designation: Pharmacist  
 PHARMACIST  
 STORES CUSTODIAN

