


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KASARGOD
2	Name of Hospital	TH PANATHADY
3	Name of Equipment with Make, Model and Serial Number	Name:- Digital BP apparatus Make:- INFI Model:-N/A SN:- 170603142
4	Equipment ID & Barcode	#128422 - 1420549
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date is 20-07-2018 As per stock book installation date is 13-01-2018 (stock book attached for reference)
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registrationof complaint through email/ Toll free)	09-08-2024 ( TOLL FREE)
9	Action taken	Checked the machine and found that main board, display, motor and pressure sensor are defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendation.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	1500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED.
16	Reasons for recommending the equipment as BER	Checked the machine and found that main board, display, pressure sensor and motor are defective. Ticket master installation date is 20-07-2018 As per stock book installation date is 13-01-2018 (stock book attached for reference) aged up to 6 years and 7 months. Spare quote not submitted since spares are not available in market and so recommending for Condemnation.
17	Name & Signature of CYRIX Authority	ABHINAV O 

\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked the machine and found main board, display, Pressure sensor and motor are defective. (Equipment installed on 13/01/2018 (stock book details attached). Cyrix service report attached. So recommending the equipment for BER.

  
 Signature of **JUNIOR CONSULTANT**  
**Biomedical Engineering**  
**Arogyakeralam (NHM)**  
**Kasaragod at Kanhangad**

Date \_\_\_\_\_

  
**Seal**

For   
**Signature of Surgeon**  
**Taluk Superintendent / Medical Officer (I/G)**  
**Kasaragod, Kerala**



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1002207

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>Taluk Hospital</u></p> <p>Address: <u>Parathady</u></p> <p style="padding-left: 40px;"><u>Kasaragod</u></p> <p>Ph: <u>9495906110</u></p>	<p>Call Registration Date: <u>09.08.2024</u></p> <p>Caller ID: <u>128422</u></p> <p>Date of Visit: <u>12.08.2024</u></p> <p>Asset No: <u>1420549</u></p> <p>EQPT Name: <u>Digital Bp Apparatus</u></p> <p>Manufacture: <u>Inf</u> Model: <u>NA</u></p> <p>S. No: <u>170603142</u> Dept: <u>Ward</u></p>
---	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not Switching on

Action Taken : Checked the machine and found that machine mainboard, display, motor and pressure sensor are defective.

Completed  Date : 12.08.24 Time : 3:15pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Abhinav O</u>	<u>12-08-24</u>	<u>2:40pm</u>	<u>3:15 pm</u>

Customer Remark Completed  Pending

<p>Service Engineer Name : <u>Abhinav O</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>12.08.2024</u></p> <p>Contact Number: <u>8943392989</u></p>	<p>Customer Name : <u>Pindin Joseph K.</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>12.08.24</u></p> <p>Contact Number: <u>9495906110</u></p> <p>Designation : <u>BND</u></p> <p>Hospital Seal: <u>[Seal]</u></p>
--	---

# BP APPARATUS DIGITAL.

270

Date	Rd	Issued	Recd	Balance	Sign
13/1/18	Rd from the store	2		2	<del>BT</del>
14/1/18	Rd from the store	1		3	<del>BT</del>
20/6/20	Ti. out to instrument book OPD vol II Pg 123. (Rd on <del>14/1/18</del> )		1	2	0
9/3/23	Go. in brown sub stock burnt down equipment register secondary cell Pno-35 (Rs-2750/each) Wrong entry	+			Ward store
26/6/23	Cr. Old to USAR Pno 22			1	1



Pa  
[Signature]  
 Assistant Surgeon  
 Taluk Hospital Poodankallu  
 Kasaragod, Kerala



For  
Assistant Surgeon  
Taluk Health Office, Podankallu  
Kasaragod, Kerala