


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	GH THALASSERY
3	Name of Equipment with Make, Model and Serial Number	PATIENT MONITOR Make: BPL Model: EXCELLO D SN: EVTD7E3630
4	Equipment ID/ Barcode	1311221 #126008
5	Date of purchase/ Year of manufacture/Installation Date	15-08-2017 As per stock book Installation Date is 11.07.2010 (stock book copy attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 27-07-2024
9	Action taken	Checked the Machine found Equipment is Not Turning ON, found that SMPS and Main Board Defective. Enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	SMPS Board -35,000 /- Main Board- 45,000/- <b>Total-80,000/-</b>



13	Asset Value	54500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	146.788%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Machine found Equipment is Not Turning ON, found that SMPS and Main Board Defective. Enquired OEM for quotation. In software Equipment installed on 15-08-2017 As per the stock book installation date is 11.7.2010 aged up to 14 years . Spare cost is 146.788% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that the machine not working due to SMPS, main board defective. Aged up to 14 years. The service cost not feasible as per the OEM's quotation. So recommended for BER as per BER protocol.

  
23/08/2024  
Signature of JG/BM (NHM)  
Radhika (NHM)  
CONSULTANT IN BIOMEDICAL  
NATIONAL HEALTH MISSION  
KANNUR - 670002



Date: 24 - 8 - 24

  
SUPERINTENDENT  
Govt. General Hospital  
Thalassery  
Superintendent/Medical Officer (i/c)



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 225206

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>G.H Thalassery</u></p> <p>Address <u>Thalassery</u> <u>Kannur</u></p> <p>Ph : <u>9947868404</u></p>	<p>Call Registration Date : <u>27-7-24</u></p> <p>Caller ID : <u>126008</u></p> <p>Date of Visit : <u>29-7-24</u></p> <p>Asset No. : <u>1311221</u></p> <p>EQPT Name : <u>Multipara monitor</u></p> <p>Manufacture <u>BPL</u> Model : <u>Excello D</u></p> <p>S. No. <u>EVSDTE3630</u> Dept. <u>SBICU</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Machine not switching ON

**Action Taken :** checked the machine and found equipment is not turning ON. found that SMPS and main board defective. Enquired OEM for quotation.

Completed  Date : 29-7-24 Time : 1:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasiboj</u>	<u>29-7-24</u>	<u>12:00pm</u>	<u>1:00pm</u>

**Customer Remark** Completed  Pending

**Service Engineer Name :** Rasiboj  
**Signature :** Rasiboj  
**Date :** 29-7-24  
**Contact Number :** 9995521436



**Customer Name :** Babitha, CP  
**Signature :** [Signature]  
**Date :** 29/7/24  
**Contact Number :** 956248199  
**Designation :** NO  
**Hospital Seal :**

**SUPERINTENDENT**  
Govt. General Hospital  
Thalassery

09.08.2024  
RINS/SA/24/4

TO,  
CYRIX HEALTHCARE PRIVATE LIMITED  
ERNAKULAM

QUOTATION FOR SMPS BOARD & MAIN BOARD FOR MONITOR EXCELLO-D

SL NO	ITEM	RATE	QTY	AMOUNT
1	SMPS BOARD FOR MONITOR EXCELLO-D (SI No: EVTD7E3630)	35000	1 no	35000.00
2	EXCELLO-D MAIN BOARD (ARM BOARD+INTERFACE PCB) (SI No: EVTD7E3630)	45000	1 no	45000.00
GST EXTRA 18%				

For , Rins Marketing & Services



Authorized Signatory

OUR BANK DETAILS:

BANK: FEDERAL BANK, A/C NO: 19975500000015, IFSC: FDRL0001997, BRANCH: TALAP



---

**QUOTATION REQUIRED FOR SMPS,MAINBOARD, PATIENT MONITOR #1311221,#126008, GH THALASSERY**

2 messages

Fri, 9 Aug, 2024 at 9:44 am

avinash t <avinashtcyrix@gmail.com>

To: rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir,

Please provide Quotation and stock availability of **SMPS and MAIN BOARD** for below mentioned machine.

EQUIPMENT NAME- PATIENT MONITOR

MAKE- BPL

MODEL- EXCELLO D

SL NO- EVTD7E3630

Thanks & Regards

Avinash T

District charge, kannur

bemp kerala

mobile: +91 6364064948

E mail: [avinashtcyrix@gmail.com](mailto:avinashtcyrix@gmail.com)

**CYRIX HEALTHCARE (P) LTD**

**30/641B | Petta | Punithura | Kochi | Kerala- 682038**

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Fri, 9 Aug, 2024 at 11:47 am

---

rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>

To: avinash t <avinashtcyrix@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir,

Please find the attached quotation.

Regards,

Anagha

**RINS**  
MARKETING & SERVICES

TM 364 N, 1st Floor, City Tower

Opp. Chinmaya Balabhavan, Kannur-2

Tel.: 0497-2704624, Fax.: 2709789

Mob.: 9846008441, 9947064624

E-mail: [rinsmarketing@gmail.com](mailto:rinsmarketing@gmail.com)

[Quoted text hidden]

ROAD SIDE MONITOR  
(CARRIAGE)

71

	R	I	B	SG
11/7/10 Received from Store + Gc. (V.N. + V.P.) NR 2001 of V.P. Reg No. 167	8	-	8	Bush /11
11/7/10 P. out to 200 [Sd. 200-100597-019. -021, 018, 020.] (Grant is untraceable again)		4	4	
8/8/10 241017 R66 - Rd 200 to Store (TWA) RS 3,61,669 / GA 3112/2029 in use 241017 issued	7	7	14	11 11
31/5/2018 condemned as per order no: B2/ 18176/2017 dtd 31/5/2018 In use		4 7	7 7	11 11 11/5/2018



*[Signature]*  
SUPERINTENDENT  
Govt. General Hospital  
Thalassery



Preventive maintenance Schedule **ktpl**

Q1 26/9/19 Q3 \_\_\_\_\_

Q2 \_\_\_\_\_ Q4 \_\_\_\_\_

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
HEALTH CARES PVT. LTD.

**TOLL FREE NO: 1800 - 425-7669**

BAR CODE - (8004-890615225) **1311221**




**CYRIX HEALTH CARE**  
**BEMP SERVICE PRO**  
**PREVENTIVE MAINTENANCE SERVICE**

PMS DONE ON: 26-9-2019

NEXT PMS DUE: 25-9-20

PERFORMED BY: \_\_\_\_\_



*[Handwritten Signature]*  
**SUPERINTENDENT**  
 Govt. General Hospital  
 Thalassery

