


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	GH KASARGOD
3	Name of Equipment with Make, Model and Serial Number	ENDOMOTOR MAKE: GUILIN WOODPECKER MODEL: E-Com+ S/N: N/A
4	Equipment ID & Barcode	#122089, 1453313
5	Date of purchase / Year of manufacture /Installation Date	23-09-2019
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	09-07-2024 (Toll free)
9	Action taken	Checked the machine, found problem with main board, display and outer cover . Need to replace these spares for further checking. Enquired supplier for spare quotation.
E	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No Recommendation
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	37663/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and Supplier EOL letter attached.
16	Reasons for recommending the equipment as BER	Checked the machine, upon examination, it has been discovered as main board, Display and outer cover defective. Machine installed on 19-09-2019 and aged up to 4 years and 9 months. In our source replacement part, we inquired with the supplier Union chemist, as per them, there are no available spares from the manufacturer because it's an obsolete model. Considering the circumstances and in adherence to the tender clause 5.3.14.2, which stipulates that spares or service not available from the vendor, the equipment should be evaluated for condemnation, we recommend that this equipment is marked for condemnation.
17	Name & Signature of CYRIX Authority	Ajith babu K 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Checked the machine and found that main board, display & outer cover defective. Equipment installed on 19-09-2019. As per supplier, equipment is obsolete model, spares & service not available. cyrix service report and supplier letter attached. so recommending the equipment for BER.

Signature of JC BM (NHM) SULTAN 

Biomedical Engineering
 Aranyakeralam (NHM)
 Kanhar


 Date 23/08/24
 Signature of Superintendent Medical Officer (SC) 
 671 121



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 208473

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>GH Kasaragod</u></p> <p>Address <u>Kasaragod</u> <u>Kerala</u></p> <p>Ph : <u>9947198297</u></p>	<p>Call Registration Date : <u>9-7-24</u></p> <p>Caller ID : <u>122089</u></p> <p>Date of Visit : <u>10-7-24</u></p> <p>Asset No. : <u>1453313</u></p> <p>EQPT Name : <u>Endo motor</u></p> <p>Manufacture <u>huilin wood</u> ^{Penkep} Model : <u>E-10mt</u></p> <p>S. No. <u>NA</u> Dept. <u>Dental department</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working,

Action Taken : checked the machine and found problems with main board, display and covers, need to replace those spare for further working condition of the equipment

Completed Date : 10-7-24 Time : 10:45 AM Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ajith Babu</u>	<u>10-7-24</u>	<u>10:00 AM</u>	<u>10:45 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Ajith Babu
Signature : [Signature]
Date : 10-7-24
Contact Number : 8848079285

Customer Name : ADUMIAT-125
Signature : [Signature]
Date : 10.07.2024
Contact Number : 8129220834
Designation : Dental Engineer
Hospital Seat : Kasaragod GH



UNION CHEMISTS

PHARMACEUTICAL AND SURGICAL DISTRIBUTORS

ANE BAGILU, KASARAGOD-671 121, KERALA, PHONE: 04994 226951

DL NO: KL-KAS-117257-20B-2020, KL-KAS-117258-21B2020

GST IN: 32AACFU1827D1ZO

TO
BIOMEDICAL ENGINEER
GENERAL HOSPITAL
KASARAGOD

Sir,

This letter is to inform you that Endomotor supplied by us with following details,

Make: Guilin Woodpecker

Model: E- Com+

This equipment was obsolete model so spares and service not available from original manufacturer.

Date: 31-07-2024

Thanking you

UNION CHEMISTS
(Pharmaceutical & Surgical Distributors)
Door No. 325/A7(1) & (2)
BELUR BUILDING, ANEBAGILU
KASARAGOD - 671 121
Phone : 04994 - 226951
35105
Yours Faithfully
CST No. 36118827



Jayesh Krishnan <jayeshcyrix@gmail.com>

PRICE ENQUIRY - 1453313

5 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, May 14, 2024 at 3:45 PM

To: UNION CHEMISTS Pharmaceutical Surgical Distributors <unionchemistsksd@gmail.com>

Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

Please share the price and availability of front cover and display with board for Endomotor

Make:Guilin wood pecker

Model: E-com+

S/N: N/A

Photo of the machine attached for your reference.

--

Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

2 attachments

**IMG-20240502-WA0005.jpg**
684K**IMG20240502124701.jpg**
1524K

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, Jul 16, 2024 at 10:10 AM

To: UNION CHEMISTS Pharmaceutical Surgical Distributors <unionchemistsksd@gmail.com>

Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Sir,

Kindly update. If the spare or service not available, Kindly explain it in your letter head.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Fri, Jul 19, 2024 at 9:57 AM

To: UNION CHEMISTS Pharmaceutical Surgical Distributors <unionchemistsksd@gmail.com>

Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Sir,

Kindly update urgently.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Sep 2, 2024 at 3:35 PM

To: UNION CHEMISTS Pharmaceutical Surgical Distributors <unionchemistsksd@gmail.com>

Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Sir kindly send the Letter at the earliest

[Quoted text hidden]

UNION CHEMISTS Pharmaceutical Surgical Distributors

Mon, Sep 2, 2024 at 4:23

<unionchemistsksd@gmail.com>

PM

To: Jayesh Krishnan <jayeshcyrix@gmail.com>


new

[Quoted text hidden]




img20240902_16121878.pdf

140K



[Handwritten Signature]
 ട്രസ്റ്റി
 സി. ജനറൽ ആശുപത്രി
 കോട്ടയം ഐ.എ. - 671 12



KASARAGOD

Handwritten text in Malayalam script, including "സംഗ്രഹ" and "കാസറഗോഡ്".

254

ENDOMOTOR & APEX LOCATER

Sl	Date	Details	Remarks
1-	23.09.19	Invoice No: 1900073004163 Endomebo & Apex Locater	Plan Fund (2019-20) Rate - 37663.36
	24.09.19	put in use	Forepage No LP IV - A page 58 H/W



സൂപ്രണ്ട്
 നവ. മനോജ് കുമാർ
 ഓഫീസറായ്
 ഓഫീസറായ് എ.ഒ. - 571 121