


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHC POOKATUKAVU
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE MAKE : SAMSO MODEL : NA SN : NA
4	Equipment ID/ Barcode	0942933 # 130680
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil (As per stock book installation date 19/10/2020, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	21/08/2024 (toll free)
9	Action taken	Checked the machine found not switching on. Found load cell and Board damaged. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS.1600/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found not switching on. Found load cell and board damaged. In ticket master installation date is Nil, As per stock book equipment installed on 19/10/2020. aged up to 3 year 9 months. Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Equipment installed on 19/10/2020 aged up to 3 year and 9 months - Quotation not submitted




29/08/24
Signature of Junior Consultant (Biomedical) NHM (Aayakalam)

Date: _____





MEDICAL OFFICER
FAMILY HEALTH CENTRE
POOKKOTTUKAVU - 679 515
 Signature of
 Superintendent/Medical Officer (i/c)

WEIGHING MACHINE (DIGITAL)

Date / Invoice no	From whom received / to whom issued	Receipt	Cumulative Initial of Balance
19.10.2018 20.10.2020	Received from store put in use	1 COP 1	1 
7/10/23	Received 1 digital weighing machine for IMAGE ROOM	1	2 
7/10/23	(put in use)	2	3 MEDICAL OFFICER FAMILY HEALTH CENTRE POOKKOTTUKAVU - 679 515




 MEDICAL OFFICER
 FAMILY HEALTH CENTRE
 POOKKOTTUKAVU - 679 515

SAPISO

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0942933



MEDICAL OFFICER
FAMILY HEALTH CENTRE
POOKKOTTURAM - 679 518