


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC MATTANNUR
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make: A&D MEDICAL Model: UM-211 SN: NA
4	Equipment ID/ Barcode	1330285 #130549
5	Date of purchase/ Year of manufacture/Installation Date	In ticket master Installation date is – 30-10-2018 As per the stock book installation date is 03-10-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll-free)	Toll free 21.08.2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. In ticket master installation date is 30.10.2018. As per the stock book installation date is 03-10-2018 (stock book copy attached for reference) Aged up to 05 years 09 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that machine not switching ON, due to main board, pressure sensor, pump are defective. Aged up to 5 years & above. its spare cost economically not feasible. I recommended for BER as per BER protocol.

 Signature of JC BM (NHM)  
30/08/2024

CONSULTANT BIOMEDICAL  
NATIONAL HEALTH MISSION  
KANNUR - 670002



Date: 31 08 2024

  
Superintendent/ Medical Officer (i/c)  
Mattannur  
മട്ടന്നൂർ-670 702





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 234861

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p><b>Health Facility</b> ..... CHC MATTANNUR .....</p> <p><b>Address</b> ..... MATTANNUR .....</p> <p style="padding-left: 40px;">RANNUR .....</p> <p><b>Ph :</b> ..... 09446517548 .....</p>	<p><b>Call Registration Date :</b> ..... 21/08/2024 .....</p> <p><b>Caller ID :</b> ..... 130549 .....</p> <p><b>Date of Visit :</b> ..... 22/08/2024 .....</p> <p><b>Asset No. :</b> ..... 1330285 .....</p> <p><b>EQPT Name :</b> ..... BP Apparatus .....</p> <p><b>Manufacture</b> A&amp;D medical.. <b>Model :</b> ..... UM-211 .....</p> <p><b>S. No.</b> ..... NA ..... <b>Dept.</b> ..... OP .....</p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... Equipment is not switching ON .....

**Action Taken :** ..... checked the equipment found equipment is not switching ON, found that Main Board, Pressure sensor and Pump Defective. Need to replace these spare for further checking .....

**Completed**  **Date :** ..... 22/08/2024 **Time :** ..... 12:10pm ..... **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
Prathyush T. Bhaskar	22/08/2024	11:35am	12:10pm

**Customer Remark**  **Completed**  **Pending**

<p><b>Service Engineer Name :</b> Prathyush T. Bhaskar</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 22/08/2024</p> <p><b>Contact Number :</b> 9947025643</p>	<p><b>Customer Name :</b> Jessykn</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 22/8/24</p> <p><b>Contact Number :</b> 9446517548</p> <p><b>Designation :</b> Sem Ng</p> <p><b>Hospital Seal :</b> </p>
---	---









RIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
 TOLL FREE NO: 1800 - 425-7669  
 UAR CODE - (8004-890615225) 1330285

SYS mmHg

DIA mmHg

PUL /min.

▲  
 ▼  
 MODE


START  
 STOP

2018 SN 5B1401147

**DIGITAL BLOOD PRESSURE MONITOR**  
 Designed by A&D Medical, Tokyo Japan Model: UM-211  
 Measuring range: 0-299mmHg  
 Rechargeable Battery: 3.6V(UM-211-20)  
 AC Adapter: 6V(TB-268)

A&D Company, Limited  
 Kitamoto-shi, Saitama JAPAN  
 A&D INSTRUMENTS LTD.  
 Abingdon, Oxon OX14 1DY, U.K.

CE 0123



31/8/24  
 മെഡിക്കൽ ഓഫീസർ  
 സാമൂഹികാരോഗ്യകേന്ദ്രം  
 മട്ടന്നൂർ-670 702