


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**


Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC MATTANNUR
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make: COOL SPIGMO Model: NA SN: NA
4	Equipment ID/ Barcode	1330287 #130552
5	Date of purchase/ Year of manufacture/Installation Date	30-06-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 21.08.2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking. Equipment installed on 30.06.2020. Aged up to 04 years 01 month. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory #Based on the period of life and value as per the BER guidelines
 * Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 checked & identified that machine not switching ON, due to main board, pressure sensor and Pump are defective. The spare cost is economically not feasible. Aged up to 4 years & above. So recommended for BER as Per BER protocol.

 30/08/2024 Signature of JC BM (NHM)



Date: 31-08-2024


 Signature of Superintendent Medical Officer (i/c)
 സാമൂഹികാരോഗ്യകേന്ദ്രം
 കൂന്തളം-670 702



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234862

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC MATTANNUR</u></p> <p>Address <u>MATTANNUR</u></p> <p style="padding-left: 40px;"><u>KANNUR</u></p> <p>Ph : <u>09446513548</u></p>	<p>Call Registration Date : <u>21/08/2024</u></p> <p>Caller ID : <u>130552</u></p> <p>Date of Visit : <u>22/08/2024</u></p> <p>Asset No. : <u>1330287</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>SOUL SPIGMO</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment is not switching on.

Action Taken : checked the equipment and found equipment is not switching on, found that Main Based, pressure sensor and pump Defective. Need to replace these spare for further checking.

Completed Date : 22/08/2024 Time : 12:45pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Prathyugh T. Bhaskeer</u>	<u>22/08/2024</u>	<u>12:10pm</u>	<u>12:45pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Prathyugh T. Bhaskeer</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>22/08/2024</u></p> <p>Contact Number : <u>9947029608</u></p>	<p>Customer Name : <u>Jessy KJ</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>22/8/24</u></p> <p>Contact Number : <u>9446517046</u></p> <p>Designation : <u>Senior Officer</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	---





മെഡിക്കൽ ഓഫീസർ
സാമൂഹികാരോഗ്യകേന്ദ്രം
മട്ടന്നൂർ-670 702