





**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
 BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: <b>DH WADAKKANCHERRY</b>	Name of District: <b>Thrissur</b>
Name of Equipment: <b>Sterlizer (Flash Autoclave)</b>	Equipment ID & Barcode: <b>125467 &amp; 0812270</b>
Make: <b>GETIDY</b>	Model: <b>JY23</b>
Serial Number: <b>201207106</b>	Warranty details: <b>No Warranty</b>
Date of purchase/Year of manufacture/Installation Date: <b>Ticket master installation date- 11-05-2012</b> As per Stock book installation date - <b>05-11-2012</b>	Present status of the equipment: <b>fully damaged</b>
Date of breakdown: <b>25-07-2024</b> Date of registration of complaint through email/ Toll free): <b>Toll Free</b>	*AMC/CAMC Period agreed at the time of purchase: <b>No AMC/CAMC</b>
Action taken: <b>Machine checked and found steamer, control board and transformer are defective. Need to replace these spares for further checking working condition of equipment and spare enquired to oem.</b>	
Recommendations for repair (required service details) - <b>No Recommendations</b>	
Cost of spares (specify parts and cost): <b>N/A (Obsolete model)</b>	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: <b>N/A</b>	Asset Value: <b>82467/-</b>
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) - <b>OEM letter attached and cyrix report attached</b>	
Reasons for recommending the equipment as BER: <b>Checked the machine and found Steamer, control board and transformer are defective. Supplier informed that these parts are no longer available due to an update in the machine's design, making the model obsolete. The new parts are only compatible with the latest version, which does not fit for this older model. In ticket master installation date is 11-05-2012 and As per stock book installation date is 05-11-2012 aged up to 08 years 11 months and also the model is obsolete as the supplier information. So as per tender clause 5.3.14.2 we recommending the equipment for condemnation.</b>	
<b>BIJO T JOY</b>  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM:  <p align="center"><i>obsolete model</i></p>	
Recommended for BER (Yes/ No):	 Signature of JC BM (NHM) <b>Jr. Consultant (Biomedical)</b> <b>NHM, Thrissur</b>
Date: <b>02/9/24</b>	
	 Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory # Based on the period of life and value as per the BER guidelines \* Attach Photograph

**SUPERINTENDENT**  
**DISTRICT HOSPITAL, WADAKKANCHERRY**  
**KUMARANELLUR PO THRISUR-680 590**



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

No : 1000183

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>DH Wadakkancherry</u>	Call Registration Date : <u>25/07/2024</u>
Address: <u>Wadakkancherry,</u>	Caller ID : <u>125467</u>
<u>Thrissur</u>	Date of Visit : <u>26/07/2024</u>
Ph: <u>9605893658</u>	Asset No : <u>0812270</u>
	EQPT Name : <u>Sterlizer</u>
	Manufacture: <u>Credity</u> Model: <u>JY23</u>
	S. No: <u>201207106</u> Dept: <u>Dental Department</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Equipment is not working.

Action Taken : checked the machine and found steamer, control board and transformer are defective. Need to replace these spares for further checking. working condition of equipment and spare required to OEM

Completed  Date : 26-07-2024 Time : 2:30 pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ASwin. A.R</u>	<u>26/7/2024</u>	<u>12:00 Pm</u>	<u>2:30 Pm</u>

Customer Remark Completed  Pending

Service Engineer Name : ASwin. A.R


Signature: [Signature]

Date: 26/7/2024

Contact Number: 85899 65265

Customer Name : Dr. Shabeer K. P. A

Signature: [Signature]  
Date: [Date]  
Contact Number: [Number]  
Designation : [Designation]  
Hospital Seal: [Seal]



[Handwritten initials]

# **JAIGURU SALES & SERVICES**

The Team For Sterilization

XIII/385, N.Paravur, Ernakulam, Kerala-683519

Ph: 8089960227  
Mob: 09447043821

To

Cyrix Health care Pvt. Ltd,Ernakulam

Ref: Qtn. /88/2024-25 Dated 19-08-2024

Sir,

**Sub: Flash Autoclave Repair at DH Vadakkancheri, Thrissur Dist.**

After inspecting your GETIDY machine (Serial No. 201207106), we found that the steamer and control boards, including the transformer, need replacement.

However, these parts are no longer available due to an update in the machine's design, making the model obsolete. The new parts are only compatible with the latest version, which does not fit your older model.

We recommend discarding the equipment and considering an upgrade to the latest version for continued support.

So please do the needful,

Expecting your valued co operation,

For **Jaiguru Sales & Service**

  
Proprietor  
Remesh Kumar





ospital - DH wadakkanchery equipment - flash autoclave barcode - 0812270 caller id - 125467

7 messages

Biyo T Joy <bijocyrix@gmail.com>  
To: Jaiguru Sales & Service Kerala <Jssindiaster@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Wed, 7 Aug, 2024 at 9:0

dear sir,

It is inform you that flash autoclave complaint registered at DH wadakkanchery hospital.kindly share the quotation of the coil and trnsfomer.  
hospital - DH wadakkanchery  
equipment - flash autoclave  
barcode - 0812270  
caller id - 125467  
make - LY PLANS SYSTEMS  
model - GETIDY&JY23  
sr.no -20120710

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Biyo T Joy <bijocyrix@gmail.com>  
To: Jaiguru Sales & Service Kerala <Jssindiaster@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 8 Aug, 2024 at 8:1

reminder

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

Biyo T Joy <bijocyrix@gmail.com>  
To: Jaiguru Sales & Service Kerala <Jssindiaster@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Mon, 12 Aug, 2024 at 11:3

reminder

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

Biyo T Joy <bijocyrix@gmail.com>  
To: Jaiguru Sales & Service Kerala <Jssindiaster@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Tue, 13 Aug, 2024 at 6:4

dear sir,

kindly share the quotation

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

**Bijo T Joy** <bijocyrix@gmail.com>  
To: Jaiguru Sales & Service Kerala <jssindiaster@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Wed, 14 Aug, 2024 at 7:5

reminder

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

**Bijo T Joy** <bijocyrix@gmail.com>  
To: Jaiguru Sales & Service Kerala <jssindiaster@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Mon, 19 Aug, 2024 at 7:5

reminder

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

**Jaiguru Sales & Service Kerala** <jssindiaster@gmail.com>  
To: Bijo T Joy <bijocyrix@gmail.com>

Mon, 19 Aug, 2024 at 9:5

[Quoted text hidden]


-  
For Jaiguru Sales & Service  
Remesh Kumar  
Proprietor  
Jaiguru Sales & Service  
N.Paravur, Ernakulam, Kerala-India  
Ph: 8089960227 Mob: 09447043821

# Autoclave:

5/11/2012	Flash Autoclave			Cost.	Store P.No 119/10
	Model No: Jy-23	1 NO			
	Made: Zhejiang Getobuy Medical Instruments Co. Ltd.			85,765.08	

## Fuji-IX - Continued

28/11/2020	FUJIX	?			
	29/11/2022		1	2	
	30-12-22		1	1	
	2-2-2023		1	0	





**DENTAL**

**23L**

**STICKER**

Supplied & Maintained by  
**KERALA MEDICAL SERVICES CORPORATION LTD.**  
 (Department of Health & Family Welfare, Govt. of Kerala)  
 Thycaud, P.O. Thiruvananthapuram-695014

Under KMSCL Scheme

Tender No. KMSCL/IT/2011/0119 Eqt. Code 445-4

Purchase order (KMSCL) 377/2011/0119

Installed on 15/01/2012 Warranty (CMC/AMC) Upto 31/03/12

For Repair/Service/Maintenance  
 Contact **ENGINEERING DIVISION, KMSCL**  
**Toll Free - 1800 425 0004**  
 Fax: 0471-4015332



STICKER  
 TOLL FREE NO: 1800-425-0004  
 0471-4015332

CYRUS HEALTH CARE PVT.LTD  
 BIDDY SERVICE PROVIDER  
 PREVENTIVE MAINTENANCE SERVICE




(4 14)68(204)08(2 1)04 00

**Getidy** **CE** 0197  
 220V-230V,50Hz,1500W  
 Model: JY-23

**Notice READ OPERATION MANUAL**

2012-07 **ECREP**



S.N.: 201207106

Nymplan Systems  
 10 Woodhill Drive, Preston  
 Manchester M25 0AD  
 浙江凯德医疗器械有限公司  
 ZHEJIANG GETIDY MEDICAL  
 ADD: TONGQIN INDUSTRIAL ZONE, WUYI, ZHEJIANG, CHINA